



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/17/2018

TEMP1 IN: 103 09/17/2018 07:35 TEMP OUT: 104 09/17/2018 08:40

TEMP2 IN: 104 09/17/2018 09:20 TEMP OUT: 104 09/17/2018 10:30

TEMP3 IN: 105 09/17/2018 11:30 TEMP OUT: 105 09/17/2018 13:10

TEMP4 IN: 105 09/17/2018 14:00 TEMP OUT: 105 09/17/2018 15:35

Run Number: LB97942

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97942BL	LB97942BL	1.4021	1.4021	100	1.4021	1.4021	1.4021	0.0000	0
2	LB97942BS	LB97942BS	1.4121	1.4121	100	1.4620	1.4620	1.4620	0.0499	499
3	J4929-13	SP-PIT-11-BASELINE-WATER	1.3877	1.3877	10	2.4937	2.4937	2.4937	1.1060	110600
4	J4931-01	091418-SIDI-E-D-B	1.4188	1.4188	200	1.4270	1.4270	1.4270	0.0082	41
5	J4931-01DUP	091418-SIDI-E-D-BDUP	1.3984	1.3984	200	1.4070	1.4070	1.4070	0.0086	43
6	J4931-02	091418-SIDI-E-D-S	1.4118	1.4118	300	1.4240	1.4240	1.4240	0.0122	40.7
7	J4931-03	091418-SIDI-E-U-B	1.3974	1.3974	300	1.4109	1.4109	1.4109	0.0135	45
8	J4931-04	091418-SIDI-E-U-S	1.4123	1.4123	300	1.4256	1.4256	1.4256	0.0133	44.3
9	J4931-05	091418-DBRI-E-D-B	1.4006	1.4006	200	1.4179	1.4179	1.4179	0.0173	86.5
10	J4931-06	091418-DBRI-E-D-S	1.4041	1.4041	300	1.4182	1.4182	1.4182	0.0141	47
11	J4931-07	091418-DBRI-E-U-B	1.3951	1.3951	300	1.4122	1.4122	1.4122	0.0171	57
12	J4931-08	091418-DBRI-E-U-S	1.4102	1.4102	300	1.4251	1.4251	1.4251	0.0149	49.7
13	J4931-09	091418-JETT-E-D-B	1.4027	1.4027	300	1.4176	1.4176	1.4176	0.0149	49.7
14	J4931-10	091418-JETT-E-D-S	1.4128	1.4128	300	1.4217	1.4217	1.4217	0.0089	29.7
15	J4931-11	091418-JETT-E-U-B	1.4218	1.4218	300	1.4362	1.4362	1.4362	0.0144	48



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/17/2018

TEMP1 IN: 103 09/17/2018 07:35 TEMP OUT: 104 09/17/2018 08:40

TEMP2 IN: 104 09/17/2018 09:20 TEMP OUT: 104 09/17/2018 10:30

TEMP3 IN: 105 09/17/2018 11:30 TEMP OUT: 105 09/17/2018 13:10

TEMP4 IN: 105 09/17/2018 14:00 TEMP OUT: 105 09/17/2018 15:35

Run Number: LB97942

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4931-12	091418-JETT-E-U-S	1.3939	1.3939	200	1.4015	1.4015	1.4015	0.0076	38
17	J4931-12DUP	091418-JETT-E-U-SDUP	1.3980	1.3980	200	1.4057	1.4057	1.4057	0.0077	38.5
18	J4933-01	001M-WILLETS-PT-BLVD (SEP)	1.4122	1.4122	100	1.4187	1.4187	1.4187	0.0065	65

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS091718-2

WorkList ID : 116770

Date : 9/17/2018 7:49:17 AM

2397942

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/19/2018	Water	J4929-13 C	TSS	Cool 4 deg C	PSEG01	I31	SP-PIT-11-BASELINE-WATEF	09/14/2018	SM2540 D
09/21/2018	Water	J4931-01 C	TSS	Cool 4 deg C	HDR103	I21	091418-SIDI-E-D-B	09/14/2018	SM2540 D
09/21/2018	Water	J4931-02 C	TSS	Cool 4 deg C	HDR103	I21	091418-SIDI-E-D-S	09/14/2018	SM2540 D
09/21/2018	Water	J4931-03 C	TSS	Cool 4 deg C	HDR103	I21	091418-SIDI-E-U-B	09/14/2018	SM2540 D
09/21/2018	Water	J4931-04 C	TSS	Cool 4 deg C	HDR103	I21	091418-SIDI-E-U-S	09/14/2018	SM2540 D
09/21/2018	Water	J4931-05 C	TSS	Cool 4 deg C	HDR103	I21	091418-DBRI-E-D-B	09/14/2018	SM2540 D
09/21/2018	Water	J4931-06 C	TSS	Cool 4 deg C	HDR103	I21	091418-DBRI-E-D-S	09/14/2018	SM2540 D
09/21/2018	Water	J4931-07 C	TSS	Cool 4 deg C	HDR103	I21	091418-DBRI-E-U-B	09/14/2018	SM2540 D
09/21/2018	Water	J4931-08 C	TSS	Cool 4 deg C	HDR103	I21	091418-DBRI-E-U-S	09/14/2018	SM2540 D
09/21/2018	Water	J4931-09 C	TSS	Cool 4 deg C	HDR103	I21	091418-JETT-E-D-B	09/14/2018	SM2540 D
09/21/2018	Water	J4931-10 C	TSS	Cool 4 deg C	HDR103	I21	091418-JETT-E-D-S	09/14/2018	SM2540 D
09/21/2018	Water	J4931-11 C	TSS	Cool 4 deg C	HDR103	I21	091418-JETT-E-U-B	09/14/2018	SM2540 D
09/21/2018	Water	J4931-12 C	TSS	Cool 4 deg C	HDR103	I21	091418-JETT-E-U-S	09/14/2018	SM2540 D
09/19/2018	Water	J4933-01 B	TSS	Cool 4 deg C	TULL01	I42	001M-WILLETS-PT-BLVD(SE)	09/12/2018	SM2540 D

Date/Time 09/17/18 10:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 09/17/18 12:30
 Received by: [Signature]
 Relinquished by: [Signature]