

LB 97948


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35788



A= 0.0000e+000

B= 2.7883e-004

C= -2.1632e-003

Rho= 0.9997256

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.008	-0.008	-21	0.000	-21				
0.2	0.200	0.210	0.010	760	0.0 %	760				
2.5	2.500	2.406	-0.094	8638	0.0 %	8638				
5.0	5.000	5.177	0.177	18576	0.0 %	18576				
7.5	7.500	7.438	-0.062	26682	0.0 %	26682				
10.0	10.000	9.977	-0.023	35788	0.0 %	35788				

LB97948 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Sep 2018 10:21:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.0000	7470A	PPB	17 Sep 2018 10:29:49	S	Std
0.2 - 1	50.2	760	-	0.2000	7470A	PPB	17 Sep 2018 10:31:58	S	Std
2.5 - 1	52.5	8638	-	2.5000	7470A	PPB	17 Sep 2018 10:34:07	S	Std
5.0 - 1	55	18576	-	5.0000	7470A	PPB	17 Sep 2018 10:36:17	S	Std
7.5 - 1	57.5	26682	-	7.5000	7470A	PPB	17 Sep 2018 10:38:25	S	Std
10.0 - 1	510	35788	-	10.0000	7470A	PPB	17 Sep 2018 10:40:34	S	Std
ICV66 - 1	ICV66	14202	3.9578	-	7470A	PPB	17 Sep 2018 10:43:09	U	SMPL
ICB66 - 1	ICB66	-178	-0.0518	-	7470A	PPB	17 Sep 2018 10:45:18	U	SMPL
CCV69 - 1	CCV69	18913	5.2714	-	7470A	PPB	17 Sep 2018 10:47:26	U	SMPL
CCB69 - 1	CCB69	-178	-0.0518	-	7470A	PPB	17 Sep 2018 10:49:33	U	SMPL
CRA - 1	CRA	841	0.2323	-	7470A	PPB	17 Sep 2018 10:54:18	U	SMPL
HighStd - 1	HighStd	36908	10.2890	-	7470A	PPB	17 Sep 2018 10:56:25	U	SMPL
ChkStd - 1	ChkStd	25636	7.1460	-	7470A	PPB	17 Sep 2018 10:58:31	U	SMPL
PB112875BL - 1	PBW	-222	-0.0641	-	7470A	PPB	17 Sep 2018 11:00:41	U	SMPL
PB112875BS - 1	LCSW	12807	3.5689	-	7470A	PPB	17 Sep 2018 11:02:51	U	SMPL
PB112875TB - 1	PB112828TB	-152	-0.0445	-	7470A	PPB	17 Sep 2018 11:05:02	U	SMPL
J4898-03 - 1	ROLLOFF01-COMP-FILL	88	0.0224	-	7470A	PPB	17 Sep 2018 11:07:17	U	SMPL
J4898-04 - 1	ROLLOFF02-COMP-FILL	150	0.0397	-	7470A	PPB	17 Sep 2018 11:09:25	U	SMPL
J4906-04 - 1	SP-PIT-3	-92	-0.0278	-	7470A	PPB	17 Sep 2018 11:11:34	U	SMPL
J4906-08 - 1	SP-PIT-1	36	0.0079	-	7470A	PPB	17 Sep 2018 11:13:40	U	SMPL
CCV70 - 1	CCV70	18745	5.2246	-	7470A	PPB	17 Sep 2018 11:15:47	U	SMPL
CCB70 - 1	CCB70	-174	-0.0507	-	7470A	PPB	17 Sep 2018 11:17:57	U	SMPL
J4906-12 - 1	SP-PIT-7	-54	-0.0172	-	7470A	PPB	17 Sep 2018 11:20:04	U	SMPL
J4906-16 - 1	SP-PIT-8	113	0.0293	-	7470A	PPB	17 Sep 2018 11:22:10	U	SMPL
J4906-20 - 1	SP-PIT-9	47	0.0109	-	7470A	PPB	17 Sep 2018 11:24:19	U	SMPL
J4906-24 - 1	SP-PIT-10	26	0.0051	-	7470A	PPB	17 Sep 2018 11:26:26	U	SMPL
J4906-28 - 1	SP-PIT-4	53	0.0126	-	7470A	PPB	17 Sep 2018 11:28:33	U	SMPL
J4906-32 - 1	SP-PIT-2	100	0.0257	-	7470A	PPB	17 Sep 2018 11:30:40	U	SMPL
J4906-32DUP - 1	SP-PIT-2DUP	89	0.0227	-	7470A	PPB	17 Sep 2018 11:32:48	U	SMPL
J4906-32MS - 1	SP-PIT-2MS	12210	3.4024	-	7470A	PPB	17 Sep 2018 11:34:56	U	SMPL
J4906-32MSD - 1	SP-PIT-2MSD	12192	3.3974	-	7470A	PPB	17 Sep 2018 11:37:06	U	SMPL
CCV71 - 1	CCV71	18906	5.2695	-	7470A	PPB	17 Sep 2018 11:39:14	U	SMPL
CCB71 - 1	CCB71	-248	-0.0713	-	7470A	PPB	17 Sep 2018 11:41:22	U	SMPL
J4906-32LX5 - 1		130	0.0341	-	7470A	PPB	17 Sep 2018 11:45:56	U	SMPL
J4906-32A - 1		14363	4.0027	-	7470A	PPB	17 Sep 2018 11:48:02	U	SMPL
CCV72 - 1	CCV72	18583	5.1794	-	7470A	PPB	17 Sep 2018 11:50:12	U	SMPL
CCB72 - 1	CCB72	-205	-0.0593	-	7470A	PPB	17 Sep 2018 11:52:20	U	SMPL