



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97949

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	09/17/2018	10:55
2	J4907-01		1	7	>100.0	>212.0	>212.0	09/17/2018	11:15
3	J4907-01DUP		1	7	>100.0	>212.0	>212.0	09/17/2018	11:35
4	J4907-03		1	8	>100.0	>212.0	>212.0	09/17/2018	11:55
5	J4907-05		1	6	>100.0	>212.0	>212.0	09/17/2018	12:20
6	J4907-07		1	8	>100.0	>212.0	>212.0	09/17/2018	12:48
7	J4907-09		1	7	>100.0	>212.0	>212.0	09/17/2018	13:15
8	J4907-11		1	9	>100.0	>212.0	>212.0	09/17/2018	13:45
9	J4929-14		1	8	>100.0	>212.0	>212.0	09/17/2018	14:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97949

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91718

WorkList ID : 116800

Date : 9/17/2018 10:27:28 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/27/2018	Water	J4907-01	Flash Point	Cool 4 deg C	TULL02	I11	CBH-591(1)	09/13/2018	1010A
09/27/2018	Water	J4907-03	Flash Point	Cool 4 deg C	TULL02	I11	CBH-591(1)HOSE	09/13/2018	1010A
09/27/2018	Water	J4907-05	Flash Point	Cool 4 deg C	TULL02	I11	CBH-128(2)	09/13/2018	1010A
09/27/2018	Water	J4907-07	Flash Point	Cool 4 deg C	TULL02	I11	CBH-594(3)	09/13/2018	1010A
09/27/2018	Water	J4907-09	Flash Point	Cool 4 deg C	TULL02	I11	CBH-140(4)	09/13/2018	1010A
09/27/2018	Water	J4907-11	Flash Point	Cool 4 deg C	TULL02	I11	CBH-139(5)	09/13/2018	1010A
09/19/2018	Water	J4929-14	Flash Point	Cool 4 deg C	PSEG01	I31	SP-PIT-11-BASELINE-WATEF	09/14/2018	1010A

Date/Time 09/17/18 10:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 09/17/18 14:15
 Received by: CP
 Relinquished by: KS