


245.1

LB 97952
INSTRUMENT ID: CY2

Linear

 μ Abs.:

34116



A= 0.0000e+000

B= 2.9402e-004

C= -3.3836e-002

Rho= 0.9999621

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	55	0.000	55				
0.05	0.050	0.059	0.009	315	0.0 %	315				
0.2	0.200	0.232	0.032	904	0.0 %	904				
2.5	2.500	2.442	-0.058	8421	0.0 %	8421				
5.0	5.000	5.049	0.049	17288	0.0 %	17288				
7.5	7.500	7.489	-0.011	25586	0.0 %	25586				
10.0	10.000	9.997	-0.003	34116	0.0 %	34116				

LB97952 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 17 Sep 2018 12:51:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	55	-	0.0000	245.1	PPB	17 Sep 2018 12:53:45	S	Std
0.05 - 1	50.05	315	-	0.0500	245.1	PPB	17 Sep 2018 12:56:03	S	Std
0.2 - 1	50.2	904	-	0.2000	245.1	PPB	17 Sep 2018 13:01:10	S	Std
2.5 - 1	52.5	8421	-	2.5000	245.1	PPB	17 Sep 2018 13:03:28	S	Std
5.0 - 1	55	17288	-	5.0000	245.1	PPB	17 Sep 2018 13:05:46	S	Std
7.5 - 1	57.5	25586	-	7.5000	245.1	PPB	17 Sep 2018 13:08:03	S	Std
10.0 - 1	510	34116	-	10.0000	245.1	PPB	17 Sep 2018 13:12:59	S	Std
ICV67 - 1	ICV67	13620	3.9707	-245.1		PPB	17 Sep 2018 13:18:18	U	SMPL
ICB67 - 1	ICB67	-70	-0.0544	-245.1		PPB	17 Sep 2018 13:20:33	U	SMPL
CCV73 - 1	CCV73	16749	4.8907	-245.1		PPB	17 Sep 2018 13:22:51	U	SMPL
CCB73 - 1	CCB73	-157	-0.0800	-245.1		PPB	17 Sep 2018 13:25:07	U	SMPL
CRA - 1	CRA	260	0.0426	-245.1		PPB	17 Sep 2018 13:27:24	U	SMPL
HighStd - 1	HighStd	33719	9.8801	-245.1		PPB	17 Sep 2018 13:29:43	U	SMPL
ChkStd - 1	ChkStd	24220	7.0873	-245.1		PPB	17 Sep 2018 13:31:59	U	SMPL
PB112876BS - 1	LCSW	13070	3.8090	-245.1		PPB	17 Sep 2018 13:36:37	U	SMPL
J4901-01 - 1	091218-TIPC-F-D-B	-92	-0.0609	-245.1		PPB	17 Sep 2018 13:38:54	U	SMPL
J4901-02 - 1	091218-TIPC-F-D-S	51	-0.0188	-245.1		PPB	17 Sep 2018 13:41:11	U	SMPL
J4901-03 - 1	091218-TIPC-F-U-B	16	-0.0291	-245.1		PPB	17 Sep 2018 13:43:29	U	SMPL
J4901-04 - 1	091218-TIPC-F-U-S	102	-0.0038	-245.1		PPB	17 Sep 2018 13:45:46	U	SMPL
J4901-05 - 1	091218-DBRI-F-D-B	147	0.0094	-245.1		PPB	17 Sep 2018 13:48:02	U	SMPL
CCV74 - 1	CCV74	16807	4.9077	-245.1		PPB	17 Sep 2018 13:50:18	U	SMPL
CCB74 - 1	CCB74	-165	-0.0823	-245.1		PPB	17 Sep 2018 13:52:34	U	SMPL
J4901-06 - 1	091218-DBRI-F-D-M	28	-0.0256	-245.1		PPB	17 Sep 2018 13:54:52	U	SMPL
J4901-07 - 1	091218-DBRI-F-D-S	128	0.0038	-245.1		PPB	17 Sep 2018 13:57:07	U	SMPL
J4901-08 - 1	091218-DBRI-F-U-B	106	-0.0027	-245.1		PPB	17 Sep 2018 13:59:23	U	SMPL
J4901-09 - 1	091218-DBRI-F-U-M	87	-0.0083	-245.1		PPB	17 Sep 2018 14:01:39	U	SMPL
J4901-10 - 1	091218-DBRI-F-U-S	116	0.0003	-245.1		PPB	17 Sep 2018 14:03:57	U	SMPL
J4901-11 - 1	091218-SIDI-F-D-B	116	0.0003	-245.1		PPB	17 Sep 2018 14:06:17	U	SMPL
J4901-12 - 1	091218-SIDI-F-D-M	144	0.0085	-245.1		PPB	17 Sep 2018 14:08:36	U	SMPL
J4901-13 - 1	091218-SIDI-F-D-S	96	-0.0056	-245.1		PPB	17 Sep 2018 14:10:54	U	SMPL
J4901-14 - 1	091218-SIDI-F-U-B	80	-0.0103	-245.1		PPB	17 Sep 2018 14:13:11	U	SMPL
J4901-15 - 1	091218-SIDI-F-U-M	56	-0.0174	-245.1		PPB	17 Sep 2018 14:15:28	U	SMPL
CCV75 - 1	CCV75	16533	4.8271	-245.1		PPB	17 Sep 2018 14:17:44	U	SMPL
CCB75 - 1	CCB75	-137	-0.0741	-245.1		PPB	17 Sep 2018 14:20:02	U	SMPL
J4901-16 - 1	091218-SIDI-F-U-S	201	0.0253	-245.1		PPB	17 Sep 2018 14:22:21	U	SMPL
J4901-16DUP - 1	091218-SIDI-F-U-SDUP	284	0.0497	-245.1		PPB	17 Sep 2018 14:24:38	U	SMPL
J4901-16MS - 1	091218-SIDI-F-U-SMS	15097	4.4049	-245.1		PPB	17 Sep 2018 14:26:55	U	SMPL
J4901-16MSD - 1	091218-SIDI-F-U-SMSD	15437	4.5049	-245.1		PPB	17 Sep 2018 14:29:12	U	SMPL
PB112877BL - 1	PBW	-169	-0.0835	-245.1		PPB	17 Sep 2018 14:31:30	U	SMPL
PB112877BS - 1	LCSW	12673	3.6922	-245.1		PPB	17 Sep 2018 14:33:48	U	SMPL
J4907-01 - 1	CBH-591(1)	654	0.1585	-245.1		PPB	17 Sep 2018 14:36:05	U	SMPL
J4907-01DUP - 1	CBH-591(1)DUP	526	0.1208	-245.1		PPB	17 Sep 2018 14:38:24	U	SMPL
J4907-01MS - 1	CBH-591(1)MS	14889	4.3438	-245.1		PPB	17 Sep 2018 14:40:41	U	SMPL
J4907-01MSD - 1	CBH-591(1)MSD	15259	4.4526	-245.1		PPB	17 Sep 2018 14:42:58	U	SMPL
CCV76 - 1	CCV76	17032	4.9739	-245.1		PPB	17 Sep 2018 14:45:15	U	SMPL
CCB76 - 1	CCB76	-71	-0.0547	-245.1		PPB	17 Sep 2018 14:47:34	U	SMPL
J4907-03 - 1	CBH-591(1)HOSE	184	0.0203	-245.1		PPB	17 Sep 2018 14:49:51	U	SMPL
J4907-05 - 1	CBH-128(2)	10843	3.1542	-245.1		PPB	17 Sep 2018 14:52:07	U	SMPL
J4907-07 - 1	CBH-594(3)	1163	0.3081	-245.1		PPB	17 Sep 2018 14:54:24	U	SMPL
J4907-09 - 1	CBH-140(4)	330	0.0632	-245.1		PPB	17 Sep 2018 14:56:42	U	SMPL
J4907-11 - 1	CBH-139(5)	46	-0.0203	-245.1		PPB	17 Sep 2018 14:58:59	U	SMPL
PB112876BL - 1	PBW	122	0.0020	-245.1		PPB	17 Sep 2018 15:01:18	U	SMPL
J4901-16LX5 - 1		113	-0.0006	-245.1		PPB	17 Sep 2018 15:03:35	U	SMPL
J4901-16A - 1		17104	4.9950	-245.1		PPB	17 Sep 2018 15:05:52	U	SMPL
J4907-01LX5 - 1		58	-0.0168	-245.1		PPB	17 Sep 2018 15:08:12	U	SMPL
J4907-01A - 1		16628	4.8551	-245.1		PPB	17 Sep 2018 15:10:30	U	SMPL
CCV77 - 1	CCV77	16924	4.9421	-245.1		PPB	17 Sep 2018 15:12:49	U	SMPL
CCB77 - 1	CCB77	-98	-0.0626	-245.1		PPB	17 Sep 2018 15:15:06	U	SMPL