


7471B

LB97958
INSTRUMENT ID: CV2

Linear

 μ Abs.:

43419



A= 0.0000e+000

B= 2.2979e-004

C= -1.2955e-002

Rho= 0.9996618

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	26	0.000	26				
0.2	0.200	0.234	0.034	1073	0.0 %	1073				
2.5	2.500	2.542	0.042	11120	0.0 %	11120				
5.0	5.000	4.825	-0.175	21052	0.0 %	21052				
7.5	7.500	7.642	0.142	33314	0.0 %	33314				
10.0	10.000	9.964	-0.036	43419	0.0 %	43419				

LB97958 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Sep 2018 16:12:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7471B	PPB	17 Sep 2018 16:14:27	S	Std
0.2 - 1	50.2	1073	-	0.2000	7471B	PPB	17 Sep 2018 16:16:44	S	Std
2.5 - 1	52.5	11120	-	2.5000	7471B	PPB	17 Sep 2018 16:19:02	S	Std
5.0 - 1	55	21052	-	5.0000	7471B	PPB	17 Sep 2018 16:21:19	S	Std
7.5 - 1	57.5	33314	-	7.5000	7471B	PPB	17 Sep 2018 16:23:37	S	Std
10.0 - 1	510	43419	-	10.0000	7471B	PPB	17 Sep 2018 16:28:28	S	Std
ICV68 - 1	ICV68	18279	4.1874	-7471B	PPB	17 Sep 2018 16:31:20	U	SMPL	
ICB68 - 1	ICB68	-311	-0.0844	-7471B	PPB	17 Sep 2018 16:33:37	U	SMPL	
CCV78 - 1	CCV78	21487	4.9245	-7471B	PPB	17 Sep 2018 16:35:54	U	SMPL	
CCB78 - 1	CCB78	-209	-0.0610	-7471B	PPB	17 Sep 2018 16:38:11	U	SMPL	
CRA - 1	CRA	1021	0.2217	-7471B	PPB	17 Sep 2018 16:40:27	U	SMPL	
HighStd - 1	HighStd	44274	10.1607	-7471B	PPB	17 Sep 2018 16:42:43	U	SMPL	
ChkStd - 1	ChkStd	31072	7.1271	-7471B	PPB	17 Sep 2018 16:44:59	U	SMPL	
PB112834BL - 1	PBS	-188	-0.0562	-7471B	PPB	17 Sep 2018 16:47:17	U	SMPL	
PB112834BS - 1	LCSS	16699	3.8243	-7471B	PPB	17 Sep 2018 16:49:35	U	SMPL	
J4775-05 - 1	OR-03-091418-A	3916	0.8869	-7471B	PPB	17 Sep 2018 16:51:54	U	SMPL	
J4775-07 - 1	OR-03-091418-B	1637	0.3632	-7471B	PPB	17 Sep 2018 16:54:11	U	SMPL	
J4898-01 - 1	ROLLOFF01-COMP-FILL	6227	1.4179	-7471B	PPB	17 Sep 2018 16:56:29	U	SMPL	
J4898-02 - 1	ROLLOFF02-COMP-FILL	7840	1.7886	-7471B	PPB	17 Sep 2018 16:58:47	U	SMPL	
J4906-01 - 1	SP-PIT-3	1221	0.2676	-7471B	PPB	17 Sep 2018 17:01:02	U	SMPL	
CCV79 - 1	CCV79	21327	4.8878	-7471B	PPB	17 Sep 2018 17:03:18	U	SMPL	
CCB79 - 1	CCB79	-107	-0.0375	-7471B	PPB	17 Sep 2018 17:05:34	U	SMPL	
J4906-01DUP - 1	SP-PIT-3DUP	1159	0.2534	-7471B	PPB	17 Sep 2018 17:07:52	U	SMPL	
J4906-01MS - 1	SP-PIT-3MS	18086	4.1430	-7471B	PPB	17 Sep 2018 17:10:09	U	SMPL	
J4906-01MSD - 1	SP-PIT-3MSD	18068	4.1389	-7471B	PPB	17 Sep 2018 17:12:25	U	SMPL	
J4906-05 - 1	SP-PIT-1	1941	0.4331	-7471B	PPB	17 Sep 2018 17:14:42	U	SMPL	
J4906-09 - 1	SP-PIT-7	489	0.0994	-7471B	PPB	17 Sep 2018 17:16:59	U	SMPL	
J4906-13 - 1	SP-PIT-8	20387	4.6718	-7471B	PPB	17 Sep 2018 17:19:18	U	SMPL	
J4906-17 - 1	SP-PIT-9	3886	0.8800	-7471B	PPB	17 Sep 2018 17:21:35	U	SMPL	
J4906-21 - 1	SP-PIT-10	397	0.0783	-7471B	PPB	17 Sep 2018 17:23:53	U	SMPL	
J4906-25 - 1	SP-PIT-4	790	0.1686	-7471B	PPB	17 Sep 2018 17:26:11	U	SMPL	
J4906-29 - 1	SP-PIT-2	5930	1.3497	-7471B	PPB	17 Sep 2018 17:28:27	U	SMPL	
CCV80 - 1	CCV80	21626	4.9565	-7471B	PPB	17 Sep 2018 17:30:44	U	SMPL	
CCB80 - 1	CCB80	-115	-0.0394	-7471B	PPB	17 Sep 2018 17:33:02	U	SMPL	
J4906-01LX5 - 1		272	0.0495	-7471B	PPB	17 Sep 2018 17:35:19	U	SMPL	
J4906-01A - 1		18899	4.3298	-7471B	PPB	17 Sep 2018 17:37:35	U	SMPL	
CCV81 - 1	CCV81	21777	4.9912	-7471B	PPB	17 Sep 2018 17:39:52	U	SMPL	
CCB81 - 1	CCB81	-152	-0.0479	-7471B	PPB	17 Sep 2018 17:42:09	U	SMPL	