



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97966

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	09/17/2018	16:30
2	CAL2	1	Water			22.4	7.01	09/17/2018	16:32
3	CAL3	1	Water			22.0	10.04	09/17/2018	16:34
4	ICV	1	Water			22.2	7.00	09/17/2018	16:36
5	CCV1	1	Water			22.1	2.00	09/17/2018	16:50
6	J4960-01	1	Solid	20.02	20	22.3	6.53	09/17/2018	16:52
7	J4960-01DUP	1	Solid	20.03	20	22.4	6.54	09/17/2018	16:54
8	CCV2	1	Water			22.3	12.00	09/17/2018	16:56

WORKLIST(Hardcopy Internal Chain)

LB 97966

WorkList Name : pH091718

WorkList ID : 116818

Date : 9/17/2018 12:40:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2018	Solid	J4960-01	pH	Cool 4 deg C	AREC01	K11	S1	09/17/2018	9045D

Date/Time 09/17/18 4:20
 Received by: SP
 Relinquished by: CP

Date/Time 09/17/18 5:10
 Received by: CP
 Relinquished by: SP