



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97973

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	09/17/2018	16:30
2	CAL2	1	Water			22.4	7.01	09/17/2018	16:32
3	CAL3	1	Water			22.0	10.04	09/17/2018	16:34
4	ICV	1	Water			22.2	7.00	09/17/2018	16:36
5	CCV1	1	Water			22.2	2.00	09/17/2018	17:00
6	J4772-02	1	Solid	20.02	20	23.5	8.53	09/17/2018	17:02
7	J4772-02DUP	1	Solid	20.04	20	23.6	8.54	09/17/2018	17:04
8	CCV2	1	Water			22.1	12.00	09/17/2018	17:06

WORKLIST(Hardcopy Internal Chain)

LB 97973

WorkList Name : CORROSIVITY091718-1

WorkList ID : 116817

Date : 9/17/2018 12:40:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2018	Solid	J4772-02	Corrosivity	Cool 4 deg C	PSEG05	I61	HD-02-091718	09/17/2018	9045D

Date/Time 09/17/18 4:20
 Received by: sl
 Relinquished by: CP

Date/Time 09/17/18 5:10
 Received by: CP
 Relinquished by: Sp