

L397974


7470A

INSTRUMENT ID: CYL

Linear

μ Abs.:

33883



A= 0.0000e+000

B= 2.9461e-004

C= -2.2394e-002

Rho= 0.9999360

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	26	0.000	26				
0.2	0.200	0.171	-0.029	656	0.0 %	656				
2.5	2.500	2.503	0.003	8571	0.0 %	8571				
5.0	5.000	5.088	0.088	17346	0.0 %	17346				
7.5	7.500	7.494	-0.006	25512	0.0 %	25512				
10.0	10.000	9.960	-0.040	33883	0.0 %	33883				

LB97974 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Sep 2018 09:50:47

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	18 Sep 2018 09:55:44	S	Std
0.2 - 1	50.2	656	-	0.2000	7470A	PPB	18 Sep 2018 09:57:55	S	Std
2.5 - 1	52.5	8571	-	2.5000	7470A	PPB	18 Sep 2018 10:00:04	S	Std
5.0 - 1	55	17346	-	5.0000	7470A	PPB	18 Sep 2018 10:02:12	S	Std
7.5 - 1	57.5	25512	-	7.5000	7470A	PPB	18 Sep 2018 10:04:20	S	Std
10.0 - 1	59.5	33883	-	10.0000	7470A	PPB	18 Sep 2018 10:11:24	S	Std
ICV69 - 1	ICV69	13562	3.9731	-	7470A	PPB	18 Sep 2018 10:14:08	U	SMPL
ICB69 - 1	ICB69	-210	-0.0843	-	7470A	PPB	18 Sep 2018 10:16:15	U	SMPL
CCV82 - 1	CCV82	17605	5.1641	-	7470A	PPB	18 Sep 2018 10:18:24	U	SMPL
CCB82 - 1	CCB82	-217	-0.0863	-	7470A	PPB	18 Sep 2018 10:20:31	U	SMPL
CRA - 1	CRA	741	0.1959	-	7470A	PPB	18 Sep 2018 10:25:18	U	SMPL
HighStd - 1	HighStd	34331	10.0917	-	7470A	PPB	18 Sep 2018 10:27:24	U	SMPL
ChkStd - 1	ChkStd	24892	7.3109	-	7470A	PPB	18 Sep 2018 10:29:33	U	SMPL
PB112937BL - 1	PBW	-70	-0.0430	-	7470A	PPB	18 Sep 2018 10:34:07	U	SMPL
PB112937BS - 1	LCSW	12787	3.7447	-	7470A	PPB	18 Sep 2018 10:36:16	U	SMPL
PB112937TB - 1	PB112878TB	-50	-0.0371	-	7470A	PPB	18 Sep 2018 10:40:52	U	SMPL
J4773-02 - 1	CL-01-091418-A	-50	-0.0371	-	7470A	PPB	18 Sep 2018 10:43:05	U	SMPL
J4773-02DUP - 1	CL-01-091418-ADUP	-54	-0.0383	-	7470A	PPB	18 Sep 2018 10:45:13	U	SMPL
J4773-02MS - 1	CL-01-091418-AMS	11848	3.4681	-	7470A	PPB	18 Sep 2018 10:47:22	U	SMPL
J4773-02MSD - 1	CL-01-091418-AMSD	12118	3.5476	-	7470A	PPB	18 Sep 2018 10:49:29	U	SMPL
CCV83 - 1	CCV83	17467	5.1235	-	7470A	PPB	18 Sep 2018 10:51:37	U	SMPL
CCB83 - 1	CCB83	-245	-0.0946	-	7470A	PPB	18 Sep 2018 10:53:46	U	SMPL
J4773-04 - 1	CL-01-091418-B	-67	-0.0421	-	7470A	PPB	18 Sep 2018 10:55:57	U	SMPL
J4773-06 - 1	CL-01-091418-C	-82	-0.0466	-	7470A	PPB	18 Sep 2018 10:58:06	U	SMPL
J4773-08 - 1	CL-01-091418-D	91	0.0044	-	7470A	PPB	18 Sep 2018 11:00:13	U	SMPL
J4773-10 - 1	CL-01-091418-E	26	-0.0147	-	7470A	PPB	18 Sep 2018 11:02:22	U	SMPL
J4774-04 - 1	OK-01-091418	-64	-0.0412	-	7470A	PPB	18 Sep 2018 11:04:29	U	SMPL
J4775-06 - 1	OR-03-091418-A	47	-0.0085	-	7470A	PPB	18 Sep 2018 11:06:37	U	SMPL
J4775-08 - 1	OR-03-091418-B	51	-0.0074	-	7470A	PPB	18 Sep 2018 11:08:45	U	SMPL
J4919-02 - 1	ROLLOFF-03_COMP_FILL	326	0.0736	-	7470A	PPB	18 Sep 2018 11:10:54	U	SMPL
J4929-04 - 1	SP-PIT-6	99	0.0068	-	7470A	PPB	18 Sep 2018 11:13:05	U	SMPL
J4929-08 - 1	SP-PIT-12	18	-0.0171	-	7470A	PPB	18 Sep 2018 11:15:14	U	SMPL
CCV84 - 1	CCV84	17523	5.1400	-	7470A	PPB	18 Sep 2018 11:17:22	U	SMPL
CCB84 - 1	CCB84	-143	-0.0645	-	7470A	PPB	18 Sep 2018 11:19:33	U	SMPL
J4929-12 - 1	SP-PIT-11	0	-0.0224	-	7470A	PPB	18 Sep 2018 11:21:41	U	SMPL
J4935-04 - 1	PHASE-1	36	-0.0118	-	7470A	PPB	18 Sep 2018 11:23:48	U	SMPL
J4936-18 - 1	138-KV	73	-0.0009	-	7470A	PPB	18 Sep 2018 11:25:56	U	SMPL
J4936-22 - 1	138-KV-2	97	0.0062	-	7470A	PPB	18 Sep 2018 11:28:03	U	SMPL
J4936-26 - 1	PHASE-1	125	0.0144	-	7470A	PPB	18 Sep 2018 11:30:11	U	SMPL
J4936-30 - 1	PHASE-2	143	0.0197	-	7470A	PPB	18 Sep 2018 11:32:18	U	SMPL
J4936-34 - 1	PHASE-3	159	0.0244	-	7470A	PPB	18 Sep 2018 11:34:26	U	SMPL
J4936-38 - 1	132-1-VOLT	13	-0.0186	-	7470A	PPB	18 Sep 2018 11:36:35	U	SMPL
J4936-42 - 1	26KV	122	0.0135	-	7470A	PPB	18 Sep 2018 11:38:45	U	SMPL
PB112941BL - 1	PBW	208	0.0389	-	7470A	PPB	18 Sep 2018 11:40:54	U	SMPL
CCV85 - 1	CCV85	17530	5.1421	-	7470A	PPB	18 Sep 2018 11:43:03	U	SMPL
CCB85 - 1	CCB85	-178	-0.0748	-	7470A	PPB	18 Sep 2018 11:45:13	U	SMPL
PB112941BS - 1	LCSW	12724	3.7262	-	7470A	PPB	18 Sep 2018 11:47:21	U	SMPL
PB112941TB - 1	PB112893TB	-12	-0.0259	-	7470A	PPB	18 Sep 2018 11:49:29	U	SMPL
J4936-02 - 1	138kv-1	117	0.0121	-	7470A	PPB	18 Sep 2018 11:51:36	U	SMPL
J4936-02DUP - 1	138kv-1DUP	26	-0.0147	-	7470A	PPB	18 Sep 2018 11:53:43	U	SMPL
J4936-02MS - 1	138kv-1MS	12721	3.7253	-	7470A	PPB	18 Sep 2018 11:55:52	U	SMPL
J4936-02MSD - 1	138kv-1MSD	11642	3.4074	-	7470A	PPB	18 Sep 2018 11:58:02	U	SMPL
J4936-04 - 1	138kv-2	-10	-0.0253	-	7470A	PPB	18 Sep 2018 12:00:13	U	SMPL
J4936-06 - 1	PHASE-1	94	0.0053	-	7470A	PPB	18 Sep 2018 12:02:23	U	SMPL
J4936-08 - 1	PHASE-2	56	-0.0059	-	7470A	PPB	18 Sep 2018 12:04:38	U	SMPL
J4936-10 - 1	PHASE-3	3	-0.0215	-	7470A	PPB	18 Sep 2018 12:06:50	U	SMPL
CCV86 - 1	CCV86	17672	5.1839	-	7470A	PPB	18 Sep 2018 12:09:01	U	SMPL
CCB86 - 1	CCB86	-169	-0.0722	-	7470A	PPB	18 Sep 2018 12:11:11	U	SMPL
J4936-12 - 1	132-1-VOLT	-56	-0.0389	-	7470A	PPB	18 Sep 2018 12:13:20	U	SMPL
J4936-14 - 1	26KV	88	0.0035	-	7470A	PPB	18 Sep 2018 12:15:27	U	SMPL
PB112942BS - 1	LCSW	12638	3.7008	-	7470A	PPB	18 Sep 2018 12:19:43	U	SMPL
PB112942TB - 1	PN112895TB	-200	-0.0813	-	7470A	PPB	18 Sep 2018 12:22:08	U	SMPL
J4940-04 - 1	COMP-1	-4	-0.0236	-	7470A	PPB	18 Sep 2018 12:24:17	U	SMPL
J4940-04DUP - 1	COMP-1DUP	-14	-0.0265	-	7470A	PPB	18 Sep 2018 12:26:28	U	SMPL
J4940-04MS - 1	COMP-1MS	12870	3.7692	-	7470A	PPB	18 Sep 2018 12:28:36	U	SMPL
J4940-04MSD - 1	COMP-1MSD	12121	3.5485	-	7470A	PPB	18 Sep 2018 12:30:45	U	SMPL
J4940-08 - 1	COMP-2	-181	-0.0757	-	7470A	PPB	18 Sep 2018 12:32:54	U	SMPL
CCV87 - 1	CCV87	17932	5.2605	-	7470A	PPB	18 Sep 2018 12:35:05	U	SMPL
CCB87 - 1	CCB87	-227	-0.0893	-	7470A	PPB	18 Sep 2018 12:37:14	U	SMPL
J4941-03 - 1	RB47199RT	3	-0.0215	-	7470A	PPB	18 Sep 2018 12:39:24	U	SMPL
J4941-06 - 1	RB47198RT	37	-0.0115	-	7470A	PPB	18 Sep 2018 12:41:33	U	SMPL
J4942-03 - 1	RB32727RT	95	0.0056	-	7470A	PPB	18 Sep 2018 12:43:42	U	SMPL
J4943-03 - 1	RBR200051	88	0.0035	-	7470A	PPB	18 Sep 2018 12:45:49	U	SMPL
J4944-03 - 1	RB48377RT	20	-0.0165	-	7470A	PPB	18 Sep 2018 12:47:56	U	SMPL
J4945-03 - 1	RB47195RT	-81	-0.0463	-	7470A	PPB	18 Sep 2018 12:50:04	U	SMPL
J4946-03 - 1	RB12983RT	66	-0.0029	-	7470A	PPB	18 Sep 2018 12:52:14	U	SMPL
J4947-03 - 1	97-20	24	-0.0153	-	7470A	PPB	18 Sep 2018 12:54:23	U	SMPL
J4948-03 - 1	RB44766RT	-137	-0.0628	-	7470A	PPB	18 Sep 2018 12:56:31	U	SMPL
J4949-03 - 1	153	-36	-0.0330	-	7470A	PPB	18 Sep 2018 12:58:42	U	SMPL
CCV88 - 1	CCV88	17755	5.2083	-	7470A	PPB	18 Sep 2018 13:00:52	U	SMPL
CCB88 - 1	CCB88	-252	-0.0966	-	7470A	PPB	18 Sep 2018 13:02:59	U	SMPL
J4950-03 - 1	RB32734RT	-143	-0.0645	-	7470A	PPB	18 Sep 2018 13:05:09	U	SMPL
J4951-03 - 1	RB48380RT	63	-0.0038	-	7470A	PPB	18 Sep 2018 13:07:17	U	SMPL
J4952-03 - 1	2015-161-20-RB36584RT-RB47184RT	15	-0.0180	-	7470A	PPB	18 Sep 2018 13:09:25	U	SMPL
PB112942BL - 1	PBW	-43	-0.0351	-	7470A	PPB	18 Sep 2018 13:11:35	U	SMPL
J4773-02LX5 - 1		-28	-0.0306	-	7470A	PPB	18 Sep 2018 13:13:44	U	SMPL
J4773-02A - 1		13957	4.0894	-	7470A	PPB	18 Sep 2018 13:15:54	U	SMPL
J4936-02LX5 - 1		-156	-0.0684	-	7470A	PPB	18 Sep 2018 13:18:02	U	SMPL
J4936-02A - 1		13725	4.0211	-	7470A	PPB	18 Sep 2018 13:20:12	U	SMPL
J4940-04LX5 - 1		-202	-0.0819	-	7470A	PPB	18 Sep 2018 13:22:22	U	SMPL
J4940-04A - 1		14442	4.2323	-	7470A	PPB	18 Sep 2018 13:24:32	U	SMPL
CCV89 - 1	CCV89	17972	5.2723	-	7470A	PPB	18 Sep 2018 13:26:42	U	SMPL
CCB89 - 1	CCB89	-115	-0.0563	-	7470A	PPB	18 Sep 2018 13:32:40	U	SMPL