



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB97975
Analysis Date: 09/18/2018
BalanceID: WC SC-3
OvenID: EX OVEN-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 09/18/2018
Extraction IN Time: 10:00
Extraction OUT Time: 10:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB97975BL	LB97975BL	WATER	2	1000	100	3.1699	3.1699	3.02	3.1699	3.1699	0.0000	0
2	LB97975BS	LB97975BS	WATER	2	1000	100	3.1022	3.1022	3.01	3.1211	3.1211	0.0189	18.9
3	LB97975BSD	LB97975BSD	WATER	2	1000	100	3.1714	3.1714	3.03	3.1905	3.1905	0.0191	19.1
4	J4929-13	SP-PIT-11-BASELINE-WAT	WATER	2	700	100	3.0357	3.0357	3.04	3.0362	3.0362	0.0005	0.71
5	J4956-02	GROUNDWATER	WATER	2	1000	100	3.0612	3.0612	3.03	3.0619	3.0619	0.0007	0.7
6	J4957-01	N48966	WATER	2	1000	100	3.1410	3.1410	3.05	3.1752	3.1752	0.0342	34.2



QC Batch# LB97975 Test: TPH Analysis Date: 09/18/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2413
pH Paper 0-14	W2373
Sodium Sulfate	EP1876
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0018</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>12:16</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>11:15</u>		
Bal Check Time:	<u>10:20</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:40</u>
		Out Time:	<u>12:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>13:46</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>13:15</u>		
Bal Check Time:	<u>14:40</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>14:15</u>
		Out Time:	<u>13:45</u>		

WORKLIST(Hardcopy Internal Chain)

LB 07975

WorkList Name : tph j4956

WorkList ID : 116859

Date : 9/18/2018 9:42:06 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/19/2018	Water	J4929-13-1h	TPH	Conc H2SO4 tk	PSEG01	I31	SP-PIT-11-BASELINE-WATEF	09/14/2018	1664A
09/20/2018	Water	J4956-02-1h	TPH	Conc H2SO4 tk	PSEG01	I61	GROUNDWATER	09/17/2018	1664A
09/20/2018	Water	J4957-01-1h	TPH	Conc H2SO4 tk	PSEG01	I61	N48966	09/17/2018	1664A

Date/Time 09.18.18 01:50 PM

Received by: Jn

Relinquished by: [Signature]

Date/Time 09.18.18 10:50 AM

Received by: [Signature]

Relinquished by: [Signature]