

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/18/2018 9:21

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.016	0.0	0.111	
ICB1	-0.002	0.0	0.004	
CCV1	0.991	0.0	0.109	
CCB1	-0.004	0.0	0.004	
PB112872BL	-0.005	0.0	0.004	
PB112872BS	1.030	0.0	0.113	
J4880-26	0.583	0.0	0.066	
CCV2	0.985	0.0	0.108	
CCB2	-0.005	0.0	0.004	

N	9
Mean	0.510
SD	0.5057
CV%	99.18

Aquakem v. 7.2AQ1

Results from time period:

Tue Sep 18 08:16:14 2018

Tue Sep 18 09:18:33 2018

AS

Sample Id	Sam/Ctr/c/	Test short nan	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0472	mg/l	9/18/2018 8:16:14	
0.1PPM	A	Phenolics-	P	0.0996	mg/l	9/18/2018 8:16:15	
0.25PPM	A	Phenolics-	P	0.2484	mg/l	9/18/2018 8:16:16	
0.50PPM	A	Phenolics-	P	0.5074	mg/l	9/18/2018 8:16:17	
1.0PPM	A	Phenolics-	P	0.9979	mg/l	9/18/2018 8:16:18	
2.0PPM	A	Phenolics-	P	1.9995	mg/l	9/18/2018 8:16:19	
ICV1	S	Phenolics-	P	1.0159	mg/l	9/18/2018 8:37:43	
ICB1	S	Phenolics-	P	-0.0025	mg/l	9/18/2018 8:37:44	
CCV1	S	Phenolics-	P	0.9911	mg/l	9/18/2018 9:18:26	
CCB1	S	Phenolics-	P	-0.0036	mg/l	9/18/2018 9:18:27	
PB112872BL	S	Phenolics-	P	-0.0052	mg/l	9/18/2018 9:18:28	
PB112872BS	S	Phenolics-	P	1.0301	mg/l	9/18/2018 9:18:29	
J4880-26	S	Phenolics-	P	0.583	mg/l	9/18/2018 9:18:30	
CCV2	S	Phenolics-	P	0.9849	mg/l	9/18/2018 9:18:32	
CCB2	S	Phenolics-	P	-0.0049	mg/l	9/18/2018 9:18:33	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/18/2018 8:16


Test Phenolics-

Accepted 9/18/2018 8:16

Factor 9.542
Bias 0.005

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0472	0.0500	
2	Phenol-2	0.015	0.0996	0.1000	
3	Phenol-2	0.031	0.2484	0.2500	
4	Phenol-2	0.058	0.5074	0.5000	
5	Phenol-2	0.109	0.9979	1.0000	
6	Phenol-2	0.214	1.9995	2.0000	