



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97978

Ambient Barometric Pressure (mmHg): 750

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	81.2	09/18/2018	09:15
2	J4956-03		1	7	>100.0	>212.0	>212.0	09/18/2018	09:45
3	J4956-03DUP		1	7	>100.0	>212.0	>212.0	09/18/2018	10:20
4	J4957-02		1	9	>100.0	>212.0	>212.0	09/18/2018	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97978

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp91818

WorkList ID : 116857

Date : 9/18/2018 8:45:48 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/20/2018	Water	J4956-03	C Flash Point	Cool 4 deg C	PSEG01	I61	GROUNDWATER	09/17/2018	1010A
09/20/2018	Water	J4957-02	C Flash Point	Cool 4 deg C	PSF	I61	N48966	09/17/2018	1010A

Date/Time 09/18/18 8:50 AM
Received by: KS
Relinquished by: CP

Date/Time 09/18/18 10:50 AM
Received by: CP
Relinquished by: KS