

21397981

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/18/2018 10:20

Test: Phenolics-*mg/d*

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.020	0.0	0.112	
ICB1	-0.005	0.0	0.004	
CCV1	1.012	0.0	0.111	
CCB1	-0.003	0.0	0.004	
PB112871BL	-0.003	0.0	0.004	
PB112871BS	1.020	0.0	0.112	
J4907-03	-0.003	0.0	0.004	
J4907-05	0.007	0.0	0.005	
J4907-07	-0.001	0.0	0.005	
J4907-09	0.009	0.0	0.006	
CCV2	0.993	0.0	0.109	
CCB2	-0.003	0.0	0.004	
J4907-11	0.017	0.0	0.006	
J4880-27	0.583	0.0	0.066	
CCV3	0.990	0.0	0.108	
CCB3	-0.008	0.0	0.004	
J4907-01	-0.006	0.0	0.004	
J4907-01DUP	-0.004	0.0	0.004	
J4907-01MS	0.884	0.0	0.097	
J4907-01MSD	0.878	0.0	0.097	
CCV3 CCV4	0.986	0.0	0.108	
CCB3 CCB4	-0.007	0.0	0.004	
N	22			
Mean	0.380			
SD	0.4762			
CV%	125.40			

AS
09/18/18

Aquakem v. 7.2AQ1

Results from time period:

Tue Sep 18 09:49:14 2018

Tue Sep 18 10:17:54 2018

AS

Sample Id	Sam/Ctr/c/	Test short na	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0472	mg/l	9/18/2018 8:16:14	
0.1PPM	A	Phenolics-	P	0.0996	mg/l	9/18/2018 8:16:15	
0.25PPM	A	Phenolics-	P	0.2484	mg/l	9/18/2018 8:16:16	
0.50PPM	A	Phenolics-	P	0.5074	mg/l	9/18/2018 8:16:17	
1.0PPM	A	Phenolics-	P	0.9979	mg/l	9/18/2018 8:16:18	
2.0PPM	A	Phenolics-	P	1.9995	mg/l	9/18/2018 8:16:19	
ICV1	S	Phenolics-	P	1.0197	mg/l	9/18/2018 9:49:14	
ICB1	S	Phenolics-	P	-0.005	mg/l	9/18/2018 9:49:15	
CCV1	S	Phenolics-	P	1.0122	mg/l	9/18/2018 9:49:16	
CCB1	S	Phenolics-	P	-0.003	mg/l	9/18/2018 9:49:17	
PB112871BL	S	Phenolics-	P	-0.0027	mg/l	9/18/2018 9:49:18	
PB112871BS	S	Phenolics-	P	1.02	mg/l	9/18/2018 9:49:19	
J4907-03	S	Phenolics-	P	-0.003	mg/l	9/18/2018 9:49:24	
J4907-05	S	Phenolics-	P	0.0066	mg/l	9/18/2018 9:49:25	
J4907-07	S	Phenolics-	P	-0.0008	mg/l	9/18/2018 9:56:28	
J4907-09	S	Phenolics-	P	0.0087	mg/l	9/18/2018 9:56:29	
CCV2	S	Phenolics-	P	0.9928	mg/l	9/18/2018 9:56:30	
CCB2	S	Phenolics-	P	-0.0027	mg/l	9/18/2018 9:56:31	
J4907-11	S	Phenolics-	P	0.0168	mg/l	9/18/2018 9:56:32	
J4880-27	S	Phenolics-	P	0.5826	mg/l	9/18/2018 9:56:34	
CCV3	S	Phenolics-	P	0.9899	mg/l	9/18/2018 9:56:35	
CCB3	S	Phenolics-	P	-0.0082	mg/l	9/18/2018 9:56:36	
J4907-01	S	Phenolics-	P	-0.0062	mg/l	9/18/2018 10:17:49	
J4907-01DUP	S	Phenolics-	P	-0.0038	mg/l	9/18/2018 10:17:50	
J4907-01MS	S	Phenolics-	P	0.8844	mg/l	9/18/2018 10:17:51	
J4907-01MSD	S	Phenolics-	P	0.8779	mg/l	9/18/2018 10:17:52	
CCV4	S	Phenolics-	P	0.9862	mg/l	9/18/2018 10:17:53	
CCB4	S	Phenolics-	P	-0.0075	mg/l	9/18/2018 10:17:54	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/18/2018 8:16

Test Phenolics-


Accepted 9/18/2018 8:16

Factor 9.542

Bias 0.005

Coeff. of det. 0.999975

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.010	0.0472	0.0500	
2	Phenol-2	0.015	0.0996	0.1000	
3	Phenol-2	0.031	0.2484	0.2500	
4	Phenol-2	0.058	0.5074	0.5000	
5	Phenol-2	0.109	0.9979	1.0000	
6	Phenol-2	0.214	1.9995	2.0000	