


LB 97990

245.1

INSTRUMENT ID: CV1

 μ Abs.:

33631



Linear

A= 0.0000e+000

B= 2.9508e-004

C= -8.7794e-004

Rho= 0.9998307

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-75	0.000	-75				
0.05	0.050	0.038	-0.012	132	0.0 %	132				
0.2	0.200	0.248	0.048	845	0.0 %	845				
2.5	2.500	2.478	-0.022	8399	0.0 %	8399				
5.0	5.000	4.944	-0.056	16758	0.0 %	16758				
7.5	7.500	7.642	0.142	25900	0.0 %	25900				
10.0	10.000	9.923	-0.077	33631	0.0 %	33631				

LB97990 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 18 Sep 2018 13:59:06					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	-75	-	0.0000	245.1	PPB	18 Sep 2018 14:04:15	S	Std
0.05 - 1	80.05	132	-	0.0500	245.1	PPB	18 Sep 2018 14:06:32	S	Std
0.2 - 1	80.2	845	-	0.2000	245.1	PPB	18 Sep 2018 14:08:51	S	Std
2.5 - 1	82.5	8399	-	2.5000	245.1	PPB	18 Sep 2018 14:11:09	S	Std
5.0 - 1	85	16758	-	5.0000	245.1	PPB	18 Sep 2018 14:13:26	S	Std
7.5 - 1	87.5	25900	-	7.5000	245.1	PPB	18 Sep 2018 14:15:45	S	Std
10.0 - 1	110	33631	-	10.0000	245.1	PPB	18 Sep 2018 14:23:27	S	Std
ICV70 - 1	ICV70	13180	3.8883	-245.1		PPB	18 Sep 2018 14:26:12	U	SMPL
ICB70 - 1	ICB70	-226	-0.0676	-245.1		PPB	18 Sep 2018 14:28:31	U	SMPL
CCV90 - 1	CCV90	16965	5.0052	-245.1		PPB	18 Sep 2018 14:30:48	U	SMPL
CCB90 - 1	CCB90	-249	-0.0744	-245.1		PPB	18 Sep 2018 14:33:04	U	SMPL
CRA - 1	CRA	125	0.0360	-245.1		PPB	18 Sep 2018 14:35:22	U	SMPL
HighStd - 1	HighStd	33374	9.8472	-245.1		PPB	18 Sep 2018 14:37:38	U	SMPL
ChkStd - 1	ChkStd	24162	7.1289	-245.1		PPB	18 Sep 2018 14:39:54	U	SMPL
PB112935BL - 1	PBW	-20	-0.0068	-245.1		PPB	18 Sep 2018 14:46:18	U	SMPL
PB112935BS - 1	LCSW	12953	3.8213	-245.1		PPB	18 Sep 2018 14:48:35	U	SMPL
J4930-01 - 1	091318-JETT-E-D-B	-4	-0.0021	-245.1		PPB	18 Sep 2018 14:50:52	U	SMPL
J4930-02 - 1	091318-JETT-E-D-S	58	0.0162	-245.1		PPB	18 Sep 2018 14:53:11	U	SMPL
J4930-03 - 1	091318-JETT-E-U-B	73	0.0207	-245.1		PPB	18 Sep 2018 14:55:29	U	SMPL
J4930-04 - 1	091318-JETT-E-U-S	118	0.0339	-245.1		PPB	18 Sep 2018 14:57:47	U	SMPL
J4930-05 - 1	091318-DBRI-E-D-B	143	0.0413	-245.1		PPB	18 Sep 2018 15:00:03	U	SMPL
CCV91 - 1	CCV91	17064	5.0344	-245.1		PPB	18 Sep 2018 15:02:20	U	SMPL
CCB91 - 1	CCB91	-273	-0.0814	-245.1		PPB	18 Sep 2018 15:04:36	U	SMPL
J4930-06 - 1	091318-DBRI-E-D-M	103	0.0295	-245.1		PPB	18 Sep 2018 15:06:54	U	SMPL
J4930-07 - 1	091318-DBRI-E-D-S	49	0.0136	-245.1		PPB	18 Sep 2018 15:09:10	U	SMPL
J4930-08 - 1	091318-DBRI-E-U-B	72	0.0204	-245.1		PPB	18 Sep 2018 15:11:26	U	SMPL
J4930-09 - 1	091318-DBRI-E-U-M	138	0.0398	-245.1		PPB	18 Sep 2018 15:13:43	U	SMPL
J4930-11 - 1	091318-TIPC-F-D-B	60	0.0168	-245.1		PPB	18 Sep 2018 15:18:18	U	SMPL
J4930-12 - 1	091318-TIPC-F-D-S	98	0.0280	-245.1		PPB	18 Sep 2018 15:20:35	U	SMPL
J4930-13 - 1	091318-TIPC-F-U-B	58	0.0162	-245.1		PPB	18 Sep 2018 15:22:53	U	SMPL
J4930-14 - 1	091318-TIPC-F-U-S	172	0.0499	-245.1		PPB	18 Sep 2018 15:25:11	U	SMPL
J4930-14DUP - 1	091318-TIPC-F-U-SDUP	136	0.0393	-245.1		PPB	18 Sep 2018 15:27:27	U	SMPL
CCV92 - 1	CCV92	16692	4.9246	-245.1		PPB	18 Sep 2018 15:29:43	U	SMPL
CCB92 - 1	CCB92	-211	-0.0631	-245.1		PPB	18 Sep 2018 15:32:00	U	SMPL
J4930-14MS - 1	091318-TIPC-F-U-SMS	15018	4.4307	-245.1		PPB	18 Sep 2018 15:34:19	U	SMPL
J4930-14MSD - 1	091318-TIPC-F-U-SMSD	15057	4.4422	-245.1		PPB	18 Sep 2018 15:36:35	U	SMPL
PB112936BL - 1	PBW	-112	-0.0339	-245.1		PPB	18 Sep 2018 15:38:51	U	SMPL
PB112936BS - 1	LCSW	12575	3.7098	-245.1		PPB	18 Sep 2018 15:41:09	U	SMPL
J4931-01 - 1	091418-SIDI-E-D-B	-6	-0.0026	-245.1		PPB	18 Sep 2018 15:43:26	U	SMPL
J4931-02 - 1	091418-SIDI-E-D-S	113	0.0325	-245.1		PPB	18 Sep 2018 15:45:43	U	SMPL
J4931-03 - 1	091418-SIDI-E-U-B	155	0.0449	-245.1		PPB	18 Sep 2018 15:48:01	U	SMPL
J4931-04 - 1	091418-SIDI-E-U-S	171	0.0496	-245.1		PPB	18 Sep 2018 15:50:18	U	SMPL
J4931-05 - 1	091418-DBRI-E-D-B	194	0.0564	-245.1		PPB	18 Sep 2018 15:52:36	U	SMPL
J4931-06 - 1	091418-DBRI-E-D-S	299	0.0874	-245.1		PPB	18 Sep 2018 15:54:53	U	SMPL
CCV93 - 1	CCV93	16754	4.9429	-245.1		PPB	18 Sep 2018 15:57:09	U	SMPL
CCB93 - 1	CCB93	-237	-0.0708	-245.1		PPB	18 Sep 2018 15:59:25	U	SMPL
J4931-07 - 1	091418-DBRI-E-U-B	396	0.1160	-245.1		PPB	18 Sep 2018 16:01:44	U	SMPL
J4931-08 - 1	091418-DBRI-E-U-S	278	0.0812	-245.1		PPB	18 Sep 2018 16:04:00	U	SMPL
J4931-09 - 1	091418-JETT-E-D-B	169	0.0490	-245.1		PPB	18 Sep 2018 16:06:17	U	SMPL
J4931-10 - 1	091418-JETT-E-D-S	129	0.0372	-245.1		PPB	18 Sep 2018 16:08:34	U	SMPL
J4931-11 - 1	091418-JETT-E-U-B	189	0.0549	-245.1		PPB	18 Sep 2018 16:10:50	U	SMPL
J4931-12 - 1	091418-JETT-E-U-S	172	0.0499	-245.1		PPB	18 Sep 2018 16:13:07	U	SMPL
J4931-12DUP - 1	091418-JETT-E-U-SDUP	165	0.0478	-245.1		PPB	18 Sep 2018 16:15:24	U	SMPL
J4931-12MS - 1	091418-JETT-E-U-SMS	14047	4.1442	-245.1		PPB	18 Sep 2018 16:17:41	U	SMPL
J4931-12MSD - 1	091418-JETT-E-U-SMSD	14802	4.3669	-245.1		PPB	18 Sep 2018 16:19:59	U	SMPL
J4930-10 - 1	091318-DBRI-E-U-S	-7	-0.0029	-245.1		PPB	18 Sep 2018 16:22:16	U	SMPL
CCV94 - 1	CCV94	16998	5.0149	-245.1		PPB	18 Sep 2018 16:24:34	U	SMPL
CCB94 - 1	CCB94	-225	-0.0673	-245.1		PPB	18 Sep 2018 16:26:52	U	SMPL
J4930-14LX5 - 1		-13	-0.0047	-245.1		PPB	18 Sep 2018 16:38:17	U	SMPL
J4930-14A - 1		16482	4.8627	-245.1		PPB	18 Sep 2018 16:40:34	U	SMPL
J4931-12LX5 - 1		-157	-0.0472	-245.1		PPB	18 Sep 2018 16:42:53	U	SMPL
J4931-12A - 1		16416	4.8432	-245.1		PPB	18 Sep 2018 16:45:12	U	SMPL
CCV95 - 1	CCV95	17079	5.0388	-245.1		PPB	18 Sep 2018 16:47:30	U	SMPL
CCB95 - 1	CCB95	-230	-0.0687	-245.1		PPB	18 Sep 2018 16:49:48	U	SMPL