



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB97996

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP67408
COD calibration std. 100 ppm	WP67405
COD calibration std. 10 ppm	WP67404
COD calibration std. 150 ppm	WP67407
COD calibration std. 50 ppm	WP67406
COD calibration std. 0 ppm	WP67403
COD Digestion Vials Low Level 0-150Mg/L	W2395
COD ICV-LCS std, 50ppm	WP67877
COD CCV std, 50ppm	WP67878

Temp In (C): 150	Date In: 09/18/2018	Time In: 11:00
Temp Out (C): 150	Date Out: 09/18/2018	Time Out: 13:00

Intercept: 0.2469

Slope: 0.9992

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	08/29/2018	14:40
2	CAL2	10	1	Water	10.000	08/29/2018	14:40
3	CAL3	50	1	Water	51.000	08/29/2018	14:41
4	CAL4	100	1	Water	100.000	08/29/2018	14:41
5	CAL5	150	1	Water	150.000	08/29/2018	14:42



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB97996

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	51.000	50.794	08/29/2018	14:42
2	ICB				1	Water	0.000	-0.247	08/29/2018	14:43
3	CCV1	50			1	Water	50.000	49.793	09/18/2018	14:00
4	CCB1				1	Water	0.000	-0.247	09/18/2018	14:00
5	LB97996BL				1	Water	0.000	-0.247	09/18/2018	14:01
6	LB97996BS	50			1	Water	49.000	48.792	09/18/2018	14:01
7	J4929-13				1	Water	80.000	79.817	09/18/2018	14:02
8	J4929-13DUP				1	Water	83.000	82.819	09/18/2018	14:02
9	J4929-13MS	50			1	Water	131.000	130.858	09/18/2018	14:03
10	J4929-13MSD	50			1	Water	131.000	130.858	09/18/2018	14:03
11	J4956-02				1	Water	4.000	3.756	09/18/2018	14:04
12	J4957-01				1	Water	112.000	111.843	09/18/2018	14:04
13	CCV2	50			1	Water	50.000	49.793	09/18/2018	14:05
14	CCB2				1	Water	0.000	-0.247	09/18/2018	14:05

2897996

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD091818

WorkList ID : 116866

Date : 9/18/2018 10:29:47 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/19/2018	Water	J4929-13	COD	Conc H2SO4 tc	PSEG01	I31	SP-PIT-11-BASELINE-WATEF	09/14/2018	SM5220 D
09/20/2018	Water	J4956-02	COD	Conc H2SO4 tc	PSEG01	I61	GROUNDWATER	09/17/2018	SM5220 D
09/20/2018	Water	J4957-01	COD	Conc H2SO4 tc	PSEG01	I61	N48966	09/17/2018	SM5220 D

Date/Time

Received by:

Relinquished by:

09/18/18 10:33

IB

28

Date/Time

Received by:

Relinquished by:

09/18/18 12:00

AD

28