

LB98000


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

45155



A= 0.0000e+000

B= 2.2244e-004

C= 2.6845e-002

Rho= 0.9999083

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	29	0.000	29				
0.2	0.200	0.228	0.028	906	0.0 %	906				
2.5	2.500	2.452	-0.048	10904	0.0 %	10904				
5.0	5.000	4.990	-0.010	22311	0.0 %	22311				
7.5	7.500	7.425	-0.075	33259	0.0 %	33259				
10.0	10.000	10.071	0.071	45155	0.0 %	45155				

LB98000 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 18 Sep 2018 17:00:00			
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type
0.0 - 1	50	29	-	0.00007471B	PPB	18 Sep 2018 17:01:07	S Std
0.2 - 1	50.2	906	-	0.20007471B	PPB	18 Sep 2018 17:03:24	S Std
2.5 - 1	52.5	10904	-	2.50007471B	PPB	18 Sep 2018 17:05:42	S Std
5.0 - 1	55	22311	-	5.00007471B	PPB	18 Sep 2018 17:07:59	S Std
7.5 - 1	57.5	33259	-	7.50007471B	PPB	18 Sep 2018 17:12:58	S Std
10.0 - 1	510	45155	-	10.00007471B	PPB	18 Sep 2018 17:18:03	S Std
ICV71 - 1	ICV71	16971	3.8019	-7471B	PPB	18 Sep 2018 17:20:43	U SMPL
ICB71 - 1	ICB71	-218	-0.0216	-7471B	PPB	18 Sep 2018 17:22:58	U SMPL
CCV01 - 1	CCV01	22445	5.0195	-7471B	PPB	18 Sep 2018 17:25:15	U SMPL
CCB01 - 1	CCB01	-204	-0.0185	-7471B	PPB	18 Sep 2018 17:27:31	U SMPL
CRA - 1	CRA	842	0.2141	-7471B	PPB	18 Sep 2018 17:29:48	U SMPL
HighStd - 1	HighStd	45742	10.2018	-7471B	PPB	18 Sep 2018 17:32:04	U SMPL
ChkStd - 1	ChkStd	30762	6.8696	-7471B	PPB	18 Sep 2018 17:34:20	U SMPL
PB112913BL - 1	PBS	-55	0.0146	-7471B	PPB	18 Sep 2018 17:36:39	U SMPL
PB112913BS - 1	LCSS	15466	3.4671	-7471B	PPB	18 Sep 2018 17:38:58	U SMPL
J4773-01 - 1	CL-01-091418-A	3290	0.7587	-7471B	PPB	18 Sep 2018 17:46:13	U SMPL
J4773-01DUP - 1	CL-01-091418-ADUP	2925	0.6775	-7471B	PPB	18 Sep 2018 17:48:29	U SMPL
J4773-01MS - 1	CL-01-091418-AMS	19354	4.3320	-7471B	PPB	18 Sep 2018 17:50:47	U SMPL
J4773-01MSD - 1	CL-01-091418-AMSD	18919	4.2352	-7471B	PPB	18 Sep 2018 17:53:06	U SMPL
J4773-03 - 1	CL-01-091418-B	3667	0.8425	-7471B	PPB	18 Sep 2018 17:55:22	U SMPL
CCV02 - 1	CCV02	22334	4.9949	-7471B	PPB	18 Sep 2018 17:57:39	U SMPL
CCB02 - 1	CCB02	-160	-0.0087	-7471B	PPB	18 Sep 2018 17:59:55	U SMPL
J4773-05 - 1	CL-01-091418-C	3214	0.7418	-7471B	PPB	18 Sep 2018 18:02:12	U SMPL
J4773-07 - 1	CL-01-091418-D	2633	0.6125	-7471B	PPB	18 Sep 2018 18:04:28	U SMPL
J4773-09 - 1	CL-01-091418-E	1969	0.4648	-7471B	PPB	18 Sep 2018 18:06:44	U SMPL
J4935-01 - 1	PHASE-1	826	0.2106	-7471B	PPB	18 Sep 2018 18:09:00	U SMPL
J4936-01 - 1	138kv-1	1987	0.4688	-7471B	PPB	18 Sep 2018 18:11:17	U SMPL
J4936-03 - 1	138kv-2	403	0.1165	-7471B	PPB	18 Sep 2018 18:13:35	U SMPL
J4936-05 - 1	PHASE-1	4760	1.0857	-7471B	PPB	18 Sep 2018 18:15:52	U SMPL
J4936-07 - 1	PHASE-2	1751	0.4163	-7471B	PPB	18 Sep 2018 18:18:10	U SMPL
J4936-09 - 1	PHASE-3	1149	0.2824	-7471B	PPB	18 Sep 2018 18:20:28	U SMPL
J4936-11 - 1	132-1-VOLT	846	0.2150	-7471B	PPB	18 Sep 2018 18:22:44	U SMPL
CCV03 - 1	CCV03	22566	5.0465	-7471B	PPB	18 Sep 2018 18:25:00	U SMPL
CCB03 - 1	CCB03	-138	-0.0039	-7471B	PPB	18 Sep 2018 18:27:17	U SMPL
J4936-13 - 1	26KV	653	0.1721	-7471B	PPB	18 Sep 2018 18:29:35	U SMPL
J4936-15 - 1	138-KV	8335	1.8809	-7471B	PPB	18 Sep 2018 18:31:51	U SMPL
J4936-19 - 1	138-KV-2	15905	3.5648	-7471B	PPB	18 Sep 2018 18:34:08	U SMPL
J4936-23 - 1	PHASE-1	65819	14.6677	-7471B	PPB	18 Sep 2018 18:36:25	U SMPL
J4936-27 - 1	PHASE-2	8030	1.8131	-7471B	PPB	18 Sep 2018 18:38:41	U SMPL
J4936-31 - 1	PHASE-3	31239	6.9757	-7471B	PPB	18 Sep 2018 18:41:03	U SMPL
J4936-35 - 1	132-1-VOLT	27409	6.1238	-7471B	PPB	18 Sep 2018 18:43:20	U SMPL
J4936-39 - 1	26KV	4510	1.0301	-7471B	PPB	18 Sep 2018 18:45:39	U SMPL
PB112914BL - 1	PBS	-14	0.0237	-7471B	PPB	18 Sep 2018 18:50:35	U SMPL
PB112914BS - 1	LCSS	15968	3.5788	-7471B	PPB	18 Sep 2018 18:52:51	U SMPL
CCV04 - 1	CCV04	22033	4.9279	-7471B	PPB	18 Sep 2018 18:55:08	U SMPL
CCB04 - 1	CCB04	-68	0.0117	-7471B	PPB	18 Sep 2018 18:57:24	U SMPL
J4774-03 - 1	OK-01-091418	1909	0.4515	-7471B	PPB	18 Sep 2018 18:59:41	U SMPL
J4919-01 - 1	ROLLOFF-03_COMP_FILL	24258	5.4228	-7471B	PPB	18 Sep 2018 19:01:57	U SMPL
J4929-01 - 1	SP-PIT-6	3556	0.8178	-7471B	PPB	18 Sep 2018 19:04:14	U SMPL
J4929-05 - 1	SP-PIT-12	911	0.2295	-7471B	PPB	18 Sep 2018 19:06:30	U SMPL
J4929-09 - 1	SP-PIT-11	3729	0.8563	-7471B	PPB	18 Sep 2018 19:08:47	U SMPL
J4940-01 - 1	COMP-1	270	0.0869	-7471B	PPB	18 Sep 2018 19:11:04	U SMPL
J4940-05 - 1	COMP-2	646	0.1705	-7471B	PPB	18 Sep 2018 19:13:22	U SMPL
J4941-01 - 1	RB47199RT	10357	2.3307	-7471B	PPB	18 Sep 2018 19:15:40	U SMPL
J4941-01DUP - 1	RB47199RTDUP	10604	2.3856	-7471B	PPB	18 Sep 2018 19:18:03	U SMPL
J4941-01MS - 1	RB47199RTMS	26707	5.9676	-7471B	PPB	18 Sep 2018 19:20:20	U SMPL
CCV05 - 1	CCV05	22179	4.9604	-7471B	PPB	18 Sep 2018 19:22:38	U SMPL
CCB05 - 1	CCB05	-133	-0.0027	-7471B	PPB	18 Sep 2018 19:24:55	U SMPL
J4941-01MSD - 1	RB47199RTMSD	26196	5.8539	-7471B	PPB	18 Sep 2018 19:27:13	U SMPL
J4941-04 - 1	RB47198RT	8522	1.9225	-7471B	PPB	18 Sep 2018 19:29:31	U SMPL
J4942-01 - 1	RB32727RT	1138	0.2800	-7471B	PPB	18 Sep 2018 19:31:48	U SMPL
J4943-01 - 1	RBR200051	3647	0.8381	-7471B	PPB	18 Sep 2018 19:34:06	U SMPL
J4944-01 - 1	RB48377RT	1737	0.4132	-7471B	PPB	18 Sep 2018 19:36:23	U SMPL
J4945-01 - 1	RB47195RT	709	0.1846	-7471B	PPB	18 Sep 2018 19:38:41	U SMPL
J4946-01 - 1	RB12983RT	17873	4.0025	-7471B	PPB	18 Sep 2018 19:40:58	U SMPL
J4947-01 - 1	97-20	17339	3.8838	-7471B	PPB	18 Sep 2018 19:43:16	U SMPL
J4948-01 - 1	RB44766RT	1141	0.2807	-7471B	PPB	18 Sep 2018 19:45:34	U SMPL
J4949-01 - 1	153	3561	0.8190	-7471B	PPB	18 Sep 2018 19:47:50	U SMPL
CCV06 - 1	CCV06	22410	5.0118	-7471B	PPB	18 Sep 2018 19:50:07	U SMPL
CCB06 - 1	CCB06	-204	-0.0185	-7471B	PPB	18 Sep 2018 19:52:24	U SMPL
J4950-01 - 1	RB32734RT	839	0.2135	-7471B	PPB	18 Sep 2018 19:54:43	U SMPL
J4951-01 - 1	RB48380RT	14144	3.1731	-7471B	PPB	18 Sep 2018 19:57:00	U SMPL
J4952-01 - 1	2015-161-20-RB36584RT-RB47184RT	1754	0.4170	-7471B	PPB	18 Sep 2018 19:59:16	U SMPL
J4773-01LX5 - 1		496	0.1372	-7471B	PPB	18 Sep 2018 20:01:35	U SMPL
J4773-01A - 1		20223	4.5253	-7471B	PPB	18 Sep 2018 20:03:52	U SMPL
J4941-01LX5 - 1		1667	0.3977	-7471B	PPB	18 Sep 2018 20:06:09	U SMPL
J4941-01A - 1		27996	6.2543	-7471B	PPB	18 Sep 2018 20:08:27	U SMPL
J4936-23DLX2 - 1		31584	7.0524	-7471B	PPB	18 Sep 2018 20:10:45	U SMPL
CCV07 - 1	CCV07	22368	5.0024	-7471B	PPB	18 Sep 2018 20:13:04	U SMPL
CCB07 - 1	CCB07	-43	0.0173	-7471B	PPB	18 Sep 2018 20:15:23	U SMPL