



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98001

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	09/18/2018	16:40
2	CAL2	1	Water			21.8	7.01	09/18/2018	16:42
3	CAL3	1	Water			22.0	10.04	09/18/2018	16:44
4	ICV	1	Water			21.9	7.00	09/18/2018	16:46
5	CCV1	1	Water			21.8	2.00	09/18/2018	16:48
6	J4974-04	1	Water			23.5	7.30	09/18/2018	16:50
7	J4974-05	1	Water			23.7	5.53	09/18/2018	16:52
8	J4979-04	1	Water			23.8	7.61	09/18/2018	16:54
9	J4982-01	1	Water			23.3	7.40	09/18/2018	16:56
10	J4982-01DUP	1	Water			23.4	7.41	09/18/2018	16:58
11	CCV2	1	Water			21.9	12.00	09/18/2018	17:00

WORKLIST(Hardcopy Internal Chain)

LB98001

WorkList Name : pH091818-1

WorkList ID : 116889

Date : 9/18/2018 3:53:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2018	Water	J4974-04	pH	Cool 4 deg C	PSEG01	I63	DPC-84010-OW-02	09/18/2018	9040C
09/18/2018	Water	J4974-05	pH	Cool 4 deg C	PSEG01	I63	DPC-84010-OW-04	09/18/2018	9040C
09/18/2018	Water	J4979-04	pH	Cool 4 deg C	PSEG01	I61	GROUND-WATER	09/18/2018	9040C
09/18/2018	Water	J4982-01	pH	Cool 4 deg C	PSEG01	I63	OILY-WATER	09/18/2018	9040C

Date/Time 09/18/18 4:30
 Received by: SP
 Relinquished by: CP

5:35
 Date/Time 09/18/18 5:35
 Received by: CP
 Relinquished by: SP