



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98008

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	09/18/2018	16:40
2	CAL2	1	Water			21.8	7.01	09/18/2018	16:42
3	CAL3	1	Water			22.0	10.04	09/18/2018	16:44
4	ICV	1	Water			21.9	7.00	09/18/2018	16:46
5	CCV1	1	Water			21.8	2.00	09/18/2018	17:04
6	J4974-02	1	Solid	20.02	20	24.1	5.47	09/18/2018	17:06
7	J4974-02DUP	1	Solid	20.03	20	24.2	5.48	09/18/2018	17:08
8	J4974-03	1	Solid	20.02	20	23.5	6.83	09/18/2018	17:10
9	J4974-10	1	Solid	20.03	20	23.5	7.12	09/18/2018	17:12
10	CCV2	1	Water			22.0	12.00	09/18/2018	17:30

WORKLIST(Hardcopy Internal Chain)

LB98008

WorkList Name : pH091818

WorkList ID : 116888

Date : 9/18/2018 3:53:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2018	Solid	J4974-02	pH	Cool 4 deg C	PSEG01	I63	DPC-84010-OW-01	09/18/2018	9045D
09/18/2018	Solid	J4974-03	pH	Cool 4 deg C	PSEG01	I63	DPC-84010-OW-03	09/18/2018	9045D
09/18/2018	Solid	J4974-10	pH	Cool 4 deg C	PSEG01	I63	DPC-84010-OW-COMP	09/18/2018	9045D

Date/Time 09/18/18 4:30
 Received by: SP
 Relinquished by: CP

Date/Time 09/18/18 5:35
 Received by: CP
 Relinquished by: SP