



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB98010

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	09/18/2018	16:40
2	CAL2	1	Water			21.8	7.01	09/18/2018	16:42
3	CAL3	1	Water			22.0	10.04	09/18/2018	16:44
4	ICV	1	Water			21.9	7.00	09/18/2018	16:46
5	CCV1	1	Water			21.8	2.00	09/18/2018	17:04
6	J4771-04	1	Solid	20.02	20	23.8	5.58	09/18/2018	17:14
7	J4771-04DUP	1	Solid	20.04	20	23.9	5.59	09/18/2018	17:16
8	J4973-01	1	Solid	20.02	20	23.8	7.46	09/18/2018	17:18
9	J4973-02	1	Solid	20.02	20	23.5	8.44	09/18/2018	17:20
10	J4973-03	1	Solid	20.03	20	23.4	7.70	09/18/2018	17:22
11	J4973-04	1	Solid	20.04	20	23.3	8.88	09/18/2018	17:24
12	J4975-04	1	Solid	20.02	20	24.0	11.15	09/18/2018	17:26
13	J4979-02	1	Solid	20.03	20	24.0	9.72	09/18/2018	17:28
14	CCV2	1	Water			22.0	12.00	09/18/2018	17:30

WORKLIST(Hardcopy Internal Chain)

LB98010

WorkList Name : corrosivity091818

WorkList ID : 116886

Date : 9/18/2018 3:53:18 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/18/2018	Solid	J4771-04	Corrosivity	Cool 4 deg C	PSEG05		NB-07-091818	09/18/2018	9045D
09/17/2018	Solid	J4973-01	Corrosivity	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-1	09/17/2018	9045D
09/17/2018	Solid	J4973-02	Corrosivity	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-2	09/17/2018	9045D
09/17/2018	Solid	J4973-03	Corrosivity	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-3	09/17/2018	9045D
09/17/2018	Solid	J4973-04	Corrosivity	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-4	09/17/2018	9045D
09/18/2018	Solid	J4975-04	Corrosivity	Cool 4 deg C	PSEG01	I62	VNJ-220	09/18/2018	9045D
09/18/2018	Solid	J4979-02	Corrosivity	Cool 4 deg C	PSEG01	I61	VNJ-202	09/18/2018	9045D

Date/Time 09/18/18 4:30

Received by: SP

Relinquished by: CP

Date/Time 09/18/18 5:35

Received by: CP

Relinquished by: SP