


LB98011

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.7271e-004

C= 1.4705e-002

Rho= 0.9999475

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.027	0.027	46	0.000	46				
0.2	0.200	0.196	-0.004	663	0.0 %	663				
2.5	2.500	2.432	-0.068	8864	0.0 %	8864				
5.0	5.000	5.052	0.052	18471	0.0 %	18471				
7.5	7.500	7.512	0.012	27491	0.0 %	27491				
10.0	10.000	9.981	-0.019	36546	0.0 %	36546				

LB98011 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Sep 2018 09:58:34

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0007470A	PPB	19 Sep 2018 10:04:05	S	Std
0.2 - 1	50.2	663	-	0.20007470A	PPB	19 Sep 2018 10:06:14	S	Std
2.5 - 1	52.5	8864	-	2.50007470A	PPB	19 Sep 2018 10:08:22	S	Std
5.0 - 1	55	18471	-	5.00007470A	PPB	19 Sep 2018 10:10:30	S	Std
7.5 - 1	57.5	27491	-	7.50007470A	PPB	19 Sep 2018 10:12:38	S	Std
10.0 - 1	51.0	36546	-	10.00007470A	PPB	19 Sep 2018 10:14:46	S	Std
ICV72 - 1	ICV72	14332	3.9232	-7470A	PPB	19 Sep 2018 10:17:20	U	SMPL
ICB72 - 1	ICB72	-142	-0.0240	-7470A	PPB	19 Sep 2018 10:19:26	U	SMPL
CCV08 - 1	CCV08	18461	5.0493	-7470A	PPB	19 Sep 2018 10:21:35	U	SMPL
CCB08 - 1	CCB08	-167	-0.0308	-7470A	PPB	19 Sep 2018 10:23:41	U	SMPL
CRA - 1	CRA	652	0.1925	-7470A	PPB	19 Sep 2018 10:25:50	U	SMPL
HighStd - 1	HighStd	36737	10.0334	-7470A	PPB	19 Sep 2018 10:27:56	U	SMPL
ChkStd - 1	ChkStd	25847	7.0635	-7470A	PPB	19 Sep 2018 10:30:05	U	SMPL
PB112972BL - 1	PBW	-208	-0.0420	-7470A	PPB	19 Sep 2018 10:32:15	U	SMPL
PB112972BS - 1	LCSW	13539	3.7070	-7470A	PPB	19 Sep 2018 10:34:25	U	SMPL
PB112972TB - 1	PB112943TB	-80	-0.0071	-7470A	PPB	19 Sep 2018 10:36:33	U	SMPL
J4772-02 - 1	HD-02-091718	10	0.0174	-7470A	PPB	19 Sep 2018 10:38:41	U	SMPL
J4772-02DUP - 1	HD-02-091718DUP	52	0.0289	-7470A	PPB	19 Sep 2018 10:40:50	U	SMPL
J4772-02MS - 1	HD-02-091718MS	13039	3.5706	-7470A	PPB	19 Sep 2018 10:42:58	U	SMPL
J4772-02MSD - 1	HD-02-091718MSD	15382	4.2096	-7470A	PPB	19 Sep 2018 10:45:05	U	SMPL
CCV09 - 1	CCV09	18119	4.9560	-7470A	PPB	19 Sep 2018 10:47:13	U	SMPL
CCB09 - 1	CCB09	-191	-0.0374	-7470A	PPB	19 Sep 2018 10:49:21	U	SMPL
J4960-02 - 1	S1	1569	0.4426	-7470A	PPB	19 Sep 2018 10:51:29	U	SMPL
PB112973BL - 1	PBW	73	0.0346	-7470A	PPB	19 Sep 2018 10:53:36	U	SMPL
PB112973BS - 1	LCSW	14066	3.8507	-7470A	PPB	19 Sep 2018 10:55:45	U	SMPL
PB112973TB - 1	PB112944TB	-11	0.0117	-7470A	PPB	19 Sep 2018 10:57:52	U	SMPL
J4929-14 - 1	SP-PIT-11-BASELINE-WATER	32	0.0234	-7470A	PPB	19 Sep 2018 11:00:01	U	SMPL
J4956-03 - 1	GROUNDWATER	121	0.0477	-7470A	PPB	19 Sep 2018 11:02:08	U	SMPL
J4957-02 - 1	N48966	128	0.0496	-7470A	PPB	19 Sep 2018 11:04:16	U	SMPL
J4957-02DUP - 1	N48966DUP	115	0.0461	-7470A	PPB	19 Sep 2018 11:06:25	U	SMPL
J4957-02MS - 1	N48966MS	13232	3.6233	-7470A	PPB	19 Sep 2018 11:08:33	U	SMPL
J4957-02MSD - 1	N48966MSD	13027	3.5673	-7470A	PPB	19 Sep 2018 11:10:40	U	SMPL
CCV10 - 1	CCV10	18183	4.9735	-7470A	PPB	19 Sep 2018 11:12:48	U	SMPL
CCB10 - 1	CCB10	-157	-0.0281	-7470A	PPB	19 Sep 2018 11:14:56	U	SMPL
PB112974BL - 1	PBW	-62	-0.0022	-7470A	PPB	19 Sep 2018 11:17:04	U	SMPL
PB112974BS - 1	LCSW	12532	3.4324	-7470A	PPB	19 Sep 2018 11:19:11	U	SMPL
J4913-02 - 1	BPS1-TT-MW305D-20180911	-157	-0.0281	-7470A	PPB	19 Sep 2018 11:21:18	U	SMPL
J4913-03 - 1	BPS1-TT-MW305I-20180911	-23	0.0084	-7470A	PPB	19 Sep 2018 11:23:26	U	SMPL
J4913-04 - 1	BP-TT-AOC22-MW09-20180911	48	0.0278	-7470A	PPB	19 Sep 2018 11:25:33	U	SMPL
J4913-05 - 1	BPS1-TT-MW303I2-20180912	-45	0.0024	-7470A	PPB	19 Sep 2018 11:27:40	U	SMPL
J4913-06 - 1	BPS1-TT-MW303I1-20180912	673	0.1982	-7470A	PPB	19 Sep 2018 11:29:49	U	SMPL
J4913-07 - 1	BPS1-TT-MW303D-20180912	84	0.0376	-7470A	PPB	19 Sep 2018 11:31:57	U	SMPL
J4913-08 - 1	BP-DUP01-20180912	39	0.0253	-7470A	PPB	19 Sep 2018 11:34:06	U	SMPL
J4913-09 - 1	BP-TT-AOC22-MW06-20180912	111	0.0450	-7470A	PPB	19 Sep 2018 11:36:13	U	SMPL
CCV11 - 1	CCV11	18111	4.9538	-7470A	PPB	19 Sep 2018 11:38:20	U	SMPL
CCB11 - 1	CCB11	-135	-0.0221	-7470A	PPB	19 Sep 2018 11:40:27	U	SMPL
J4913-10 - 1	BP-TT-AOC22-MW08-20180912	-14	0.0109	-7470A	PPB	19 Sep 2018 11:42:35	U	SMPL
J4913-11 - 1	BPS1-TT-MW302I1-20180912	95	0.0406	-7470A	PPB	19 Sep 2018 11:44:42	U	SMPL
J4913-12 - 1	BPS1-TT-MW302S-20180912	-45	0.0024	-7470A	PPB	19 Sep 2018 11:46:49	U	SMPL
J4913-13 - 1	BPS1-TT-MW302I2-20180912	-46	0.0022	-7470A	PPB	19 Sep 2018 11:48:57	U	SMPL
J4913-14 - 1	BP-TT-EB01-20180913	96	0.0409	-7470A	PPB	19 Sep 2018 11:51:05	U	SMPL
J4913-15 - 1	BPS1-TT-MW306D-20180913	61	0.0313	-7470A	PPB	19 Sep 2018 11:53:13	U	SMPL
J4913-16 - 1	BPS1-TT-MW302D-20180913	38	0.0251	-7470A	PPB	19 Sep 2018 11:55:21	U	SMPL
J4913-17 - 1	BPS1-TT-MW307I-20180913	18	0.0196	-7470A	PPB	19 Sep 2018 11:57:29	U	SMPL
J4913-18 - 1	BPS1-TT-MW307S-20180913	48	0.0278	-7470A	PPB	19 Sep 2018 11:59:38	U	SMPL
J4913-19 - 1	BPS1-TT-MW303S-20180913	1	0.0150	-7470A	PPB	19 Sep 2018 12:01:45	U	SMPL
CCV12 - 1	CCV12	18637	5.0973	-7470A	PPB	19 Sep 2018 12:03:52	U	SMPL
CCB12 - 1	CCB12	-132	-0.0213	-7470A	PPB	19 Sep 2018 12:05:59	U	SMPL
J4913-20 - 1	BP-TT-AOC22-MW07-20180913	-3	0.0139	-7470A	PPB	19 Sep 2018 12:08:07	U	SMPL
J4913-20DUP - 1	BP-TT-AOC22-MW07-20180913DUP	-53	0.0003	-7470A	PPB	19 Sep 2018 12:10:15	U	SMPL
J4913-20MS - 1	BP-TT-AOC22-MW07-20180913MS	13407	3.6710	-7470A	PPB	19 Sep 2018 12:12:22	U	SMPL
J4913-20MSD - 1	BP-TT-AOC22-MW07-20180913MSD	13026	3.5671	-7470A	PPB	19 Sep 2018 12:14:30	U	SMPL
PB112975BL - 1	PBW	-123	-0.0188	-7470A	PPB	19 Sep 2018 12:16:39	U	SMPL
PB112975BS - 1	LCSW	14450	3.9554	-7470A	PPB	19 Sep 2018 12:22:20	U	SMPL
J4929-13 - 1	SP-PIT-11-BASELINE-WATER	24309	6.6441	-7470A	PPB	19 Sep 2018 12:24:28	U	SMPL
J4956-02 - 1	GROUNDWATER	-16	0.0103	-7470A	PPB	19 Sep 2018 12:29:18	U	SMPL
J4957-01 - 1	N48966	3778	1.0450	-7470A	PPB	19 Sep 2018 12:31:26	U	SMPL
J4957-01DUP - 1	N48966DUP	3156	0.8754	-7470A	PPB	19 Sep 2018 12:33:33	U	SMPL
CCV13 - 1	CCV13	18755	5.1295	-7470A	PPB	19 Sep 2018 12:35:41	U	SMPL
CCB13 - 1	CCB13	-92	-0.0104	-7470A	PPB	19 Sep 2018 12:37:48	U	SMPL
J4957-01MS - 1	N48966MS	18159	4.9669	-7470A	PPB	19 Sep 2018 12:39:56	U	SMPL
J4957-01MSD - 1	N48966MSD	18224	4.9846	-7470A	PPB	19 Sep 2018 12:42:03	U	SMPL
PB112976BL - 1	PBW	-121	-0.0183	-7470A	PPB	19 Sep 2018 12:44:12	U	SMPL
PB112976BS - 1	LCSW	13225	3.6213	-7470A	PPB	19 Sep 2018 12:49:32	U	SMPL
J4926-01 - 1	BPS1-TT-MW305D-20180911	-144	-0.0246	-7470A	PPB	19 Sep 2018 12:51:39	U	SMPL
J4926-01DUP - 1	BPS1-TT-MW305D-20180911DUP	-34	0.0054	-7470A	PPB	19 Sep 2018 12:53:48	U	SMPL
J4926-01MS - 1	BPS1-TT-MW305D-20180911MS	12713	3.4817	-7470A	PPB	19 Sep 2018 12:55:56	U	SMPL
J4926-01MSD - 1	BPS1-TT-MW305D-20180911MSD	12208	3.3440	-7470A	PPB	19 Sep 2018 12:58:04	U	SMPL
J4926-02 - 1	BPS1-TT-MW305I-20180911	-112	-0.0158	-7470A	PPB	19 Sep 2018 13:00:13	U	SMPL
J4926-03 - 1	BP-TT-AOC22-MW09-20180911	6	0.0163	-7470A	PPB	19 Sep 2018 13:02:21	U	SMPL
CCV14 - 1	CCV14	18514	5.0637	-7470A	PPB	19 Sep 2018 13:04:28	U	SMPL
CCB14 - 1	CCB14	-110	-0.0153	-7470A	PPB	19 Sep 2018 13:06:36	U	SMPL
J4926-04 - 1	BPS1-TT-MW303I2-20180912	-40	0.0038	-7470A	PPB	19 Sep 2018 13:08:45	U	SMPL
J4926-05 - 1	BPS1-TT-MW303I1-20180912	62	0.0316	-7470A	PPB	19 Sep 2018 13:10:53	U	SMPL
J4926-06 - 1	BPS1-TT-MW303D-20180912	2	0.0153	-7470A	PPB	19 Sep 2018 13:13:01	U	SMPL
J4926-07 - 1	BP-DUP01-20180912	1	0.0150	-7470A	PPB	19 Sep 2018 13:15:08	U	SMPL
J4926-08 - 1	BP-TT-AOC22-MW06-20180912	38	0.0251	-7470A	PPB	19 Sep 2018 13:17:16	U	SMPL
J4926-09 - 1	BP-TT-AOC22-MW08-20180912	-5	0.0133	-7470A	PPB	19 Sep 2018 13:19:24	U	SMPL
J4926-10 - 1	BPS1-TT-MW302I1-20180912	16	0.0191	-7470A	PPB	19 Sep 2018 13:21:32	U	SMPL
J4926-11 - 1	BPS1-TT-MW302S-20180912	-4	0.0136	-7470A	PPB	19 Sep 2018 13:23:40	U	SMPL
J4926-12 - 1	BPS1-TT-MW302I2-20180912	18	0.0196	-7470A	PPB	19 Sep 2018 13:25:49	U	SMPL
J4926-13 - 1	BPS1-TT-MW306D-20180913	5	0.0161	-7470A	PPB	19 Sep 2018 13:27:56	U	SMPL
CCV15 - 1	CCV15	18188	4.9748	-7470A	PPB	19 Sep 2018 13:30:05	U	SMPL

LB98011 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Sep 2018 09:58:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB15 - 1	CCB15	-117	-0.0172		-7470A	PPB	19 Sep 2018 13:32:12	U	SMPL
J4926-14 - 1	BPS1-TT-MW302D-20180913	-24	0.0082		-7470A	PPB	19 Sep 2018 13:34:21	U	SMPL
J4926-15 - 1	BPS1-TT-MW307I-20180913	7	0.0166		-7470A	PPB	19 Sep 2018 13:36:29	U	SMPL
J4926-16 - 1	BPS1-TT-MW307S-20180913	-1	0.0144		-7470A	PPB	19 Sep 2018 13:38:37	U	SMPL
J4926-17 - 1	BPS1-TT-MW303S-20180913	69	0.0335		-7470A	PPB	19 Sep 2018 13:40:44	U	SMPL
J4926-18 - 1	BP-TT-AOC22-MW07-20180913	10	0.0174		-7470A	PPB	19 Sep 2018 13:42:52	U	SMPL
J4926-22 - 1	BP-TT-EB01-20180913	0	0.0147		-7470A	PPB	19 Sep 2018 13:45:00	U	SMPL
J4956-01 - 1	GROUNDWATER-GRAB	-15	0.0106		-7470A	PPB	19 Sep 2018 13:47:08	U	SMPL
CCV16 - 1	CCV16	18167	4.9691		-7470A	PPB	19 Sep 2018 13:49:17	U	SMPL
CCB16 - 1	CCB16	-135	-0.0221		-7470A	PPB	19 Sep 2018 13:51:25	U	SMPL
J4772-02LX5 - 1		57	0.0302		-7470A	PPB	19 Sep 2018 13:53:34	U	SMPL
J4772-02A - 1		14441	3.9530		-7470A	PPB	19 Sep 2018 13:55:42	U	SMPL
J4957-02LX5 - 1		-51	0.0008		-7470A	PPB	19 Sep 2018 13:57:50	U	SMPL
J4957-02A - 1		14156	3.8752		-7470A	PPB	19 Sep 2018 13:59:58	U	SMPL
J4913-20LX5 - 1		-156	-0.0278		-7470A	PPB	19 Sep 2018 14:02:07	U	SMPL
J4913-20A - 1		13933	3.8144		-7470A	PPB	19 Sep 2018 14:04:15	U	SMPL
J4957-01LX5 - 1		642	0.1898		-7470A	PPB	19 Sep 2018 14:06:23	U	SMPL
J4957-01A - 1		20400	5.5781		-7470A	PPB	19 Sep 2018 14:08:30	U	SMPL
J4926-01LX5 - 1		-4	0.0136		-7470A	PPB	19 Sep 2018 14:13:31	U	SMPL
J4926-01A - 1		14297	3.9137		-7470A	PPB	19 Sep 2018 14:15:39	U	SMPL
CCV17 - 1	CCV17	18157	4.9664		-7470A	PPB	19 Sep 2018 14:17:48	U	SMPL
CCB17 - 1	CCB17	-121	-0.0183		-7470A	PPB	19 Sep 2018 14:19:56	U	SMPL