

Analytical Summary Report

Analysis Method: 9038

ANALYST: AMANDEEP

Parameter: Sulfate

SUPERVISOR REVIEW BY: apatel

Run Number: LB98014

Reagent/Standard	Lot/Log #
sulfate ICV 20ppm	WP67916
Sulfate LCS std, 20ppm	WP67917
sulfate 0.0ppm	WP67906
sulfate 5ppm	WP67907
sulfate 10ppm	WP67908
sulfate 20ppm	WP67910
sulfate 30ppm	WP67912
sulfate 40ppm	WP67914
sulfate CCV 20ppm	WP67915
Sulfate std, 15ppm	WP67909
Sulfate std, 35ppm	WP67913
Sulfate std, 25ppm	WP67911
conditioning reagent	WP65047
BARIUM CHLORIDE DIHYDRATE ACS 100G	W2210

Intercept: -0.072

Slope: 2.9829

Regression: 0.999839

Seq	Lab ID	True Val (mg/l)	DF	MATRIX	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	Water	0.277	0.298	0.021	09/19/2018	11:03
2	CAL2	5	1	Water	0.275	14.200	13.925	09/19/2018	11:07
3	CAL3	10	1	Water	0.276	30.400	30.124	09/19/2018	11:14
4	CAL4	15	1	Water	0.264	45.600	45.336	09/19/2018	11:21
5	CAL5	20	1	Water	0.278	60.200	59.922	09/19/2018	11:27
6	CAL6	25	1	Water	0.261	75.400	75.139	09/19/2018	11:33
7	CAL7	30	1	Water	0.280	88.400	88.120	09/19/2018	11:38
8	CAL8	35	1	Water	0.296	104.100	103.804	09/19/2018	11:45
9	CAL9	40	1	Water	0.221	120.100	119.879	09/19/2018	11:52



Analytical Summary Report

Analysis Method: 9038
Parameter: Sulfate
Run Number: LB98014

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	MATRIX	Initial Reading	Final Reading	Difference	Result (mg/L)	AnalDate	Anal Time
1	ICV			20	1	Water	0.297	58.200	57.903	19.436	09/19/2018	11:58
2	ICB				1	Water	0.278	0.384	0.106	0.060	09/19/2018	12:05
3	CCV1			20	1	Water	0.266	60.000	59.734	20.050	09/19/2018	12:13
4	CCB1				1	Water	0.256	0.365	0.109	0.061	09/19/2018	12:19
5	LB98014BL	50	50		1	Water	0.274	0.369	0.095	0.056	09/19/2018	12:27
6	LB98014BS	50	50	20	1	Water	0.297	58.200	57.903	19.436	09/19/2018	12:35
7	J4880-06	50	50		1	Water	0.336	102.200	101.864	34.173	09/19/2018	12:42
8	CCV2			20	1	Water	0.245	59.900	59.655	20.023	09/19/2018	12:49
9	CCB2				1	Water	0.265	0.379	0.114	0.062	09/19/2018	12:55