



Analytical Summary Report

Analysis Method: 1030
Parameter: Ignitability
Run Number: LB98016

Reviewed By: sweta
Supervisor Review By: ketankumar

Seq	LabID	ClientID	DF	matrix	Result Status	Burning Rate	Anal Date	Anal Time
1	J4771-04	NB-07-091818	1	Solid	NO	0.00	09/19/2018	09:45
2	J4771-04DUP	NB-07-091818DUP	1	Solid	NO	0.00	09/19/2018	09:50
3	J4973-01	CB-0918-S-TCLP-GRI	1	Solid	NO	0.00	09/19/2018	09:55
4	J4973-02	CB-0918-S-TCLP-GRI	1	Solid	NO	0.00	09/19/2018	10:01
5	J4973-03	CB-0918-S-TCLP-GRI	1	Solid	NO	0.00	09/19/2018	10:05
6	J4973-04	CB-0918-S-TCLP-GRI	1	Solid	NO	0.00	09/19/2018	10:10
7	J4975-04	VNJ-220	1	Solid	NO	0.00	09/19/2018	10:16
8	J4979-02	VNJ-202	1	Solid	NO	0.00	09/19/2018	10:20

Burning Rate = $\frac{\text{Length (mm)}}{\text{Total Time (sec)}}$

WORKLIST(Hardcopy Internal Chain)

LB98016

WorkList Name : IGT091918

WorkList ID : 116934

Date : 9/19/2018 9:56:54 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2018	Solid	J4771-04 A	Ignitability	Cool 4 deg C	PSEG05		NB-07-091818	09/18/2018	1030
09/21/2018	Solid	J4973-01 B	Ignitability	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-1	09/17/2018	1030
09/21/2018	Solid	J4973-02 B	Ignitability	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-2	09/17/2018	1030
09/21/2018	Solid	J4973-03 B	Ignitability	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-3	09/17/2018	1030
09/21/2018	Solid	J4973-04 B	Ignitability	Cool 4 deg C	USAR03	K21	CB-0918-S-TCLP-GRID56-4	09/17/2018	1030
09/21/2018	Solid	J4975-04 A	Ignitability	Cool 4 deg C	PSEG01	I62	VNJ-220	09/18/2018	1030
09/21/2018	Solid	J4979-02 B	Ignitability	Cool 4 deg C	PSEG01	I61	VNJ-202	09/18/2018	1030

Date/Time 09/19/18 9:30
 Received by: SP
 Relinquished by: CP

Date/Time 09/19/18 11:00
 Received by: CP
 Relinquished by: SP