


7471B

LB98025
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44742



A= 0.0000e+000

B= 2.2566e-004

C= -5.5529e-003

Rho= 0.9998329

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-17	0.000	-17				
0.2	0.200	0.228	0.028	1034	0.0 %	1034				
2.5	2.500	2.526	0.026	11217	0.0 %	11217				
5.0	5.000	4.998	-0.002	22171	0.0 %	22171				
7.5	7.500	7.368	-0.132	32674	0.0 %	32674				
10.0	10.000	10.091	0.091	44742	0.0 %	44742				

LB98025 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Sep 2018 14:33:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	52	-17	-	0.00007471B	PPB		19 Sep 2018 14:41:40	S	Std
0.2 - 1	50.2	1034	-	0.20007471B	PPB		19 Sep 2018 14:43:58	S	Std
2.5 - 1	52.5	11217	-	2.50007471B	PPB		19 Sep 2018 14:46:16	S	Std
5.0 - 1	55	22171	-	5.00007471B	PPB		19 Sep 2018 14:48:33	S	Std
7.5 - 1	57.5	32674	-	7.50007471B	PPB		19 Sep 2018 14:50:50	S	Std
10.0 - 1	510	44742	-	10.00007471B	PPB		19 Sep 2018 14:55:45	S	Std
ICV73 - 1	ICV73	17707	3.9902	-7471B	PPB		19 Sep 2018 14:58:27	U	SMPL
ICB73 - 1	ICB73	68	0.0098	-7471B	PPB		19 Sep 2018 15:03:21	U	SMPL
CCV18 - 1	CCV18	22187	5.0011	-7471B	PPB		19 Sep 2018 15:05:36	U	SMPL
CCB18 - 1	CCB18	-94	-0.0268	-7471B	PPB		19 Sep 2018 15:07:52	U	SMPL
CRA - 1	CRA	984	0.2165	-7471B	PPB		19 Sep 2018 15:10:10	U	SMPL
HighStd - 1	HighStd	44909	10.1285	-7471B	PPB		19 Sep 2018 15:12:26	U	SMPL
ChkStd - 1	ChkStd	30760	6.9357	-7471B	PPB		19 Sep 2018 15:14:45	U	SMPL
PB112995BL - 1	PBS	-274	-0.0674	-7471B	PPB		19 Sep 2018 15:17:07	U	SMPL
PB112995BS - 1	LCSS	16646	3.7508	-7471B	PPB		19 Sep 2018 15:19:24	U	SMPL
J4771-03 - 1	NB-07-091818	9071	2.0414	-7471B	PPB		19 Sep 2018 15:21:41	U	SMPL
J4772-01 - 1	HD-02-091718	1720	0.3826	-7471B	PPB		19 Sep 2018 15:24:00	U	SMPL
J4772-01DUP - 1	HD-02-091718DUP	1506	0.3343	-7471B	PPB		19 Sep 2018 15:26:17	U	SMPL
J4772-01MS - 1	HD-02-091718MS	19018	4.2860	-7471B	PPB		19 Sep 2018 15:31:07	U	SMPL
J4772-01MSD - 1	HD-02-091718MSD	18820	4.2413	-7471B	PPB		19 Sep 2018 15:33:23	U	SMPL
CCV19 - 1	CCV19	22290	5.0244	-7471B	PPB		19 Sep 2018 15:35:39	U	SMPL
CCB19 - 1	CCB19	98	0.0166	-7471B	PPB		19 Sep 2018 15:40:29	U	SMPL
J4959-01 - 1	BACKFILL-SW2	2938	0.6574	-7471B	PPB		19 Sep 2018 15:42:45	U	SMPL
J4960-01 - 1	S1	5500	1.2356	-7471B	PPB		19 Sep 2018 15:45:01	U	SMPL
J4962-01 - 1	291-11984	3565	0.7989	-7471B	PPB		19 Sep 2018 15:47:17	U	SMPL
J4967-01 - 1	SUP-UST-091718-01	3384	0.7581	-7471B	PPB		19 Sep 2018 15:49:36	U	SMPL
J4967-02 - 1	SUP-UST-091718-02	10067	2.2661	-7471B	PPB		19 Sep 2018 15:51:53	U	SMPL
J4967-03 - 1	SUP-UST-091718-03	3768	0.8447	-7471B	PPB		19 Sep 2018 15:54:10	U	SMPL
J4975-01 - 1	VNJ-220	23774	5.3592	-7471B	PPB		19 Sep 2018 15:56:27	U	SMPL
J4979-01 - 1	VNJ-202	1379	0.3056	-7471B	PPB		19 Sep 2018 15:58:45	U	SMPL
J4772-01LX5 - 1		351	0.0737	-7471B	PPB		19 Sep 2018 16:01:03	U	SMPL
J4772-01A - 1		19150	4.3158	-7471B	PPB		19 Sep 2018 16:03:22	U	SMPL
CCV20 - 1	CCV20	22393	5.0476	-7471B	PPB		19 Sep 2018 16:05:38	U	SMPL
CCB20 - 1	CCB20	-36	-0.0137	-7471B	PPB		19 Sep 2018 16:07:55	U	SMPL