



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98031

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.3	4.01	09/19/2018	12:10
2	CAL2	1	Water			22.4	7.00	09/19/2018	12:12
3	CAL3	1	Water			22.5	10.04	09/19/2018	12:14
4	ICV	1	Water			22.4	7.01	09/19/2018	12:16
5	CCV1	1	Water			22.3	2.00	09/19/2018	12:18
6	J4986-01	1	Water			23.5	5.37	09/19/2018	12:20
7	J4986-01DUP	1	Water			23.6	5.38	09/19/2018	12:22
8	J4997-01	1	Water			23.3	5.79	09/19/2018	16:24
9	CCV2	1	Water			22.5	12.00	09/19/2018	16:26

WORKLIST(Hardcopy Internal Chain)

LB 98031

WorkList Name : pH091918

WorkList ID : 116939

Date : 9/19/2018 10:56:19 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/17/2018	Water	J4986-01 A	pH	Cool 4 deg C	STER04	L12	POOL-WATER-ANALYSIS	09/17/2018	SM 4500-PH B
09/18/2018	Water	J4997-01 A	pH	Cool 4 deg C	STER01	L23	POOL-WATER	09/18/2018	SM 4500-PH B

Date/Time 09/19/18 12:00
 Received by: SP
 Relinquished by: CP

Date/Time 09/19/18 12:30
 Received by: CP
 Relinquished by: SP