



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB98033

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.7	4.01	09/19/2018	16:20
2	CAL2	1	Water			22.6	7.00	09/19/2018	16:22
3	CAL3	1	Water			22.9	10.03	09/19/2018	16:24
4	ICV	1	Water			22.7	7.01	09/19/2018	16:26
5	CCV1	1	Water			22.8	2.00	09/19/2018	16:28
6	J4993-07	1	Water			24.8	6.54	09/19/2018	16:30
7	J4993-07DUP	1	Water			24.9	6.55	09/19/2018	16:32
8	CCV2	1	Water			22.6	12.00	09/19/2018	16:34

LB 94033

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity091918-1

WorkList ID : 116956

Date : 9/19/2018 3:32:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/19/2018	Water	J4993-07	Corrosivity	Cool 4 deg C	PSEG01	J41	NS-OILY-WATER	09/19/2018	9040C

Date/Time 09/19/18 4:10
Received by: SP
Relinquished by: CP

Date/Time 09/19/18 4:40
Received by: CP
Relinquished by: SP