

LB98043


7470A

INSTRUMENT ID: CVI

Linear

μ Abs.:

35553



A= 0.0000e+000

B= 2.8061e-004

C= 5.2009e-003

Rho= 0.9999368

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	-23	0.000	-23				
0.2	0.200	0.180	-0.020	624	0.0 %	624				
2.5	2.500	2.556	0.056	9090	0.0 %	9090				
5.0	5.000	4.935	-0.065	17569	0.0 %	17569				
7.5	7.500	7.548	0.048	26879	0.0 %	26879				
10.0	10.000	9.982	-0.018	35553	0.0 %	35553				

LB98043 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Sep 2018 08:50:21

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-23	-	0.00007470A	PPB		20 Sep 2018 09:01:22	S	Std
0.2 - 1	50.2	624	-	0.20007470A	PPB		20 Sep 2018 09:03:30	S	Std
2.5 - 1	52.5	9090	-	2.50007470A	PPB		20 Sep 2018 09:05:39	S	Std
5.0 - 1	55	17569	-	5.00007470A	PPB		20 Sep 2018 09:07:47	S	Std
7.5 - 1	57.5	26879	-	7.50007470A	PPB		20 Sep 2018 09:09:56	S	Std
10.0 - 1	516	35553	-	10.00007470A	PPB		20 Sep 2018 09:12:03	S	Std
ICV74 - 1	ICV74	14372	4.0382	-7470A	PPB		20 Sep 2018 09:14:42	U	SMPL
ICB74 - 1	ICB74	35	0.0150	-7470A	PPB		20 Sep 2018 09:19:15	U	SMPL
CCV21 - 1	CCV21	17591	4.9415	-7470A	PPB		20 Sep 2018 09:21:21	U	SMPL
CCB21 - 1	CCB21	-145	-0.0355	-7470A	PPB		20 Sep 2018 09:23:27	U	SMPL
CRA - 1	CRA	564	0.1635	-7470A	PPB		20 Sep 2018 09:25:35	U	SMPL
HighStd - 1	HighStd	35659	10.0116	-7470A	PPB		20 Sep 2018 09:27:42	U	SMPL
ChkStd - 1	ChkStd	24950	7.0065	-7470A	PPB		20 Sep 2018 09:29:49	U	SMPL
PB113038BL - 1	PBW	7	0.0072	-7470A	PPB		20 Sep 2018 09:34:26	U	SMPL
PB113038BS - 1	LCSW	13276	3.7306	-7470A	PPB		20 Sep 2018 09:36:32	U	SMPL
PB113038TB - 1	PB112983TB	-166	-0.0414	-7470A	PPB		20 Sep 2018 09:38:40	U	SMPL
J4771-04 - 1	NB-07-091818	-22	-0.0010	-7470A	PPB		20 Sep 2018 09:40:48	U	SMPL
J4771-04DUP - 1	NB-07-091818DUP	20	0.0108	-7470A	PPB		20 Sep 2018 09:42:56	U	SMPL
J4771-04MS - 1	NB-07-091818MS	12157	3.4166	-7470A	PPB		20 Sep 2018 09:45:05	U	SMPL
J4771-04MSD - 1	NB-07-091818MSD	12185	3.4245	-7470A	PPB		20 Sep 2018 09:47:12	U	SMPL
CCV22 - 1	CCV22	17605	4.9454	-7470A	PPB		20 Sep 2018 09:49:18	U	SMPL
CCB22 - 1	CCB22	-154	-0.0380	-7470A	PPB		20 Sep 2018 09:51:26	U	SMPL
J4973-01 - 1	CB-0918-S-TCLP-GRID56-1	19	0.0105	-7470A	PPB		20 Sep 2018 09:53:35	U	SMPL
J4973-02 - 1	CB-0918-S-TCLP-GRID56-2	33	0.0145	-7470A	PPB		20 Sep 2018 09:55:41	U	SMPL
J4973-03 - 1	CB-0918-S-TCLP-GRID56-3	149	0.0470	-7470A	PPB		20 Sep 2018 09:57:49	U	SMPL
J4973-04 - 1	CB-0918-S-TCLP-GRID56-4	-31	-0.0035	-7470A	PPB		20 Sep 2018 09:59:56	U	SMPL
J4974-10 - 1	DPC-84010-OW-COMP	42	0.0170	-7470A	PPB		20 Sep 2018 10:02:03	U	SMPL
J4975-04 - 1	VNJ-220	15	0.0094	-7470A	PPB		20 Sep 2018 10:04:11	U	SMPL
J4979-02 - 1	VNJ-202	30	0.0136	-7470A	PPB		20 Sep 2018 10:06:19	U	SMPL
PB113039BL - 1	PBW	40	0.0164	-7470A	PPB		20 Sep 2018 10:08:27	U	SMPL
PB113039BS - 1	LCSW	12253	3.4436	-7470A	PPB		20 Sep 2018 10:10:36	U	SMPL
PB113039TB - 1	PB112993TB	-95	-0.0215	-7470A	PPB		20 Sep 2018 10:12:43	U	SMPL
CCV23 - 1	CCV23	17494	4.9143	-7470A	PPB		20 Sep 2018 10:14:51	U	SMPL
CCB23 - 1	CCB23	-144	-0.0352	-7470A	PPB		20 Sep 2018 10:16:58	U	SMPL
J4979-05 - 1	GROUND-WATER	27	0.0128	-7470A	PPB		20 Sep 2018 10:19:06	U	SMPL
J4979-05DUP - 1	GROUND-WATERDUP	9	0.0077	-7470A	PPB		20 Sep 2018 10:21:13	U	SMPL
J4979-05MS - 1	GROUND-WATERMS	13128	3.6891	-7470A	PPB		20 Sep 2018 10:23:20	U	SMPL
J4979-05MSD - 1	GROUND-WATERMSD	13360	3.7542	-7470A	PPB		20 Sep 2018 10:25:27	U	SMPL
J4982-02 - 1	OILY-WATER	-208	-0.0532	-7470A	PPB		20 Sep 2018 10:27:36	U	SMPL
PB113040BL - 1	PBW	-25	-0.0018	-7470A	PPB		20 Sep 2018 10:29:45	U	SMPL
PB113040BS - 1	LCSW	13432	3.7744	-7470A	PPB		20 Sep 2018 10:31:53	U	SMPL
J4958-01 - 1	BPSI-TT-MW308D-20180914	-128	-0.0307	-7470A	PPB		20 Sep 2018 10:34:01	U	SMPL
J4958-01DUP - 1	BPSI-TT-MW308D-20180914DUP	8	0.0074	-7470A	PPB		20 Sep 2018 10:36:10	U	SMPL
J4958-01MS - 1	BPSI-TT-MW308D-20180914MS	13742	3.8614	-7470A	PPB		20 Sep 2018 10:38:17	U	SMPL
CCV24 - 1	CCV24	17615	4.9482	-7470A	PPB		20 Sep 2018 10:40:24	U	SMPL
CCB24 - 1	CCB24	76	0.0265	-7470A	PPB		20 Sep 2018 10:45:14	U	SMPL
J4958-01MSD - 1	BPSI-TT-MW308D-20180914MSD	13436	3.7755	-7470A	PPB		20 Sep 2018 10:47:21	U	SMPL
J4958-04 - 1	BPSI-HN-MW24S-20180914	-97	-0.0220	-7470A	PPB		20 Sep 2018 10:49:28	U	SMPL
J4958-05 - 1	BPSI-HN-MW24IR-20180914	9	0.0077	-7470A	PPB		20 Sep 2018 10:51:36	U	SMPL
J4958-06 - 1	BP-DUP02-20180914	-20	-0.0004	-7470A	PPB		20 Sep 2018 10:53:44	U	SMPL
J4958-07 - 1	BPSI-TT-MW306I-20180914	6	0.0069	-7470A	PPB		20 Sep 2018 10:55:51	U	SMPL
J4958-08 - 1	BPSI-TT-MW306S-20180914	-51	-0.0091	-7470A	PPB		20 Sep 2018 10:57:59	U	SMPL
J4958-12 - 1	BP-TT-AOC22-MW04-20180914	154	0.0484	-7470A	PPB		20 Sep 2018 11:00:06	U	SMPL
J4958-13 - 1	BPSI-TT-MW308I-20180915	7	0.0072	-7470A	PPB		20 Sep 2018 11:02:15	U	SMPL
J4958-14 - 1	BP-AOC22-MW03-20180915	1021	0.2917	-7470A	PPB		20 Sep 2018 11:04:24	U	SMPL
J4958-15 - 1	BPSI-TT-MW309S-20180915	-128	-0.0307	-7470A	PPB		20 Sep 2018 11:06:31	U	SMPL
CCV25 - 1	CCV25	18974	5.3296	-7470A	PPB		20 Sep 2018 11:08:38	U	SMPL
CCB25 - 1	CCB25	-133	-0.0321	-7470A	PPB		20 Sep 2018 11:10:45	U	SMPL
J4958-16 - 1	BP-TT-AOC22-MW10-20180915	24	0.0119	-7470A	PPB		20 Sep 2018 11:12:55	U	SMPL
J4958-17 - 1	BP-MH-SW4001-SOUTH-20180915	33	0.0145	-7470A	PPB		20 Sep 2018 11:15:02	U	SMPL
J4958-18 - 1	BP-DUP03-20180915	50	0.0192	-7470A	PPB		20 Sep 2018 11:17:09	U	SMPL
J4958-19 - 1	BP-TT-SW4002-20180915	-54	-0.0100	-7470A	PPB		20 Sep 2018 11:19:16	U	SMPL
J4958-22 - 1	BPSI-TT-MW309D-20180915	-47	-0.0080	-7470A	PPB		20 Sep 2018 11:21:24	U	SMPL
J4958-23 - 1	BPSI-TT-MW309I-20180915	152	0.0479	-7470A	PPB		20 Sep 2018 11:23:31	U	SMPL
J4979-04 - 1	GROUND-WATER	116	0.0378	-7470A	PPB		20 Sep 2018 11:25:39	U	SMPL
PB113041BL - 1	PBW	13	0.0088	-7470A	PPB		20 Sep 2018 11:27:48	U	SMPL
PB113041BS - 1	LCSW	13430	3.7738	-7470A	PPB		20 Sep 2018 11:29:56	U	SMPL
J4980-01 - 1	BPSI-HN-MW29IR-20180916	-98	-0.0223	-7470A	PPB		20 Sep 2018 11:32:04	U	SMPL
CCV26 - 1	CCV26	18805	5.2821	-7470A	PPB		20 Sep 2018 11:34:12	U	SMPL
CCB26 - 1	CCB26	-40	-0.0060	-7470A	PPB		20 Sep 2018 11:36:19	U	SMPL
J4980-01DUP - 1	BPSI-HN-MW29IR-20180916DUP	6	0.0069	-7470A	PPB		20 Sep 2018 11:38:29	U	SMPL
J4980-01MS - 1	BPSI-HN-MW29IR-20180916MS	12844	3.6094	-7470A	PPB		20 Sep 2018 11:40:36	U	SMPL
J4980-01MSD - 1	BPSI-HN-MW29IR-20180916MSD	12762	3.5864	-7470A	PPB		20 Sep 2018 11:42:43	U	SMPL
J4980-02 - 1	BPSI-HN-MW29D-20180916	-140	-0.0341	-7470A	PPB		20 Sep 2018 11:44:52	U	SMPL
J4980-03 - 1	BP-FW-MW03-20180916	103	0.0341	-7470A	PPB		20 Sep 2018 11:47:00	U	SMPL
J4980-04 - 1	BP-FW-MW02-20180916	-31	-0.0035	-7470A	PPB		20 Sep 2018 11:49:07	U	SMPL
J4980-05 - 1	BP-FW-MW01-20180916	28	0.0131	-7470A	PPB		20 Sep 2018 11:51:15	U	SMPL
J4980-06 - 1	BP-DUP04-20180916	-39	-0.0057	-7470A	PPB		20 Sep 2018 11:53:23	U	SMPL
J4980-07 - 1	BPSI-TT-MW308S-20180917	41	0.0167	-7470A	PPB		20 Sep 2018 11:55:32	U	SMPL
J4980-08 - 1	BPSI-TT-MW304S-20180917	-29	-0.0029	-7470A	PPB		20 Sep 2018 11:57:40	U	SMPL
CCV27 - 1	CCV27	18815	5.2849	-7470A	PPB		20 Sep 2018 11:59:48	U	SMPL
CCB27 - 1	CCB27	-111	-0.0259	-7470A	PPB		20 Sep 2018 12:01:55	U	SMPL
J4980-09 - 1	BPSI-TT-MW304D-20180917	35	0.0150	-7470A	PPB		20 Sep 2018 12:04:05	U	SMPL
J4980-11 - 1	BPSI-TT-MW304I1-20180917	3	0.0060	-7470A	PPB		20 Sep 2018 12:06:13	U	SMPL
J4980-12 - 1	BPSI-TT-MW304I2-20180917	63	0.0229	-7470A	PPB		20 Sep 2018 12:08:20	U	SMPL
J4982-01 - 1	OILY-WATER	-58	-0.0111	-7470A	PPB		20 Sep 2018 12:10:28	U	SMPL
PB113042BL - 1	PBW	55	0.0206	-7470A	PPB		20 Sep 2018 12:13:37	U	SMPL
PB113042BS - 1	LCSW	13736	3.8597	-7470A	PPB		20 Sep 2018 12:15:44	U	SMPL
J4958-24 - 1	BPSI-TT-MW308D-20180914	-111	-0.0259	-7470A	PPB		20 Sep 2018 12:17:52	U	SMPL
J4958-24DUP - 1	BPSI-TT-MW308D-20180914DUP	-6	0.0035	-7470A	PPB		20 Sep 2018 12:20:01	U	SMPL
J4958-24MS - 1	BPSI-TT-MW308D-20180914MS	13759	3.8662	-7470A	PPB		20 Sep 2018 12:22:09	U	SMPL
J4958-24MSD - 1	BPSI-TT-MW308D-20180914MSD	14657	4.1182	-7470A	PPB		20 Sep 2018 12:24:17	U	SMPL
CCV28 - 1	CCV28	17522	4.9221	-7470A	PPB		20 Sep 2018 12:28:58	U	SMPL

LB98043 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Sep 2018 08:50:21

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Cond	Method	Units	Date	Type	Type
CCB28 - 1	CCB28	-124	-0.0296	-7470A		PPB	20 Sep 2018 12:31:05	U	SMPL
J4958-25 - 1	BPSI-HN-MW24S-20180914	-9	0.0027	-7470A		PPB	20 Sep 2018 12:33:15	U	SMPL
J4958-26 - 1	BPSI-HN-MW24IR-20180914	-10	0.0024	-7470A		PPB	20 Sep 2018 12:35:23	U	SMPL
J4958-27 - 1	BP-DUP02-20180914	147	0.0465	-7470A		PPB	20 Sep 2018 12:40:50	U	SMPL
J4958-28 - 1	BPSI-TT-MW306I-20180914	105	0.0347	-7470A		PPB	20 Sep 2018 12:42:57	U	SMPL
J4958-29 - 1	BPSI-TT-MW306S-20180914	-1	0.0049	-7470A		PPB	20 Sep 2018 12:45:06	U	SMPL
J4958-30 - 1	BP-TT-AOC22-MW04-20180914	88	0.0299	-7470A		PPB	20 Sep 2018 12:47:13	U	SMPL
J4958-31 - 1	BPSI-TT-MW308I-20180915	-38	-0.0055	-7470A		PPB	20 Sep 2018 12:49:22	U	SMPL
J4958-32 - 1	BP-AOC22-MW03-20180915	293	0.0874	-7470A		PPB	20 Sep 2018 12:54:19	U	SMPL
J4958-33 - 1	BPSI-TT-MW309S-20180915	-53	-0.0097	-7470A		PPB	20 Sep 2018 12:56:27	U	SMPL
J4958-34 - 1	BP-TT-AOC22-MW10-20180915	119	0.0386	-7470A		PPB	20 Sep 2018 12:58:36	U	SMPL
CCV29 - 1	CCV29	17409	4.8904	-7470A		PPB	20 Sep 2018 13:00:44	U	SMPL
CCB29 - 1	CCB29	-75	-0.0158	-7470A		PPB	20 Sep 2018 13:02:51	U	SMPL
J4958-35 - 1	BP-MH-SW4001-SOUTH-20180915	32	0.0142	-7470A		PPB	20 Sep 2018 13:05:01	U	SMPL
J4958-36 - 1	BP-DUP03-20180915	69	0.0246	-7470A		PPB	20 Sep 2018 13:07:09	U	SMPL
J4958-37 - 1	BP-TT-SW4002-20180915	10	0.0080	-7470A		PPB	20 Sep 2018 13:09:17	U	SMPL
J4958-38 - 1	BPSI-TT-MW309D-20180915	-45	-0.0074	-7470A		PPB	20 Sep 2018 13:11:24	U	SMPL
J4958-39 - 1	BPSI-TT-MW309I-20180915	25	0.0122	-7470A		PPB	20 Sep 2018 13:13:32	U	SMPL
PB113043BL - 1	PBW	-17	0.0004	-7470A		PPB	20 Sep 2018 13:15:40	U	SMPL
PB113043BS - 1	LCSW	14063	3.9515	-7470A		PPB	20 Sep 2018 13:17:48	U	SMPL
J4980-13 - 1	BPSI-HN-MW29IR-20180916	-20	-0.0004	-7470A		PPB	20 Sep 2018 13:19:56	U	SMPL
J4980-13DUP - 1	BPSI-HN-MW29IR-20180916DUP	-8	0.0030	-7470A		PPB	20 Sep 2018 13:22:06	U	SMPL
J4980-13MS - 1	BPSI-HN-MW29IR-20180916MS	13510	3.7963	-7470A		PPB	20 Sep 2018 13:24:14	U	SMPL
CCV30 - 1	CCV30	17464	4.9058	-7470A		PPB	20 Sep 2018 13:26:22	U	SMPL
CCB30 - 1	CCB30	-166	-0.0414	-7470A		PPB	20 Sep 2018 13:28:32	U	SMPL
J4980-13MSD - 1	BPSI-HN-MW29IR-20180916MSD	12835	3.6069	-7470A		PPB	20 Sep 2018 13:30:40	U	SMPL
J4980-14 - 1	BPSI-HN-MW29D-20180916	-150	-0.0369	-7470A		PPB	20 Sep 2018 13:32:48	U	SMPL
J4980-15 - 1	BP-FW-MW03-20180916	-44	-0.0071	-7470A		PPB	20 Sep 2018 13:34:56	U	SMPL
J4980-16 - 1	BP-FW-MW02-20180916	19	0.0105	-7470A		PPB	20 Sep 2018 13:37:04	U	SMPL
J4980-17 - 1	BP-FW-MW01-20180916	-8	0.0030	-7470A		PPB	20 Sep 2018 13:39:12	U	SMPL
J4980-18 - 1	BP-DUP04-20180916	27	0.0128	-7470A		PPB	20 Sep 2018 13:41:21	U	SMPL
J4980-19 - 1	BPSI-TT-MW308S-20180917	-11	0.0021	-7470A		PPB	20 Sep 2018 13:43:28	U	SMPL
J4980-20 - 1	BPSI-TT-MW304S-20180917	-2	0.0046	-7470A		PPB	20 Sep 2018 13:45:37	U	SMPL
J4980-21 - 1	BPSI-TT-MW304D-20180917	-31	-0.0035	-7470A		PPB	20 Sep 2018 13:47:45	U	SMPL
J4980-22 - 1	BPSI-TT-MW304I1-20180917	27	0.0128	-7470A		PPB	20 Sep 2018 13:49:53	U	SMPL
CCV31 - 1	CCV31	18469	5.1879	-7470A		PPB	20 Sep 2018 13:52:02	U	SMPL
CCB31 - 1	CCB31	-138	-0.0335	-7470A		PPB	20 Sep 2018 13:54:10	U	SMPL
J4980-23 - 1	BPSI-TT-MW304I2-20180917	87	0.0296	-7470A		PPB	20 Sep 2018 13:56:21	U	SMPL
J4771-04LX5 - 1		79	0.0274	-7470A		PPB	20 Sep 2018 13:58:29	U	SMPL
J4771-04A - 1		13480	3.7879	-7470A		PPB	20 Sep 2018 14:00:37	U	SMPL
J4979-05LX5 - 1		-111	-0.0259	-7470A		PPB	20 Sep 2018 14:02:46	U	SMPL
J4979-05A - 1		13731	3.8583	-7470A		PPB	20 Sep 2018 14:04:56	U	SMPL
J4958-01LX5 - 1		-87	-0.0192	-7470A		PPB	20 Sep 2018 14:07:06	U	SMPL
J4958-01A - 1		15104	4.2436	-7470A		PPB	20 Sep 2018 14:09:16	U	SMPL
J4980-01LX5 - 1		-150	-0.0369	-7470A		PPB	20 Sep 2018 14:11:25	U	SMPL
J4980-01A - 1		11655	3.2758	-7470A		PPB	20 Sep 2018 14:13:33	U	SMPL
CCV32 - 1	CCV32	18469	5.1879	-7470A		PPB	20 Sep 2018 14:15:42	U	SMPL
CCB32 - 1	CCB32	-148	-0.0363	-7470A		PPB	20 Sep 2018 14:17:52	U	SMPL
J4958-24LX5 - 1		-38	-0.0055	-7470A		PPB	20 Sep 2018 14:20:02	U	SMPL
J4958-24A - 1		14323	4.0244	-7470A		PPB	20 Sep 2018 14:22:11	U	SMPL
J4980-13LX5 - 1		-174	-0.0436	-7470A		PPB	20 Sep 2018 14:24:19	U	SMPL
J4980-13A - 1		13943	3.9178	-7470A		PPB	20 Sep 2018 14:26:28	U	SMPL
CCV33 - 1	CCV33	17884	5.0237	-7470A		PPB	20 Sep 2018 14:28:37	U	SMPL
CCB33 - 1	CCB33	-166	-0.0414	-7470A		PPB	20 Sep 2018 14:30:45	U	SMPL