



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB98045

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.7	4.01	09/19/2018	16:20
2	CAL2	1	Water			22.6	7.00	09/19/2018	16:22
3	CAL3	1	Water			22.9	10.03	09/19/2018	16:24
4	ICV	1	Water			22.7	7.01	09/19/2018	16:26
5	CCV1	1	Water			22.7	2.00	09/19/2018	17:00
6	J4987-03	1	Solid	20.02	20	24.0	5.41	09/19/2018	17:02
7	J4987-03DUP	1	Solid	20.03	20	24.1	5.42	09/19/2018	17:04
8	J4993-05	1	Solid	20.02	20	23.7	6.61	09/19/2018	17:06
9	J5001-01	1	Solid	20.02	20	23.5	5.65	09/19/2018	17:08
10	J5001-04	1	Solid	20.04	20	23.4	6.75	09/19/2018	17:10
11	CCV2	1	Water			22.8	12.00	09/19/2018	17:12

WORKLIST(Hardcopy Internal Chain)

LB98045

WorkList Name : corrosivity091918

WorkList ID : 116955

Date : 9/19/2018 3:32:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/19/2018	Solid	J4987-03	Corrosivity	Cool 4 deg C	PSEG01	I62	RBR200036	09/19/2018	9045D
09/19/2018	Solid	J4993-05	Corrosivity	Cool 4 deg C	PSEG01	J41	NS-SLUDGE	09/19/2018	9045D
09/19/2018	Solid	J5001-01	Corrosivity	Cool 4 deg C	PSEG01	I31	BS-OILY-DEBRI	09/19/2018	9045D
09/19/2018	Solid	J5001-04	Corrosivity	Cool 4 deg C	PSEG01	I31	BS-OILY-STONE	09/19/2018	9045D

Date/Time 09/19/18 4:00

Received by: SP

Relinquished by: CP

Date/Time 09/19/18 5:15

Received by: CP

Relinquished by: SP