



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98050

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 051106

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	09/20/2018	09:20
2	J4993-07		1	9	>100.0	>212.0	>212.0	09/20/2018	09:50
3	J4993-07DUP		1	9	>100.0	>212.0	>212.0	09/20/2018	10:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB98050

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92018

WorkList ID : 116989

Date : 9/20/2018 8:54:21 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/24/2018	Water	J4993-07	Flash Point	Cool 4 deg C	PSEG01	J41	NS-OILY-WATER	09/19/2018	1010A

Date/Time 09/20/18 8:55 AM
Received by: KS
Relinquished by: CP

Date/Time 09/20/18 10:30 AM
Received by: CP
Relinquished by: KS