



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98054

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP67975
Turbidity std, 40NTU	WP67968
turbidity std 5 NTU	WP67971
turbidity std 20 NTU	WP67969
Turbidity std 1 NTU	WP67972
turbidity std 0 NTU	WP67974
turbidity std 10 NTU	WP67970
10 NTU Standard 500 ml	W2297

Intercept: 0.1942

Slope: 0.9059

Regression: 0.9996

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.244	09/20/2018	12:17
2	CAL2	1	1	1.110	09/20/2018	12:20
3	CAL3	5	1	4.800	09/20/2018	12:23
4	CAL4	10	1	9.560	09/20/2018	12:27
5	CAL5	20	1	17.600	09/20/2018	12:30
6	CAL6	40	1	36.700	09/20/2018	12:34



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98054

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.810	10.615	09/20/2018	12:38
2	ICB		1	0.260	0.073	09/20/2018	12:42
3	CCV1	10	1	9.440	10.206	09/20/2018	12:47
4	CCB1		1	0.250	0.062	09/20/2018	12:51
5	LB98054BL		1	0.241	0.052	09/20/2018	12:54
6	J4880-37		1	2.740	2.810	09/20/2018	12:59
7	J5007-01		1	15.300	16.675	09/20/2018	13:03
8	J5007-01DUP		1	15.700	17.116	09/20/2018	13:07
9	CCV2	10	1	9.510	10.283	09/20/2018	13:11
10	CCB2		1	0.257	0.069	09/20/2018	13:16

LB98054

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity092018

WorkList ID : 116991

Date : 9/20/2018 9:39:57 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2018	Water	J5007-01	Turbidity	Cool 4 deg C	LIR001	L23	UDS-CH-4-01	09/19/2018	SM2130 B

Date/Time 9-20-18 11:00
 Received by: AL
 Relinquished by: OP

Date/Time 9-20-18 13:20
 Received by: CP
 Relinquished by: AL