



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98055

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP67975
Turbidity std, 40NTU	WP67968
turbidity std 5 NTU	WP67971
turbidity std 20 NTU	WP67969
Turbidity std 1 NTU	WP67972
turbidity std 0 NTU	WP67974
turbidity std 10 NTU	WP67970
10 NTU Standard 500 ml	W2297

Intercept: 0.1942

Slope: 0.9059

Regression: 0.9996

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.244	09/20/2018	12:17
2	CAL2	1	1	1.110	09/20/2018	12:20
3	CAL3	5	1	4.800	09/20/2018	12:23
4	CAL4	10	1	9.560	09/20/2018	12:27
5	CAL5	20	1	17.600	09/20/2018	12:30
6	CAL6	40	1	36.700	09/20/2018	12:34



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB98055

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.810	10.615	09/20/2018	12:38
2	ICB		1	0.260	0.073	09/20/2018	12:42
3	CCV1	10	1	9.510	10.283	09/20/2018	13:11
4	CCB1		1	0.257	0.069	09/20/2018	13:16
5	LB98055BL		1	0.249	0.060	09/20/2018	13:25
6	J4880-38		1	2.750	2.821	09/20/2018	13:45
7	J4880-38DUP		1	2.730	2.799	09/20/2018	13:48
8	CCV2	10	1	9.460	10.228	09/20/2018	13:51
9	CCB2		1	0.261	0.074	09/20/2018	13:54