



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB98060

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.2	4.01	09/20/2018	11:45
2	CAL2	1	Water			22.3	7.00	09/20/2018	11:47
3	CAL3	1	Water			22.5	10.04	09/20/2018	11:49
4	ICV	1	Water			22.5	7.01	09/20/2018	11:51
5	CCV1	1	Water			22.4	2.00	09/20/2018	11:53
6	J5008-01	1	Water			22.2	5.07	09/20/2018	11:55
7	J5008-01DUP	1	Water			22.3	5.08	09/20/2018	11:57
8	J5008-02	1	Water			22.1	4.78	09/20/2018	11:59
9	J5008-03	1	Water			21.7	5.28	09/20/2018	12:02
10	CCV2	1	Water			22.5	12.00	09/20/2018	12:03

LB 98060

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH092018

WorkList ID : 117000

Date : 9/20/2018 11:15:06 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/19/2018	Water	J5008-01 A	pH	Cool 4 deg C	STER07	L21	CELL-A-POOL-WATER	09/19/2018	SM 4500-PH B
09/19/2018	Water	J5008-02 A	pH	Cool 4 deg C	STER07	L21	CELL-B-C-POOL-WATER	09/19/2018	SM 4500-PH B
09/19/2018	Water	J5008-03 A	pH	Cool 4 deg C	STER07	L21	CELL-D-POOL-WATER	09/19/2018	SM 4500-PH B

Date/Time 09/20/18 11:30
Received by: SP
Relinquished by: CP

Date/Time 09/20/18 12:30
Received by: CP
Relinquished by: SP