

7471B


LB98070

INSTRUMENT ID: CV2

Linear

 μ Abs.:

43232



A= 0.0000e+000

B= 2.2973e-004

C= 4.5062e-002

Rho= 0.9996846

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	54	0.000	54				
0.2	0.200	0.214	0.014	734	0.0 %	734				
2.5	2.500	2.474	-0.026	10575	0.0 %	10575				
5.0	5.000	4.836	-0.164	20855	0.0 %	20855				
7.5	7.500	7.641	0.141	33066	0.0 %	33066				
10.0	10.000	9.977	-0.023	43232	0.0 %	43232				

LB98070 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Sep 2018 17:01:09

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	54	-	0.0000	7471B	PPB	20 Sep 2018 17:02:21	S	Std
0.2 - 1	50.2	734	-	0.2000	7471B	PPB	20 Sep 2018 17:04:39	S	Std
2.5 - 1	52.5	10575	-	2.5000	7471B	PPB	20 Sep 2018 17:06:57	S	Std
5.0 - 1	53	20855	-	5.0000	7471B	PPB	20 Sep 2018 17:09:14	S	Std
7.5 - 1	57.5	33066	-	7.5000	7471B	PPB	20 Sep 2018 17:11:31	S	Std
10.0 - 1	510	43232	-	10.0000	7471B	PPB	20 Sep 2018 17:16:25	S	Std
ICV75 - 1	ICV75	17772	4.1279	-	7471B	PPB	20 Sep 2018 17:19:11	U	SMPL
ICB75 - 1	ICB75	-154	0.0097	-	7471B	PPB	20 Sep 2018 17:21:27	U	SMPL
CCV36 - 1	CCV36	21142	4.9021	-	7471B	PPB	20 Sep 2018 17:23:44	U	SMPL
CCB36 - 1	CCB36	-124	0.0166	-	7471B	PPB	20 Sep 2018 17:25:59	U	SMPL
CRA - 1	CRA	690	0.2036	-	7471B	PPB	20 Sep 2018 17:28:17	U	SMPL
HighStd - 1	HighStd	43402	10.0159	-	7471B	PPB	20 Sep 2018 17:30:33	U	SMPL
ChkStd - 1	ChkStd	30359	7.0195	-	7471B	PPB	20 Sep 2018 17:32:49	U	SMPL
PB113053BL - 1	PBS	-230	-0.0078	-	7471B	PPB	20 Sep 2018 17:35:08	U	SMPL
PB113053BS - 1	LCSS	15816	3.6785	-	7471B	PPB	20 Sep 2018 17:37:27	U	SMPL
J4778-01 - 1	SA-01-091918	2481	0.6150	-	7471B	PPB	20 Sep 2018 17:44:40	U	SMPL
J4778-01DUP - 1	SA-01-091918DUP	2290	0.5711	-	7471B	PPB	20 Sep 2018 17:46:57	U	SMPL
J4778-01MS - 1	SA-01-091918MS	17782	4.1302	-	7471B	PPB	20 Sep 2018 17:49:14	U	SMPL
J4778-01MSD - 1	SA-01-091918MSD	18598	4.3176	-	7471B	PPB	20 Sep 2018 17:51:32	U	SMPL
J4985-01 - 1	EDISON-1	433	0.1445	-	7471B	PPB	20 Sep 2018 17:53:49	U	SMPL
CCV37 - 1	CCV37	20977	4.8642	-	7471B	PPB	20 Sep 2018 17:56:05	U	SMPL
CCB37 - 1	CCB37	-141	0.0127	-	7471B	PPB	20 Sep 2018 17:58:21	U	SMPL
J4985-03 - 1	EDISON-2	819	0.2332	-	7471B	PPB	20 Sep 2018 18:00:37	U	SMPL
J4987-01 - 1	RBR200036	1338	0.3524	-	7471B	PPB	20 Sep 2018 18:02:53	U	SMPL
J4988-01 - 1	METUCHEN-1.5	927	0.2580	-	7471B	PPB	20 Sep 2018 18:05:09	U	SMPL
J4988-03 - 1	METUCHEN-2	597	0.1822	-	7471B	PPB	20 Sep 2018 18:07:25	U	SMPL
J4988-05 - 1	METUCHEN-3	546	0.1705	-	7471B	PPB	20 Sep 2018 18:09:42	U	SMPL
J4988-07 - 1	METUCHEN-4	882	0.2477	-	7471B	PPB	20 Sep 2018 18:11:59	U	SMPL
J4993-04 - 1	NS-SLUDGE	2980	0.7297	-	7471B	PPB	20 Sep 2018 18:14:16	U	SMPL
J4998-01 - 1	YG3-SLOAN-PARK-BRIDGE	70247	16.1831	-	7471B	PPB	20 Sep 2018 18:16:34	U	SMPL
J5001-02 - 1	BS-OILY-STONE	353	0.1262	-	7471B	PPB	20 Sep 2018 18:18:51	U	SMPL
J5002-01 - 1	BACKFILL-12TH-AVE	5209	1.2417	-	7471B	PPB	20 Sep 2018 18:21:10	U	SMPL
CCV38 - 1	CCV38	21048	4.8805	-	7471B	PPB	20 Sep 2018 18:23:27	U	SMPL
CCB38 - 1	CCB38	-161	0.0081	-	7471B	PPB	20 Sep 2018 18:25:43	U	SMPL
J5004-01 - 1	FK-GF-02-026-A	1601	0.4129	-	7471B	PPB	20 Sep 2018 18:27:59	U	SMPL
J5004-03 - 1	FK-GF-02-026-B	1609	0.4147	-	7471B	PPB	20 Sep 2018 18:30:15	U	SMPL
J4778-01LX5 - 1		439	0.1459	-	7471B	PPB	20 Sep 2018 18:32:31	U	SMPL
J4778-01A - 1		19550	4.5363	-	7471B	PPB	20 Sep 2018 18:34:47	U	SMPL
J4998-01DLX5 - 1		14461	3.3672	-	7471B	PPB	20 Sep 2018 18:37:04	U	SMPL
CCV39 - 1	CCV39	21041	4.8789	-	7471B	PPB	20 Sep 2018 18:39:21	U	SMPL
CCB39 - 1	CCB39	-175	0.0049	-	7471B	PPB	20 Sep 2018 18:41:38	U	SMPL