



Analytical Summary Report

Analysis Method: 9045D
Parameter: pH
Run Number: LB98079
BalanceID: WC SC-4

Analyst By : Sweta
Supervisor Review By : apatel
Slope : 99.1
pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].
True Value of CCV1 = 2.00 Control Limits[+/- 0.1].
True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.5	4.01	09/21/2018	09:40
2	CAL2	1	Water			22.2	7.00	09/21/2018	09:42
3	CAL3	1	Water			22.5	10.03	09/21/2018	09:44
4	ICV	1	Water			22.3	7.01	09/21/2018	09:46
5	CCV1	1	Water			22.4	2.00	09/21/2018	09:48
6	J5025-01	1	Solid	20.02	20	23.3	5.28	09/21/2018	09:50
7	J5025-01DUP	1	Solid	20.03	20	23.4	5.29	09/21/2018	09:52
8	CCV2	1	Water			22.3	12.00	09/21/2018	09:54

WORKLIST(Hardcopy Internal Chain)

LB98079

WorkList Name : pH092118

WorkList ID : 117043

Date : 9/21/2018 9:11:50 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/20/2018	Solid	J5025-01	pH	Cool 4 deg C	TULL02	M11	FK-PS-01-041	09/20/2018	9045D

Date/Time 09/21/18 9:15
 Received by: SP
 Relinquished by: CP

Date/Time 09/21/18 10:00
 Received by: CP
 Relinquished by: SP