

245.1


LB98083

INSTRUMENT ID: CY1

Linear

 μ Abs.:

32396



A= 0.0000e+000

B= 3.0869e-004

C= -4.8587e-002

Rho= 0.9999533

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.007	0.007	181	0.000	181				
0.05	0.050	0.048	-0.002	312	0.0 %	312				
0.2	0.200	0.148	-0.052	636	0.0 %	636				
2.5	2.500	2.519	0.019	8317	0.0 %	8317				
5.0	5.000	5.055	0.055	16534	0.0 %	16534				
7.5	7.500	7.522	0.022	24524	0.0 %	24524				
10.0	10.000	9.952	-0.048	32396	0.0 %	32396				

LB98083 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 21 Sep 2018 12:17:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	181	-	0.0000	245.1	PPB	21 Sep 2018 12:24:22	S	Std
0.05 - 1	50.05	312	-	0.0500	245.1	PPB	21 Sep 2018 12:26:40	S	Std
0.2 - 1	50.2	636	-	0.2000	245.1	PPB	21 Sep 2018 12:28:57	S	Std
2.5 - 1	52.5	8317	-	2.5000	245.1	PPB	21 Sep 2018 12:31:15	S	Std
5.0 - 1	55	16534	-	5.0000	245.1	PPB	21 Sep 2018 12:33:32	S	Std
7.5 - 1	57.5	24524	-	7.5000	245.1	PPB	21 Sep 2018 12:35:48	S	Std
10.0 - 1	510	32396	-	10.0000	245.1	PPB	21 Sep 2018 12:41:22	S	Std
ICV77 - 1	ICV77	13462	4.1070	-245.1		PPB	21 Sep 2018 12:44:27	U	SMPL
ICB77 - 1	ICB77	10	-0.0455	-245.1		PPB	21 Sep 2018 12:46:43	U	SMPL
CCV45 - 1	CCV45	16845	5.1512	-245.1		PPB	21 Sep 2018 12:54:16	U	SMPL
CCB45 - 1	CCB45	-31	-0.0582	-245.1		PPB	21 Sep 2018 12:56:32	U	SMPL
CRA - 1	CRA	351	0.0598	-245.1		PPB	21 Sep 2018 13:01:29	U	SMPL
HighStd - 1	HighStd	32185	9.8865	-245.1		PPB	21 Sep 2018 13:03:44	U	SMPL
ChkStd - 1	ChkStd	22618	6.9333	-245.1		PPB	21 Sep 2018 13:06:00	U	SMPL
PB113072BL - 1	PBW	141	-0.0051	-245.1		PPB	21 Sep 2018 13:10:56	U	SMPL
PB113072BS - 1	LCSW	14388	4.3928	-245.1		PPB	21 Sep 2018 13:13:13	U	SMPL
J4930-01 - 1	091318-JETT-E-D-B	-138	-0.0912	-245.1		PPB	21 Sep 2018 13:15:30	U	SMPL
J4930-02 - 1	091318-JETT-E-D-S	27	-0.0403	-245.1		PPB	21 Sep 2018 13:17:49	U	SMPL
J4930-03 - 1	091318-JETT-E-U-B	34	-0.0381	-245.1		PPB	21 Sep 2018 13:20:06	U	SMPL
J4930-04 - 1	091318-JETT-E-U-S	43	-0.0353	-245.1		PPB	21 Sep 2018 13:22:24	U	SMPL
J4930-05 - 1	091318-DBRI-E-D-B	32	-0.0387	-245.1		PPB	21 Sep 2018 13:24:40	U	SMPL
CCV46 - 1	CCV46	16417	5.0191	-245.1		PPB	21 Sep 2018 13:26:56	U	SMPL
CCB46 - 1	CCB46	-66	-0.0690	-245.1		PPB	21 Sep 2018 13:29:12	U	SMPL
J4930-06 - 1	091318-DBRI-E-D-M	0	-0.0486	-245.1		PPB	21 Sep 2018 13:31:29	U	SMPL
J4930-07 - 1	091318-DBRI-E-D-S	1	-0.0483	-245.1		PPB	21 Sep 2018 13:33:45	U	SMPL
J4930-08 - 1	091318-DBRI-E-U-B	21	-0.0421	-245.1		PPB	21 Sep 2018 13:36:01	U	SMPL
J4930-09 - 1	091318-DBRI-E-U-M	35	-0.0378	-245.1		PPB	21 Sep 2018 13:38:18	U	SMPL
J4930-10 - 1	091318-DBRI-E-U-S	37	-0.0372	-245.1		PPB	21 Sep 2018 13:40:34	U	SMPL
J4930-11 - 1	091318-TIPC-F-D-B	-19	-0.0545	-245.1		PPB	21 Sep 2018 13:42:51	U	SMPL
J4930-12 - 1	091318-TIPC-F-D-S	-3	-0.0495	-245.1		PPB	21 Sep 2018 13:45:09	U	SMPL
J4930-13 - 1	091318-TIPC-F-U-B	20	-0.0424	-245.1		PPB	21 Sep 2018 13:47:26	U	SMPL
J4930-14 - 1	091318-TIPC-F-U-S	-12	-0.0523	-245.1		PPB	21 Sep 2018 13:49:44	U	SMPL
PB113073BL - 1	PBW	-3	-0.0495	-245.1		PPB	21 Sep 2018 13:52:00	U	SMPL
CCV47 - 1	CCV47	16601	5.0759	-245.1		PPB	21 Sep 2018 13:54:16	U	SMPL
CCB47 - 1	CCB47	-38	-0.0603	-245.1		PPB	21 Sep 2018 13:56:32	U	SMPL
PB113073BS - 1	LCSW	13550	4.1341	-245.1		PPB	21 Sep 2018 13:58:51	U	SMPL
J4931-01 - 1	091418-SIDI-E-D-B	-91	-0.0767	-245.1		PPB	21 Sep 2018 14:01:07	U	SMPL
J4931-02 - 1	091418-SIDI-E-D-S	-14	-0.0529	-245.1		PPB	21 Sep 2018 14:03:24	U	SMPL
J4931-03 - 1	091418-SIDI-E-U-B	50	-0.0332	-245.1		PPB	21 Sep 2018 14:05:41	U	SMPL
J4931-04 - 1	091418-SIDI-E-U-S	17	-0.0433	-245.1		PPB	21 Sep 2018 14:07:57	U	SMPL
J4931-05 - 1	091418-DBRI-E-D-B	121	-0.0112	-245.1		PPB	21 Sep 2018 14:10:15	U	SMPL
J4931-06 - 1	091418-DBRI-E-D-S	54	-0.0319	-245.1		PPB	21 Sep 2018 14:12:32	U	SMPL
J4931-07 - 1	091418-DBRI-E-U-B	20	-0.0424	-245.1		PPB	21 Sep 2018 14:14:49	U	SMPL
J4931-08 - 1	091418-DBRI-E-U-S	-7	-0.0507	-245.1		PPB	21 Sep 2018 14:17:07	U	SMPL
J4931-09 - 1	091418-JETT-E-D-B	-9	-0.0514	-245.1		PPB	21 Sep 2018 14:19:24	U	SMPL
CCV48 - 1	CCV48	16183	4.9469	-245.1		PPB	21 Sep 2018 14:21:40	U	SMPL
CCB48 - 1	CCB48	-121	-0.0859	-245.1		PPB	21 Sep 2018 14:23:57	U	SMPL
J4931-10 - 1	091418-JETT-E-D-S	-54	-0.0653	-245.1		PPB	21 Sep 2018 14:26:15	U	SMPL
J4931-11 - 1	091418-JETT-E-U-B	19	-0.0427	-245.1		PPB	21 Sep 2018 14:28:32	U	SMPL
J4931-12 - 1	091418-JETT-E-U-S	41	-0.0359	-245.1		PPB	21 Sep 2018 14:30:48	U	SMPL
PB113074BL - 1	PBW	34	-0.0381	-245.1		PPB	21 Sep 2018 14:33:05	U	SMPL
PB113074BS - 1	LCSW	12603	3.8418	-245.1		PPB	21 Sep 2018 14:35:22	U	SMPL
J4984-01 - 1	091718-DBRI-E-D-B	-125	-0.0872	-245.1		PPB	21 Sep 2018 14:37:39	U	SMPL
J4984-02 - 1	091718-DBRI-E-D-S	27	-0.0403	-245.1		PPB	21 Sep 2018 14:39:56	U	SMPL
J4984-03 - 1	091718-DBRI-E-U-B	28	-0.0399	-245.1		PPB	21 Sep 2018 14:42:14	U	SMPL
J4984-04 - 1	091718-DBRI-E-U-S	35	-0.0378	-245.1		PPB	21 Sep 2018 14:44:31	U	SMPL
J4984-05 - 1	091718-SIDI-E-D-B	42	-0.0356	-245.1		PPB	21 Sep 2018 14:46:48	U	SMPL
CCV49 - 1	CCV49	16162	4.9404	-245.1		PPB	21 Sep 2018 14:49:04	U	SMPL
CCB49 - 1	CCB49	-185	-0.1057	-245.1		PPB	21 Sep 2018 14:51:21	U	SMPL
J4984-06 - 1	091718-SIDI-E-D-S	-43	-0.0619	-245.1		PPB	21 Sep 2018 14:53:39	U	SMPL
J4984-07 - 1	091718-SIDI-E-U-B	42	-0.0356	-245.1		PPB	21 Sep 2018 14:55:55	U	SMPL
J4984-08 - 1	091718-SIDI-E-U-S	12	-0.0449	-245.1		PPB	21 Sep 2018 14:58:12	U	SMPL
J4984-09 - 1	091718-TIPC-E-D-M	39	-0.0365	-245.1		PPB	21 Sep 2018 15:00:28	U	SMPL
J4984-10 - 1	091718-TIPC-E-U-M	50	-0.0332	-245.1		PPB	21 Sep 2018 15:02:44	U	SMPL
J4984-10DUP - 1	091718-TIPC-E-U-MDUP	-11	-0.0520	-245.1		PPB	21 Sep 2018 15:05:01	U	SMPL
J4984-10MS - 1	091718-TIPC-E-U-MMS	13959	4.2604	-245.1		PPB	21 Sep 2018 15:07:18	U	SMPL
J4984-10MSD - 1	091718-TIPC-E-U-MMSD	13250	4.0415	-245.1		PPB	21 Sep 2018 15:09:36	U	SMPL
J4984-10LX5 - 1		-28	-0.0572	-245.1		PPB	21 Sep 2018 15:11:55	U	SMPL
J4984-10A - 1		15811	4.8321	-245.1		PPB	21 Sep 2018 15:14:11	U	SMPL
CCV50 - 1	CCV50	16718	5.1120	-245.1		PPB	21 Sep 2018 15:16:27	U	SMPL
CCB50 - 1	CCB50	-65	-0.0687	-245.1		PPB	21 Sep 2018 15:18:45	U	SMPL