



PERCENT SOLID

Supervisor: sweta
Analyst: JIGNESH
Date: 9/24/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 09/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/22/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB98094

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J4770-01	JC-02-092118-A	1	1.14	9.74	8.69	87.8
J4770-03	JC-02-092118-B	2	1.13	9.82	8.96	90.1
J4770-05	JC-02-092118-C	3	1.18	9.85	8.97	89.9
J4770-07	JC-02-092118-D	4	1.13	9.97	9.4	93.6
J4770-09	JC-02-092118-E	5	1.18	9.98	9.27	91.9
J4770-11	JC-02-092118-F	6	1.13	9.54	8.67	89.7
J4770-13	JC-02-092118-G	7	1.13	9.81	8.8	88.4
J4770-15	JC-02-092118-H	8	1.1	9.97	9.11	90.3
J4770-17	JC-02-092118-I	9	1.17	9.92	8.79	87.1
J4770-19	JC-02-092118-J	10	1.16	9.85	8.9	89.1
J5042-01	2018-NYSEG-CLARK-PH4-ARI	11	1.17	9.88	7.55	73.2
J5042-02	2018-NYSEG-CLARK-PH4-ARI	12	1.11	9.98	8.5	83.3
J5042-03	2018-NYSEG-CLARK-PH4-ARI	13	1.16	9.57	8.22	83.9
J5042-04	2018-NYSEG-CLARK-PH4-ARI	14	1.16	9.83	8.05	79.5
J5044-01	GRID01-COMP-092018	15	1.13	9.65	8.01	80.8
J5044-02	GRID02-COMP-092018	16	1.14	9.9	8.42	83.1
J5044-03	GRID03-COMP-092018	17	1.13	9.91	8.06	78.9
J5044-04	GRID04-COMP-092018	18	1.16	9.82	7.77	76.3
J5044-05	GRID05-COMP-092018	19	1.18	9.97	8.35	81.6
J5044-06	GRID06-COMP-092018	20	1.17	9.6	7.97	80.7
J5044-07	GRID07-COMP-092018	21	1.17	9.87	7.88	77.1
J5044-08	GRID08-COMP-092018	22	1.13	9.95	7.96	77.4
J5044-09	GRID09-COMP-092018	23	1.16	9.92	8.38	82.4
J5046-01	S1	24	1.13	9.97	9.12	90.4
J5046-02	S-1A	25	1.17	9.88	9.1	91.0
J5050-01	A370591	26	1.00	2.00	2.00	100.0
J5050-02	B370592	27	1.00	2.00	2.00	100.0
J5057-01	HC-SOIL	28	1.16	9.95	9.2	91.5
J5057-02	HC-SOIL-DRO-GRO-1	29	1.16	9.95	9.2	91.5
J5057-03	HC-SOIL-DRO-GRO-2	30	1.16	9.95	9.2	91.5
J5057-04	HC-SOIL-DRO-GRO-3	31	1.16	9.95	9.2	91.5
J5057-05	HC-SOIL-DRO-GRO-4	32	1.16	9.95	9.2	91.5
J5057-06	HC-SOIL-DRO-GRO-5	33	1.16	9.95	9.2	91.5
J5057-07	HC-SOIL-DRO-GRO-6	34	1.16	9.95	9.2	91.5
J5057-08	HC-SOIL-DRO-GRO-7	35	1.16	9.95	9.2	91.5
J5057-09	HC-SOIL-DRO-GRO-8	36	1.16	9.95	9.2	91.5
J5058-01	DRUM-1	37	1.00	2.00	2.00	100.0
J5059-01	PHASE-2	38	1.17	9.98	8.63	84.7
J5059-02	PHASE-2-GRAB-1	39	1.13	9.72	8.62	87.2



PERCENT SOLID

Supervisor: sweta
Analyst: JIGNESH
Date: 9/24/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:30
In Date: 09/21/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 09/22/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB98094

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J5059-03	PHASE-2-GRAB-2	40	1.13	9.72	8.62	87.2
J5060-05	SVOC-GPC-BLANK	41	1.00	2.00	2.00	100.0
J5060-06	PEST-GPC-BLANK	42	1.00	2.00	2.00	100.0
J5060-07	PEST-GPC-BLANK-SPIKE	43	1.00	2.00	2.00	100.0
J5060-08	PCB-GPC-BLANK	44	1.00	2.00	2.00	100.0
J5060-09	PCB-GPC-BLANK-SPIKE	45	1.00	2.00	2.00	100.0
J5060-10	SVOC-GPC2-BLANK	46	1.00	2.00	2.00	100.0
J5060-11	PEST-GPC2-BLANK	47	1.00	2.00	2.00	100.0
J5060-12	PEST-GPC2-BLANK-SPIKE	48	1.00	2.00	2.00	100.0
J5060-13	PCB-GPC2-BLANK	49	1.00	2.00	2.00	100.0
J5060-14	PCB-GPC2-BLANK-SPIKE	50	1.00	2.00	2.00	100.0
J5061-01	230-KV-A	51	1.00	2.00	2.00	100.0
J5061-02	230-KV-A-GRAB-1	52	1.00	2.00	2.00	100.0
J5061-03	230-KV-A-GRAB-2	53	1.00	2.00	2.00	100.0
J5061-06	230-KV-B	54	1.00	2.00	2.00	100.0
J5061-07	230-KV-B-GRAB-1	55	1.00	2.00	2.00	100.0
J5061-08	230-KV-B-GRAB-2	56	1.00	2.00	2.00	100.0
J5061-11	230-KV-C	57	1.00	2.00	2.00	100.0
J5061-12	230-KV-C-GRAB-1	58	1.00	2.00	2.00	100.0
J5061-13	230-KV-C-GRAB-2	59	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$