



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/24/2018

TEMP1 IN: 103 09/24/2018 07:00 TEMP OUT: 104 09/24/2018 08:10

TEMP2 IN: 104 09/24/2018 09:00 TEMP OUT: 104 09/24/2018 10:10

TEMP3 IN: 105 09/24/2018 12:00 TEMP OUT: 105 09/24/2018 13:35

TEMP4 IN: 105 09/24/2018 14:15 TEMP OUT: 105 09/24/2018 16:00

Run Number: LB98101

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB98101BL	LB98101BL	1.4060	1.4060	100	1.4060	1.4060	1.4060	0.0000	0
2	LB98101BS	LB98101BS	1.3901	1.3901	100	1.4409	1.4409	1.4409	0.0508	508
3	J5052-01	092018-JETT-F-D-B	1.3908	1.3908	200	1.3988	1.3988	1.3988	0.0080	40
4	J5052-01DUP	092018-JETT-F-D-BDUP	1.3970	1.3970	200	1.4049	1.4049	1.4049	0.0079	39.5
5	J5052-02	092018-JETT-F-D-S	1.4114	1.4114	500	1.4161	1.4161	1.4161	0.0047	9.4
6	J5052-03	092018-JETT-F-U-B	1.4088	1.4088	500	1.4211	1.4211	1.4211	0.0123	24.6
7	J5052-04	092018-JETT-F-U-S	1.3997	1.3997	500	1.4053	1.4053	1.4053	0.0056	11.2
8	J5052-05	092018-DBRI-E-D-B	1.4044	1.4044	500	1.4142	1.4142	1.4142	0.0098	19.6
9	J5052-06	092018-DBRI-E-D-M	1.3912	1.3912	500	1.3993	1.3993	1.3993	0.0081	16.2
10	J5052-07	092018-DBRI-E-D-S	1.3975	1.3975	500	1.4033	1.4033	1.4033	0.0058	11.6
11	J5052-08	092018-DBRI-E-U-B	1.4034	1.4034	500	1.4122	1.4122	1.4122	0.0088	17.6
12	J5052-09	092018-DBRI-E-U-M	1.3996	1.3996	500	1.4050	1.4050	1.4050	0.0054	10.8
13	J5052-10	092018-DBRI-E-U-S	1.3972	1.3972	500	1.4025	1.4025	1.4025	0.0053	10.6
14	J5052-11	092018-TIPC-E-D-B	1.3966	1.3966	500	1.4075	1.4075	1.4075	0.0109	21.8
15	J5052-12	092018-TIPC-E-D-S	1.3899	1.3899	500	1.3960	1.3960	1.3960	0.0061	12.2

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 09/24/2018

TEMP1 IN: 103 09/24/2018 07:00 TEMP OUT: 104 09/24/2018 08:10

TEMP2 IN: 104 09/24/2018 09:00 TEMP OUT: 104 09/24/2018 10:10

TEMP3 IN: 105 09/24/2018 12:00 TEMP OUT: 105 09/24/2018 13:35

TEMP4 IN: 105 09/24/2018 14:15 TEMP OUT: 105 09/24/2018 16:00

Run Number: LB98101

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J5052-13	092018-TIPC-E-U-B	1.4067	1.4067	500	1.4148	1.4148	1.4148	0.0081	16.2
17	J5052-14	092018-TIPC-E-U-S	1.4033	1.4033	500	1.4105	1.4105	1.4105	0.0072	14.4
18	J5052-15	092018-SIDI-E-D-B	1.3950	1.3950	300	1.4058	1.4058	1.4058	0.0108	36
19	J5052-16	092018-SIDI-E-D-S	1.3944	1.3944	500	1.4029	1.4029	1.4029	0.0085	17
20	J5052-17	092018-SIDI-E-U-B	1.4072	1.4072	500	1.4197	1.4197	1.4197	0.0125	25
21	J5052-18	092018-SIDI-E-U-S	1.3942	1.3942	200	1.3964	1.3964	1.3964	0.0022	11
22	J5052-18DUP	092018-SIDI-E-U-SDUP	1.3937	1.3937	200	1.3958	1.3958	1.3958	0.0021	10.5

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tss092418-2

WorkList ID : 117073

Date : 9/24/2018 8:07:48 AM

2398101

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/27/2018	Water	J5052-01 C	TSS	Cool 4 deg C	HDR103	M41	092018-JETT-F-D-B	09/20/2018	SM2540 D
09/27/2018	Water	J5052-02 C	TSS	Cool 4 deg C	HDR103	M41	092018-JETT-F-D-S	09/20/2018	SM2540 D
09/27/2018	Water	J5052-03 C	TSS	Cool 4 deg C	HDR103	M41	092018-JETT-F-U-B	09/20/2018	SM2540 D
09/27/2018	Water	J5052-04 C	TSS	Cool 4 deg C	HDR103	M41	092018-JETT-F-U-S	09/20/2018	SM2540 D
09/27/2018	Water	J5052-05 C	TSS	Cool 4 deg C	HDR103	M41	092018-DBRI-E-D-B	09/20/2018	SM2540 D
09/27/2018	Water	J5052-06 C	TSS	Cool 4 deg C	HDR103	M41	092018-DBRI-E-D-M	09/20/2018	SM2540 D
09/27/2018	Water	J5052-07 C	TSS	Cool 4 deg C	HDR103	M41	092018-DBRI-E-D-S	09/20/2018	SM2540 D
09/27/2018	Water	J5052-08 C	TSS	Cool 4 deg C	HDR103	M41	092018-DBRI-E-U-B	09/20/2018	SM2540 D
09/27/2018	Water	J5052-09 C	TSS	Cool 4 deg C	HDR103	M41	092018-DBRI-E-U-M	09/20/2018	SM2540 D
09/27/2018	Water	J5052-10 C	TSS	Cool 4 deg C	HDR103	M41	092018-DBRI-E-U-S	09/20/2018	SM2540 D
09/25/2018	Water	J5052-11 C	TSS	Cool 4 deg C	HDR103	M41	092018-TIPC-E-D-B	09/20/2018	SM2540 D
09/25/2018	Water	J5052-12 C	TSS	Cool 4 deg C	HDR103	M41	092018-TIPC-E-D-S	09/20/2018	SM2540 D
09/25/2018	Water	J5052-13 C	TSS	Cool 4 deg C	HDR103	M41	092018-TIPC-E-U-B	09/20/2018	SM2540 D
09/25/2018	Water	J5052-14 C	TSS	Cool 4 deg C	HDR103	M41	092018-TIPC-E-U-S	09/20/2018	SM2540 D
09/27/2018	Water	J5052-15 C	TSS	Cool 4 deg C	HDR103	M41	092018-SIDI-E-D-B	09/20/2018	SM2540 D
09/27/2018	Water	J5052-16 C	TSS	Cool 4 deg C	HDR103	M41	092018-SIDI-E-D-S	09/20/2018	SM2540 D
09/27/2018	Water	J5052-17 C	TSS	Cool 4 deg C	HDR103	M41	092018-SIDI-E-U-B	09/20/2018	SM2540 D
09/27/2018	Water	J5052-18 C	TSS	Cool 4 deg C	HDR103	M41	092018-SIDI-E-U-S	09/20/2018	SM2540 D

FP-9823888

Date/Time

Received by:

Relinquished by:

09/24/18, 08:00
CP

Date/Time

Received by:

Relinquished by:

09/24/18, 12:30
AS