


7471B

LB98102
INSTRUMENT ID: CV2

Linear

 μ Abs.:

40952



A= 0.0000e+000

B= 2.4457e-004

C= -3.0346e-003

Rho= 0.9999774

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	29	0.000	29				
0.2	0.200	0.180	-0.020	748	0.0 %	748				
2.5	2.500	2.545	0.045	10420	0.0 %	10420				
5.0	5.000	4.970	-0.030	20333	0.0 %	20333				
7.5	7.500	7.488	-0.012	30630	0.0 %	30630				
10.0	10.000	10.013	0.013	40952	0.0 %	40952				

LB98102 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Sep 2018 08:46:35

Sample ID	Extended ID	µ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	29	-	0.00007471B	PPB	24 Sep 2018 08:52:30	S	Std
0.2 - 1	50.2	748	-	0.20007471B	PPB	24 Sep 2018 08:54:47	S	Std
2.5 - 1	52.5	10420	-	2.50007471B	PPB	24 Sep 2018 08:57:05	S	Std
5.0 - 1	85	20333	-	5.00007471B	PPB	24 Sep 2018 08:59:23	S	Std
7.5 - 1	57.5	30630	-	7.50007471B	PPB	24 Sep 2018 09:01:40	S	Std
10.0 - 1	57.5	40952	-	10.00007471B	PPB	24 Sep 2018 09:06:36	S	Std
ICV78 - 1	ICV78	16588	4.0539	-7471B	PPB	24 Sep 2018 09:09:29	U	SMPL
ICB78 - 1	ICB78	-209	-0.0541	-7471B	PPB	24 Sep 2018 09:11:45	U	SMPL
CCV51 - 1	CCV51	21564	5.2709	-7471B	PPB	24 Sep 2018 09:14:02	U	SMPL
CCB51 - 1	CCB51	-174	-0.0456	-7471B	PPB	24 Sep 2018 09:16:18	U	SMPL
CRA - 1	CRA	706	0.1696	-7471B	PPB	24 Sep 2018 09:18:34	U	SMPL
HighStd - 1	HighStd	41779	10.2149	-7471B	PPB	24 Sep 2018 09:20:50	U	SMPL
ChkStd - 1	ChkStd	29644	7.2470	-7471B	PPB	24 Sep 2018 09:23:06	U	SMPL
PB113102BL - 1	PBS	-7	-0.0047	-7471B	PPB	24 Sep 2018 09:28:05	U	SMPL
PB113102BS - 1	LCSS	16485	4.0287	-7471B	PPB	24 Sep 2018 09:30:22	U	SMPL
J4777-11 - 1	PL-02-0920-18-A	799	0.1924	-7471B	PPB	24 Sep 2018 09:37:29	U	SMPL
J4777-11DUP - 1	PL-02-0920-18-ADUP	784	0.1887	-7471B	PPB	24 Sep 2018 09:39:46	U	SMPL
J4777-11MS - 1	PL-02-0920-18-AMS	17468	4.2691	-7471B	PPB	24 Sep 2018 09:42:04	U	SMPL
J4777-11MSD - 1	PL-02-0920-18-AMSD	17592	4.2995	-7471B	PPB	24 Sep 2018 09:44:21	U	SMPL
J4777-13 - 1	PL-02-0920-18-B	586	0.1403	-7471B	PPB	24 Sep 2018 09:46:37	U	SMPL
CCV52 - 1	CCV52	21017	5.1371	-7471B	PPB	24 Sep 2018 09:48:54	U	SMPL
CCB52 - 1	CCB52	-121	-0.0326	-7471B	PPB	24 Sep 2018 09:51:09	U	SMPL
J4777-15 - 1	PL-02-0920-18-C	1629	0.3954	-7471B	PPB	24 Sep 2018 09:53:26	U	SMPL
J4777-17 - 1	PL-02-0920-18-D	266	0.0620	-7471B	PPB	24 Sep 2018 09:55:42	U	SMPL
J4777-19 - 1	PL-03-092018-A	2179	0.5299	-7471B	PPB	24 Sep 2018 09:57:58	U	SMPL
J4777-21 - 1	PL-03-092018-B	1783	0.4330	-7471B	PPB	24 Sep 2018 10:00:14	U	SMPL
J4777-23 - 1	PL-03-092018-C	1366	0.3310	-7471B	PPB	24 Sep 2018 10:02:31	U	SMPL
J4777-25 - 1	PL-04-092018-A	2203	0.5358	-7471B	PPB	24 Sep 2018 10:04:48	U	SMPL
J4777-27 - 1	PL-04-092018-B	900	0.2171	-7471B	PPB	24 Sep 2018 10:07:05	U	SMPL
J4777-29 - 1	PL-04-092018-C	668	0.1603	-7471B	PPB	24 Sep 2018 10:09:23	U	SMPL
J4777-31 - 1	PL-04-092018-D	1616	0.3922	-7471B	PPB	24 Sep 2018 10:11:41	U	SMPL
J5021-01 - 1	LS-SOIL-A	4471	1.0904	-7471B	PPB	24 Sep 2018 10:13:57	U	SMPL
CCV53 - 1	CCV53	21508	5.2572	-7471B	PPB	24 Sep 2018 10:16:14	U	SMPL
CCB53 - 1	CCB53	-104	-0.0285	-7471B	PPB	24 Sep 2018 10:18:30	U	SMPL
J5021-05 - 1	LS-SOIL-B	4809	1.1731	-7471B	PPB	24 Sep 2018 10:20:47	U	SMPL
J5021-09 - 1	LS-SOIL-C	5197	1.2680	-7471B	PPB	24 Sep 2018 10:23:03	U	SMPL
J5021-13 - 1	LS-SOIL-D	5198	1.2682	-7471B	PPB	24 Sep 2018 10:25:20	U	SMPL
J5021-17 - 1	LS-SOIL-E	3961	0.9657	-7471B	PPB	24 Sep 2018 10:27:36	U	SMPL
PB113103BL - 1	PBS	-72	-0.0206	-7471B	PPB	24 Sep 2018 10:29:52	U	SMPL
PB113103BS - 1	LCSS	16325	3.9896	-7471B	PPB	24 Sep 2018 10:32:09	U	SMPL
J5017-01 - 1	RA-091918-NORTHEX-NW-055	5093	1.2426	-7471B	PPB	24 Sep 2018 10:34:27	U	SMPL
J5017-02 - 1	RA-091918-NORTHEX-FL-056	1088	0.2631	-7471B	PPB	24 Sep 2018 10:36:45	U	SMPL
J5017-03 - 1	RA-091918-NORTHEX-EW-057	2830	0.6891	-7471B	PPB	24 Sep 2018 10:39:02	U	SMPL
J5017-04 - 1	RA-091918-SOUTHEX-SW-058	4163	1.0151	-7471B	PPB	24 Sep 2018 10:41:19	U	SMPL
CCV54 - 1	CCV54	21552	5.2680	-7471B	PPB	24 Sep 2018 10:43:35	U	SMPL
CCB54 - 1	CCB54	-108	-0.0294	-7471B	PPB	24 Sep 2018 10:45:52	U	SMPL
J5017-05 - 1	RA-091918-SOUTHEX-EW-059	2517	0.6126	-7471B	PPB	24 Sep 2018 10:48:08	U	SMPL
J5017-06 - 1	RA-091918-SOUTHEX-NW-E-060	2723	0.6629	-7471B	PPB	24 Sep 2018 10:50:24	U	SMPL
J5017-07 - 1	RA-091918-SOUTHEX-NW-LOW-061	2988	0.7277	-7471B	PPB	24 Sep 2018 10:52:41	U	SMPL
J5017-07DUP - 1	RA-091918-SOUTHEX-NW-LOW-061DUP	2689	0.6546	-7471B	PPB	24 Sep 2018 10:57:38	U	SMPL
J5017-08 - 1	J5017-07MS	18941	4.6294	-7471B	PPB	24 Sep 2018 10:59:54	U	SMPL
J5017-09 - 1	J5017-07MSD	18390	4.4946	-7471B	PPB	24 Sep 2018 11:02:11	U	SMPL
J5017-10 - 1	RA-091918-NORTHEX-WW-062	3629	0.8845	-7471B	PPB	24 Sep 2018 11:04:29	U	SMPL
J5023-01 - 1	NB-PIT-1	7101	1.7337	-7471B	PPB	24 Sep 2018 11:06:47	U	SMPL
J5023-05 - 1	NB-PIT-2	9374	2.2896	-7471B	PPB	24 Sep 2018 11:09:05	U	SMPL
J5023-09 - 1	NB-PIT-3	1892	0.4597	-7471B	PPB	24 Sep 2018 11:11:22	U	SMPL
CCV55 - 1	CCV55	21681	5.2995	-7471B	PPB	24 Sep 2018 11:13:38	U	SMPL
CCB55 - 1	CCB55	-114	-0.0309	-7471B	PPB	24 Sep 2018 11:15:55	U	SMPL
J5025-01 - 1	FK-PS-01-041	779	0.1875	-7471B	PPB	24 Sep 2018 11:18:12	U	SMPL
PB113123BL - 1	PBS	-7	-0.0047	-7471B	PPB	24 Sep 2018 11:24:19	U	SMPL
PB113123BS - 1	LCSS	14884	3.6372	-7471B	PPB	24 Sep 2018 11:26:35	U	SMPL
J5044-01 - 1	GRID01-COMP-092018	3774	0.9200	-7471B	PPB	24 Sep 2018 11:28:51	U	SMPL
J5044-01DUP - 1	GRID01-COMP-092018DUP	3740	0.9117	-7471B	PPB	24 Sep 2018 11:31:08	U	SMPL
J5044-01MS - 1	GRID01-COMP-092018MS	20181	4.9327	-7471B	PPB	24 Sep 2018 11:33:25	U	SMPL
J5044-01MSD - 1	GRID01-COMP-092018MSD	20678	5.0542	-7471B	PPB	24 Sep 2018 11:35:42	U	SMPL
J5044-02 - 1	GRID02-COMP-092018	6220	1.5182	-7471B	PPB	24 Sep 2018 11:37:59	U	SMPL
J5044-03 - 1	GRID03-COMP-092018	3250	0.7918	-7471B	PPB	24 Sep 2018 11:40:18	U	SMPL
J5044-04 - 1	GRID04-COMP-092018	2791	0.6796	-7471B	PPB	24 Sep 2018 11:42:35	U	SMPL
CCV56 - 1	CCV56	21485	5.2516	-7471B	PPB	24 Sep 2018 11:44:51	U	SMPL
CCB56 - 1	CCB56	-67	-0.0194	-7471B	PPB	24 Sep 2018 11:47:08	U	SMPL
J5044-05 - 1	GRID05-COMP-092018	3539	0.8625	-7471B	PPB	24 Sep 2018 11:49:24	U	SMPL
J5044-06 - 1	GRID06-COMP-092018	3109	0.7573	-7471B	PPB	24 Sep 2018 11:51:41	U	SMPL
J5044-07 - 1	GRID07-COMP-092018	28234	6.9022	-7471B	PPB	24 Sep 2018 11:53:58	U	SMPL
J5044-08 - 1	GRID08-COMP-092018	5342	1.3035	-7471B	PPB	24 Sep 2018 11:56:15	U	SMPL
J5044-09 - 1	GRID09-COMP-092018	2825	0.6879	-7471B	PPB	24 Sep 2018 11:58:33	U	SMPL
J5046-01 - 1	S1	3494	0.8515	-7471B	PPB	24 Sep 2018 12:00:50	U	SMPL
J4777-11LX5 - 1		131	0.0290	-7471B	PPB	24 Sep 2018 12:03:07	U	SMPL
J4777-11A - 1		17863	4.3657	-7471B	PPB	24 Sep 2018 12:05:25	U	SMPL
J5017-07LX5 - 1		424	0.1007	-7471B	PPB	24 Sep 2018 12:07:43	U	SMPL
J5017-07A - 1		20042	4.8987	-7471B	PPB	24 Sep 2018 12:09:59	U	SMPL
CCV57 - 1	CCV57	21396	5.2298	-7471B	PPB	24 Sep 2018 12:12:17	U	SMPL
CCB57 - 1	CCB57	-123	-0.0331	-7471B	PPB	24 Sep 2018 12:14:33	U	SMPL
J5044-01LX5 - 1		740	0.1779	-7471B	PPB	24 Sep 2018 12:16:51	U	SMPL
J5044-01A - 1		21499	5.2550	-7471B	PPB	24 Sep 2018 12:19:08	U	SMPL
CCV58 - 1	CCV58	21642	5.2900	-7471B	PPB	24 Sep 2018 12:21:25	U	SMPL
CCB58 - 1	CCB58	-132	-0.0353	-7471B	PPB	24 Sep 2018 12:23:42	U	SMPL