



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB98105

Slope : 99.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.9	4.01	09/21/2018	18:00
2	CAL2	1	Water			22.5	7.00	09/21/2018	18:02
3	CAL3	1	Water			22.3	10.04	09/21/2018	18:04
4	ICV	1	Water			22.7	7.01	09/21/2018	18:06
5	CCV1	1	Water			22.7	2.00	09/21/2018	18:46
6	J5056-02	1	Water			22.7	6.94	09/21/2018	18:48
7	J5056-02DUP	1	Water			22.8	6.95	09/21/2018	18:50
8	CCV2	1	Water			22.8	12.00	09/21/2018	18:52

LB 98105

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity092118-1

WorkList ID : 117058

Date : 9/21/2018 5:01:35 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/21/2018	Water	J5056-02	Corrosivity	Cool 4 deg C	PSEG01	M33	RA-BASELINE-WATER	09/21/2018	9040C

Date/Time 09/21/18 5:50
 Received by: SP
 Relinquished by: CP

Date/Time 09/21/18 7:00
 Received by: CP
 Relinquished by: SP