



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB98115

Ambient Barometric Pressure (mmHg): 765

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.3	09/24/2018	10:10
2	J5056-02		1	9	>100.0	>212.0	>212.0	09/24/2018	10:40
3	J5056-02DUP		1	9	>100.0	>212.0	>212.0	09/24/2018	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92418

WorkList ID : 117087

Date : 9/24/2018 9:17:03 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/26/2018	Water	J5056-02	C Flash Point	Cool 4 deg C	PSEG01	M33	RA-BASELINE-WATER	09/21/2018	1010A

LB 98715

Date/Time 09/24/18 9:25AM
 Received by: KS
 Relinquished by: CP

Date/Time 09/24/18 11:25AM
 Received by: CP
 Relinquished by: KS