

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/24/2018 14:38

Test: *Reactive*
Total CN

Sample Id	<i>Hyd</i> Result	Dil. 1 +	Response	Errors
ICV1	97.225	0.0	0.090	
ICB1	-0.051	0.0	0.004	
CCV1	244.392	0.0	0.221	
CCB1	-0.162	0.0	0.004	
PB113145BL	-0.216	0.0	0.004	
J4770-02	-0.314	0.0	0.004	
J4770-02DUP	-0.096	0.0	0.004	
J4770-04	-0.224	0.0	0.004	
J4770-06	-0.016	0.0	0.004	
J4770-08	-0.073	0.0	0.004	
J4770-10	-0.191	0.0	0.004	
J4770-12	-0.067	0.0	0.004	
J4770-14	-0.232	0.0	0.004	
CCV2	243.184	0.0	0.220	
CCB2	0.194	0.0	0.004	
J4770-16	-0.056	0.0	0.004	
J4770-18	-0.306	0.0	0.004	
J5057-10	0.085	0.0	0.004	
J5059-04	0.011	0.0	0.004	
J5061-04	0.815	0.0	0.005	
J5061-09	-0.002	0.0	0.004	
J5061-14	-0.035	0.0	0.004	
PB113149BL	-0.065	0.0	0.004	
CCV3	248.108	0.0	0.224	
CCB3	-0.076	0.0	0.004	
J5056-02	-0.027	0.0	0.004	
J5056-02DUP	-0.077	0.0	0.004	
J4770-20	-0.080	0.0	0.004	
CCV4	259.913	0.0	0.235	
CCB4	0.052	0.0	0.004	

N	30
Mean	36.387
SD	86.6447
CV%	238.12

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

9/24/2018 13:30


Test Total CN

Accepted 9/24/2018 13:30

Factor 1125
Bias 0.004

Coeff. of det. 0.999961

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-0.4630	0.0000	
2	5.0PPBCN	0.008	4.5823	5.0000	
3	10PPBCN	0.013	9.7681	10.0000	
4	50PPBCN	0.048	50.1619	50.0000	
5	100PPBCN	0.092	99.6625	100.0000	
6	250PPBCN	0.228	252.4563	250.0000	
7	500PPBCN	0.447	498.8320	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Mon Sep 24 12:21:40 2018

Mon Sep 24 14:36:48 2018

AS

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-0.463	µg/l	9/24/2018 13:30:37	
5.0PPBCN	A	Total CN	P	4.5823	µg/l	9/24/2018 13:30:38	
10PPBCN	A	Total CN	P	9.7681	µg/l	9/24/2018 13:30:39	
50PPBCN	A	Total CN	P	50.1619	µg/l	9/24/2018 13:30:40	
100PPBCN	A	Total CN	P	99.6625	µg/l	9/24/2018 13:30:41	
250PPBCN	A	Total CN	P	252.4563	µg/l	9/24/2018 13:30:42	
500PPBCN	A	Total CN	P	498.832	µg/l	9/24/2018 13:30:43	
ICV1	S	Total CN	P	97.2253	µg/l	9/24/2018 14:06:20	
ICB1	S	Total CN	P	-0.0512	µg/l	9/24/2018 14:06:21	
CCV1	S	Total CN	P	244.3918	µg/l	9/24/2018 14:06:22	
CCB1	S	Total CN	P	-0.1619	µg/l	9/24/2018 14:06:23	
PB113145BL	S	Total CN	P	-0.2158	µg/l	9/24/2018 14:06:24	
J4770-02	S	Total CN	P	-0.3145	µg/l	9/24/2018 14:06:25	
J4770-02DUP	S	Total CN	P	-0.0958	µg/l	9/24/2018 14:06:26	
J4770-04	S	Total CN	P	-0.2238	µg/l	9/24/2018 14:06:27	
J4770-06	S	Total CN	P	-0.0164	µg/l	9/24/2018 14:06:28	
J4770-08	S	Total CN	P	-0.0732	µg/l	9/24/2018 14:06:29	
J4770-10	S	Total CN	P	-0.1909	µg/l	9/24/2018 14:06:30	
J4770-12	S	Total CN	P	-0.0668	µg/l	9/24/2018 14:13:55	
J4770-14	S	Total CN	P	-0.2324	µg/l	9/24/2018 14:13:56	
CCV2	S	Total CN	P	243.1839	µg/l	9/24/2018 14:13:57	
CCB2	S	Total CN	P	0.1938	µg/l	9/24/2018 14:13:58	
J4770-16	S	Total CN	P	-0.0556	µg/l	9/24/2018 14:13:59	
J4770-18	S	Total CN	P	-0.306	µg/l	9/24/2018 14:14:00	
J5057-10	S	Total CN	P	0.0854	µg/l	9/24/2018 14:14:02	
J5059-04	S	Total CN	P	0.0108	µg/l	9/24/2018 14:14:03	
J5061-04	S	Total CN	P	0.8148	µg/l	9/24/2018 14:14:04	
J5061-09	S	Total CN	P	-0.0021	µg/l	9/24/2018 14:14:05	
J5061-14	S	Total CN	P	-0.0346	µg/l	9/24/2018 14:20:05	
PB113149BL	S	Total CN	P	-0.0652	µg/l	9/24/2018 14:20:06	
CCV3	S	Total CN	P	248.1083	µg/l	9/24/2018 14:20:07	
CCB3	S	Total CN	P	-0.076	µg/l	9/24/2018 14:20:08	
J5056-02	S	Total CN	P	-0.0266	µg/l	9/24/2018 14:20:09	
J5056-02DUP	S	Total CN	P	-0.0775	µg/l	9/24/2018 14:20:10	
J4770-20	S	Total CN	P	-0.0803	µg/l	9/24/2018 14:36:46	
CCV4	S	Total CN	P	259.9133	µg/l	9/24/2018 14:36:47	
CCB4	S	Total CN	P	0.0522	µg/l	9/24/2018 14:36:48	