

245.1


LB98125

INSTRUMENT ID: CV1

Linear

 μ Abs.:

33369



A= 0.0000e+000

B= 2.9992e-004

C= -6.4059e-003

Rho= 0.9997960

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	56	0.000	56				
0.05	0.050	0.046	-0.004	175	0.0 %	175				
0.2	0.200	0.167	-0.033	578	0.0 %	578				
2.5	2.500	2.466	-0.034	8244	0.0 %	8244				
5.0	5.000	5.162	0.162	17233	0.0 %	17233				
7.5	7.500	7.397	-0.103	24683	0.0 %	24683				
10.0	10.000	10.002	0.002	33369	0.0 %	33369				

LB98125 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 25 Sep 2018 09:00:52				
Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type
0.0 - 1	50	56	-	0.000245.1	PPB		25 Sep 2018 09:08:40	S Std
0.05 - 1	50.2	175	-	0.0500245.1	PPB		25 Sep 2018 09:11:00	S Std
0.2 - 1	52.5	578	-	0.2000245.1	PPB		25 Sep 2018 09:13:17	S Std
2.5 - 1	55	8244	-	2.5000245.1	PPB		25 Sep 2018 09:15:35	S Std
5.0 - 1		17233	-	5.0000245.1	PPB		25 Sep 2018 09:17:52	S Std
7.5 - 1	57.5	24683	-	7.5000245.1	PPB		25 Sep 2018 09:20:09	S Std
10.0 - 1	510	33369	-	10.0000245.1	PPB		25 Sep 2018 09:35:44	S Std
ICV80 - 1	ICV80	13040	3.9046	-245.1	PPB		25 Sep 2018 09:38:34	U SMPL
ICB80 - 1	ICB80	-41	-0.0187	-245.1	PPB		25 Sep 2018 09:43:28	U SMPL
CCV63 - 1	CCV63	17780	5.3262	-245.1	PPB		25 Sep 2018 09:45:43	U SMPL
CCB63 - 1	CCB63	-135	-0.0469	-245.1	PPB		25 Sep 2018 09:47:59	U SMPL
CRA - 1	CRA	181	0.0479	-245.1	PPB		25 Sep 2018 09:50:16	U SMPL
HighStd - 1	HighStd	33666	10.0908	-245.1	PPB		25 Sep 2018 09:52:32	U SMPL
ChkStd - 1	ChkStd	23662	7.0904	-245.1	PPB		25 Sep 2018 09:54:48	U SMPL
PB113170BL - 1	PBW	81	0.0179	-245.1	PPB		25 Sep 2018 09:59:50	U SMPL
PB113170BS - 1	LCSW	12633	3.7825	-245.1	PPB		25 Sep 2018 10:02:07	U SMPL
J5033-24 - 1	BPSI-TT-MW314D-20180919	-138	-0.0478	-245.1	PPB		25 Sep 2018 10:04:23	U SMPL
J5033-24DUP - 1	BPSI-TT-MW314D-20180919DUP	39	0.0053	-245.1	PPB		25 Sep 2018 10:06:42	U SMPL
J5033-24MS - 1	BPSI-TT-MW314D-20180919MS	13247	3.9667	-245.1	PPB		25 Sep 2018 10:09:00	U SMPL
J5033-24MSD - 1	BPSI-TT-MW314D-20180919MSD	13604	4.0737	-245.1	PPB		25 Sep 2018 10:11:19	U SMPL
J5033-25 - 1	BPSI-TT-MW301S-20180918	-182	-0.0610	-245.1	PPB		25 Sep 2018 10:13:37	U SMPL
CCV64 - 1	CCV64	17672	5.2938	-245.1	PPB		25 Sep 2018 10:15:53	U SMPL
CCB64 - 1	CCB64	-160	-0.0544	-245.1	PPB		25 Sep 2018 10:18:09	U SMPL
J5033-26 - 1	BB-TT-DUP05-20180918	-26	-0.0142	-245.1	PPB		25 Sep 2018 10:20:26	U SMPL
J5033-27 - 1	BP-HN-MW27I-20180918	78	0.0170	-245.1	PPB		25 Sep 2018 10:22:42	U SMPL
J5033-28 - 1	BPSI-TT-MW301D-20180918	51	0.0089	-245.1	PPB		25 Sep 2018 10:24:58	U SMPL
J5033-29 - 1	BPSI-TT-MW301I-20180918	57	0.0107	-245.1	PPB		25 Sep 2018 10:27:15	U SMPL
J5033-30 - 1	BP-MH-SW400I-20180918	133	0.0335	-245.1	PPB		25 Sep 2018 10:29:32	U SMPL
J5033-31 - 1	BP-TT-SW4004-20180918	0	-0.0064	-245.1	PPB		25 Sep 2018 10:31:49	U SMPL
J5033-32 - 1	BP-TT-SW4002-20180918	-65	-0.0259	-245.1	PPB		25 Sep 2018 10:34:06	U SMPL
J5033-33 - 1	BPSI-TT-MW312I-20180918	-106	-0.0382	-245.1	PPB		25 Sep 2018 10:36:24	U SMPL
J5033-34 - 1	BPSI-TT-MW312S-20180918	-58	-0.0238	-245.1	PPB		25 Sep 2018 10:38:42	U SMPL
J5033-35 - 1	BPSI-TT-MW310S-20180919	-19	-0.0121	-245.1	PPB		25 Sep 2018 10:40:58	U SMPL
CCV65 - 1	CCV65	16881	5.0566	-245.1	PPB		25 Sep 2018 10:43:15	U SMPL
CCB65 - 1	CCB65	-167	-0.0565	-245.1	PPB		25 Sep 2018 10:45:31	U SMPL
J5033-36 - 1	BPSI-TT-MW311S-20180919	-89	-0.0331	-245.1	PPB		25 Sep 2018 10:47:49	U SMPL
J5033-37 - 1	BPSI-TT-MW311I-20180919	12	-0.0028	-245.1	PPB		25 Sep 2018 10:50:05	U SMPL
J5033-38 - 1	BPSI-TT-MW314S-20180919	105	0.0251	-245.1	PPB		25 Sep 2018 10:52:22	U SMPL
J5033-39 - 1	BPSI-TT-MW313S-2018020	106	0.0254	-245.1	PPB		25 Sep 2018 10:54:38	U SMPL
J5033-40 - 1	BB-TT-DUP06-20180920	57	0.0107	-245.1	PPB		25 Sep 2018 10:56:55	U SMPL
J5033-41 - 1	BPSI-TT-MW307D-2018020	40	0.0056	-245.1	PPB		25 Sep 2018 10:59:12	U SMPL
PB113171BL - 1	PBW	37	0.0047	-245.1	PPB		25 Sep 2018 11:01:29	U SMPL
PB113171BS - 1	LCSW	12318	3.6880	-245.1	PPB		25 Sep 2018 11:03:49	U SMPL
J5051-01 - 1	092118-SIDI-F-D-B	61	0.0119	-245.1	PPB		25 Sep 2018 11:06:08	U SMPL
J5051-02 - 1	092118-SIDI-F-D-M	8	-0.0040	-245.1	PPB		25 Sep 2018 11:08:25	U SMPL
CCV66 - 1	CCV66	17362	5.2008	-245.1	PPB		25 Sep 2018 11:10:42	U SMPL
CCB66 - 1	CCB66	-236	-0.0772	-245.1	PPB		25 Sep 2018 11:12:59	U SMPL
J5051-03 - 1	092118-SIDI-F-D-S	-202	-0.0670	-245.1	PPB		25 Sep 2018 11:15:17	U SMPL
J5051-04 - 1	092118-SIDI-F-U-B	-77	-0.0295	-245.1	PPB		25 Sep 2018 11:17:33	U SMPL
J5051-05 - 1	092118-SIDI-F-U-M	-85	-0.0319	-245.1	PPB		25 Sep 2018 11:19:50	U SMPL
J5051-06 - 1	092118-SIDI-F-U-S	1	-0.0061	-245.1	PPB		25 Sep 2018 11:22:06	U SMPL
J5051-07 - 1	092118-TIPC-F-D-B	-13	-0.0103	-245.1	PPB		25 Sep 2018 11:24:23	U SMPL
J5051-08 - 1	092118-TIPC-F-D-S	102	0.0242	-245.1	PPB		25 Sep 2018 11:26:40	U SMPL
J5051-09 - 1	092118-TIPC-F-U-B	12	-0.0028	-245.1	PPB		25 Sep 2018 11:28:57	U SMPL
J5051-10 - 1	092118-TIPC-F-U-S	-35	-0.0169	-245.1	PPB		25 Sep 2018 11:31:15	U SMPL
J5051-10DUP - 1	092118-TIPC-F-U-SDUP	-27	-0.0145	-245.1	PPB		25 Sep 2018 11:33:33	U SMPL
J5051-10MS - 1	092118-TIPC-F-U-SMS	14282	4.2771	-245.1	PPB		25 Sep 2018 11:35:49	U SMPL
CCV67 - 1	CCV67	17339	5.1939	-245.1	PPB		25 Sep 2018 11:38:06	U SMPL
CCB67 - 1	CCB67	-161	-0.0547	-245.1	PPB		25 Sep 2018 11:40:22	U SMPL
J5051-10MSD - 1	092118-TIPC-F-U-SMSD	14652	4.3881	-245.1	PPB		25 Sep 2018 11:42:41	U SMPL
PB113172BL - 1	PBW	-226	-0.0742	-245.1	PPB		25 Sep 2018 11:44:58	U SMPL
PB113172BS - 1	LCSW	13442	4.0251	-245.1	PPB		25 Sep 2018 11:47:15	U SMPL
J5052-01 - 1	092018-JETT-F-D-B	-97	-0.0355	-245.1	PPB		25 Sep 2018 11:49:31	U SMPL
J5052-02 - 1	092018-JETT-F-D-S	-26	-0.0142	-245.1	PPB		25 Sep 2018 11:51:48	U SMPL
J5052-03 - 1	092018-JETT-F-U-B	14	-0.0022	-245.1	PPB		25 Sep 2018 11:54:06	U SMPL
J5052-04 - 1	092018-JETT-F-U-S	-31	-0.0157	-245.1	PPB		25 Sep 2018 11:56:23	U SMPL
J5052-05 - 1	092018-DBRI-E-D-B	64	0.0128	-245.1	PPB		25 Sep 2018 11:58:40	U SMPL
J5052-06 - 1	092018-DBRI-E-D-M	-24	-0.0136	-245.1	PPB		25 Sep 2018 12:00:59	U SMPL
J5052-07 - 1	092018-DBRI-E-D-S	31	0.0029	-245.1	PPB		25 Sep 2018 12:03:16	U SMPL
CCV68 - 1	CCV68	16852	5.0479	-245.1	PPB		25 Sep 2018 12:05:33	U SMPL
CCB68 - 1	CCB68	-116	-0.0412	-245.1	PPB		25 Sep 2018 12:07:49	U SMPL
J5052-08 - 1	092018-DBRI-E-U-B	16	-0.0016	-245.1	PPB		25 Sep 2018 12:10:09	U SMPL
J5052-09 - 1	092018-DBRI-E-U-M	28	0.0020	-245.1	PPB		25 Sep 2018 12:12:26	U SMPL
J5052-10 - 1	092018-DBRI-E-U-S	34	0.0038	-245.1	PPB		25 Sep 2018 12:14:43	U SMPL
J5052-11 - 1	092018-TIPC-E-D-B	60	0.0116	-245.1	PPB		25 Sep 2018 12:17:00	U SMPL
J5052-12 - 1	092018-TIPC-E-D-S	-30	-0.0154	-245.1	PPB		25 Sep 2018 12:19:16	U SMPL
J5052-13 - 1	092018-TIPC-E-U-B	35	0.0041	-245.1	PPB		25 Sep 2018 12:21:33	U SMPL
J5052-14 - 1	092018-TIPC-E-U-S	93	0.0215	-245.1	PPB		25 Sep 2018 12:23:51	U SMPL
J5052-15 - 1	092018-SIDI-E-D-B	34	0.0038	-245.1	PPB		25 Sep 2018 12:26:09	U SMPL
J5052-16 - 1	092018-SIDI-E-D-S	23	0.0005	-245.1	PPB		25 Sep 2018 12:28:27	U SMPL
J5052-17 - 1	092018-SIDI-E-U-B	47	0.0077	-245.1	PPB		25 Sep 2018 12:30:44	U SMPL
CCV69 - 1	CCV69	16321	4.8886	-245.1	PPB		25 Sep 2018 12:33:01	U SMPL
CCB69 - 1	CCB69	-148	-0.0508	-245.1	PPB		25 Sep 2018 12:35:18	U SMPL
J5052-18 - 1	092018-SIDI-E-U-S	17	-0.0013	-245.1	PPB		25 Sep 2018 12:37:36	U SMPL
J5052-18DUP - 1	092018-SIDI-E-U-SDUP	42	0.0062	-245.1	PPB		25 Sep 2018 12:39:53	U SMPL
J5052-18MS - 1	092018-SIDI-E-U-SMS	13779	4.1262	-245.1	PPB		25 Sep 2018 12:42:10	U SMPL
J5052-18MSD - 1	092018-SIDI-E-U-SMSD	15525	4.6499	-245.1	PPB		25 Sep 2018 12:44:27	U SMPL
PB113173BL - 1	PBW	-64	-0.0256	-245.1	PPB		25 Sep 2018 12:46:45	U SMPL
PB113173BS - 1	LCSW	12756	3.8194	-245.1	PPB		25 Sep 2018 12:49:02	U SMPL
J5019-01 - 1	091918-DBRI-E-D-B	-30	-0.0154	-245.1	PPB		25 Sep 2018 12:51:20	U SMPL
J5019-02 - 1	091918-DBRI-E-D-M	132	0.0332	-245.1	PPB		25 Sep 2018 12:53:39	U SMPL
J5019-03 - 1	091918-DBRI-E-D-S	75	0.0161	-245.1	PPB		25 Sep 2018 12:55:57	U SMPL
J5019-04 - 1	091918-DBRI-E-U-B	19	-0.0007	-245.1	PPB		25 Sep 2018 12:58:14	U SMPL

LB98125 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 25 Sep 2018 09:00:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCV70 - 1	CCV70	16354	4.8985	-245.1		PPB	25 Sep 2018 13:00:31	U	SMPL
CCB70 - 1	CCB70	-202	-0.0670	-245.1		PPB	25 Sep 2018 13:02:48	U	SMPL
J5019-05 - 1	091918-DBRI-E-U-M	36	0.0044	-245.1		PPB	25 Sep 2018 13:05:07	U	SMPL
J5019-06 - 1	091918-DBRI-E-U-S	57	0.0107	-245.1		PPB	25 Sep 2018 13:07:24	U	SMPL
J5019-07 - 1	091918-SIDI-E-D-B	101	0.0239	-245.1		PPB	25 Sep 2018 13:09:41	U	SMPL
J5019-08 - 1	091918-SIDI-E-D-S	86	0.0194	-245.1		PPB	25 Sep 2018 13:11:58	U	SMPL
J5019-09 - 1	091918-SIDI-E-U-B	42	0.0062	-245.1		PPB	25 Sep 2018 13:14:15	U	SMPL
J5019-10 - 1	091918-SIDI-E-U-S	129	0.0323	-245.1		PPB	25 Sep 2018 13:16:32	U	SMPL
J5019-11 - 1	091918-JETT-E-D-B	126	0.0314	-245.1		PPB	25 Sep 2018 13:18:50	U	SMPL
J5019-12 - 1	091918-JETT-E-D-M	176	0.0464	-245.1		PPB	25 Sep 2018 13:21:07	U	SMPL
J5019-13 - 1	091918-JETT-E-D-S	133	0.0335	-245.1		PPB	25 Sep 2018 13:23:25	U	SMPL
CCV71 - 1	CCV71	16516	4.9471	-245.1		PPB	25 Sep 2018 13:27:59	U	SMPL
CCB71 - 1	CCB71	-160	-0.0544	-245.1		PPB	25 Sep 2018 13:30:16	U	SMPL
J5019-15 - 1	091918-JETT-E-U-M	30	0.0026	-245.1		PPB	25 Sep 2018 13:32:35	U	SMPL
J5019-16 - 1	091918-JETT-E-U-S	51	0.0089	-245.1		PPB	25 Sep 2018 13:34:52	U	SMPL
J5019-17 - 1	091918-TIPC-F-D-B	47	0.0077	-245.1		PPB	25 Sep 2018 13:37:10	U	SMPL
J5019-18 - 1	091918-TIPC-F-D-S	132	0.0332	-245.1		PPB	25 Sep 2018 13:39:27	U	SMPL
J5019-19 - 1	091918-TIPC-F-U-B	167	0.0437	-245.1		PPB	25 Sep 2018 13:41:45	U	SMPL
J5019-20 - 1	091918-TIPC-F-U-S	152	0.0392	-245.1		PPB	25 Sep 2018 13:44:01	U	SMPL
J5019-20DUP - 1	091918-TIPC-F-U-SDUP	119	0.0293	-245.1		PPB	25 Sep 2018 13:46:19	U	SMPL
J5019-20MS - 1	091918-TIPC-F-U-SMS	12488	3.7390	-245.1		PPB	25 Sep 2018 13:51:11	U	SMPL
J5019-20MSD - 1	091918-TIPC-F-U-SMSD	13058	3.9100	-245.1		PPB	25 Sep 2018 13:53:28	U	SMPL
J5019-14 - 1	091918-JETT-EU-B	-87	-0.0325	-245.1		PPB	25 Sep 2018 13:55:48	U	SMPL
CCV72 - 1	CCV72	16762	5.0209	-245.1		PPB	25 Sep 2018 13:58:07	U	SMPL
CCB72 - 1	CCB72	-183	-0.0613	-245.1		PPB	25 Sep 2018 14:00:24	U	SMPL
J5033-24LX5 - 1		-36	-0.0172	-245.1		PPB	25 Sep 2018 14:02:44	U	SMPL
J5033-24A - 1		12843	3.8455	-245.1		PPB	25 Sep 2018 14:05:01	U	SMPL
J5051-10LX5 - 1		-118	-0.0418	-245.1		PPB	25 Sep 2018 14:07:19	U	SMPL
J5051-10A - 1		14403	4.3134	-245.1		PPB	25 Sep 2018 14:09:36	U	SMPL
J5052-18LX5 - 1		-84	-0.0316	-245.1		PPB	25 Sep 2018 14:11:53	U	SMPL
J5052-18A - 1		13844	4.1457	-245.1		PPB	25 Sep 2018 14:14:12	U	SMPL
J5019-20LX5 - 1		-49	-0.0211	-245.1		PPB	25 Sep 2018 14:16:29	U	SMPL
J5019-20A - 1		12793	3.8305	-245.1		PPB	25 Sep 2018 14:44:51	U	SMPL
CCV73 - 1	CCV73	16148	4.8367	-245.1		PPB	25 Sep 2018 14:47:09	U	SMPL
CCB73 - 1	CCB73	-184	-0.0616	-245.1		PPB	25 Sep 2018 14:49:28	U	SMPL