

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB98130

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP68078
calibration std. phosphate 0.5 ppm	WP68077
calibration std. phosphate 0.3 ppm	WP68076
calibration std. 0 ppm	WP68073
phosphate ICV std.	WP68079
phosphate CCV std.	WP68080
calibration std. phosphate 0.1 ppm	WP68075
calibration std. phosphate 0.05 ppm	WP68074
Combined reagent	WP68083
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	09/25/2018	12:16
2	CAL2	0.05	1	50	50	0.033	09/25/2018	12:18
3	CAL3	0.10	1	50	50	0.065	09/25/2018	12:18
4	CAL4	0.30	1	50	50	0.186	09/25/2018	12:19
5	CAL5	0.50	1	50	50	0.330	09/25/2018	12:19
6	CAL6	1.00	1	50	50	0.649	09/25/2018	12:20



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB98130

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.315	0.486	09/25/2018	12:21
2	ICB		1	50	50	0.000	0.001	09/25/2018	12:22
3	CCV1	0.50	1	50	50	0.328	0.506	09/25/2018	12:22
4	CCB1		1	50	50	0.000	0.001	09/25/2018	12:23
5	PB113195BL		1	50	50	0.000	0.001	09/25/2018	12:24
6	PB113195BS	0.50	1	50	50	0.318	0.490	09/25/2018	12:24
7	J4880-13		1	50	50	1.611	2.478	09/25/2018	12:24
8	J4880-13DUP		1	50	50	1.577	2.426	09/25/2018	12:25
9	J4880-13		10	50	50	0.588	0.905	09/25/2018	12:39
10	J4880-13DUP		10	50	50	0.601	0.925	09/25/2018	12:40
11	CCV2	0.50	1	50	50	0.328	0.506	09/25/2018	12:40
12	CCB2		1	50	50	0.000	0.001	09/25/2018	12:41