


7471B

LB 98134
INSTRUMENT ID: LV2

Linear

 μ Abs.:

4154E



A= 0.0000e+000

B= 2.4069e-004

C= 9.0348e-003

Rho= 0.9999452

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	-22	0.000	-22				
0.2	0.200	0.202	0.002	801	0.0 %	801				
2.5	2.500	2.538	0.038	10509	0.0 %	10509				
5.0	5.000	4.919	-0.081	20399	0.0 %	20399				
7.5	7.500	7.528	0.028	31240	0.0 %	31240				
10.0	10.000	10.009	0.009	41546	0.0 %	41546				

LB98134 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Sep 2018 15:10:41

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.00007471B	PPB	25 Sep 2018 15:11:50	S	Std
0.2 - 1	50.2	801	-	0.20007471B	PPB	25 Sep 2018 15:14:08	S	Std
2.5 - 1	52.5	10509	-	2.50007471B	PPB	25 Sep 2018 15:16:25	S	Std
5.0 - 1	55	20399	-	5.00007471B	PPB	25 Sep 2018 15:18:43	S	Std
7.5 - 1	57.5	31240	-	7.50007471B	PPB	25 Sep 2018 15:21:01	S	Std
10.0 - 1	510	41546	-	10.00007471B	PPB	25 Sep 2018 15:23:18	S	Std
ICV81 - 1	ICV81	15840	3.8216	-7471B	PPB	25 Sep 2018 15:25:58	U	SMPL
ICB81 - 1	ICB81	-225	-0.0451	-7471B	PPB	25 Sep 2018 15:28:16	U	SMPL
CCV74 - 1	CCV74	20459	4.9333	-7471B	PPB	25 Sep 2018 15:30:32	U	SMPL
CCB74 - 1	CCB74	-257	-0.0528	-7471B	PPB	25 Sep 2018 15:32:48	U	SMPL
CRA - 1	CRA	739	0.1869	-7471B	PPB	25 Sep 2018 15:35:05	U	SMPL
HighStd - 1	HighStd	41384	9.9698	-7471B	PPB	25 Sep 2018 15:37:21	U	SMPL
ChkStd - 1	ChkStd	28523	6.8743	-7471B	PPB	25 Sep 2018 15:39:37	U	SMPL
PB113174BL - 1	PBS	-62	-0.0059	-7471B	PPB	25 Sep 2018 15:41:56	U	SMPL
PB113174BS - 1	LCSS	15171	3.6606	-7471B	PPB	25 Sep 2018 15:44:14	U	SMPL
J4770-01 - 1	JC-02-092118-A	14814	3.5746	-7471B	PPB	25 Sep 2018 15:46:31	U	SMPL
J4770-01DUP - 1	JC-02-092118-ADUP	13053	3.1508	-7471B	PPB	25 Sep 2018 15:48:50	U	SMPL
J4770-01MS - 1	JC-02-092118-AMS	29953	7.2185	-7471B	PPB	25 Sep 2018 15:51:09	U	SMPL
J4770-01MSD - 1	JC-02-092118-AMSD	30878	7.4411	-7471B	PPB	25 Sep 2018 15:53:32	U	SMPL
J4770-03 - 1	JC-02-092118-B	3116	0.7590	-7471B	PPB	25 Sep 2018 15:55:51	U	SMPL
CCV75 - 1	CCV75	20411	4.9218	-7471B	PPB	25 Sep 2018 15:58:09	U	SMPL
CCB75 - 1	CCB75	-163	-0.0302	-7471B	PPB	25 Sep 2018 16:00:24	U	SMPL
J4770-05 - 1	JC-02-092118-C	5777	1.3995	-7471B	PPB	25 Sep 2018 16:02:44	U	SMPL
J4770-07 - 1	JC-02-092118-D	9029	2.1822	-7471B	PPB	25 Sep 2018 16:05:00	U	SMPL
J4770-09 - 1	JC-02-092118-E	9897	2.3912	-7471B	PPB	25 Sep 2018 16:07:16	U	SMPL
J4770-11 - 1	JC-02-092118-F	15782	3.8076	-7471B	PPB	25 Sep 2018 16:09:34	U	SMPL
J4770-13 - 1	JC-02-092118-G	10344	2.4988	-7471B	PPB	25 Sep 2018 16:11:51	U	SMPL
J4770-15 - 1	JC-02-092118-H	9191	2.2212	-7471B	PPB	25 Sep 2018 16:14:09	U	SMPL
J4770-17 - 1	JC-02-092118-I	13863	3.3457	-7471B	PPB	25 Sep 2018 16:16:26	U	SMPL
J4770-19 - 1	JC-02-092118-J	5776	1.3993	-7471B	PPB	25 Sep 2018 16:18:44	U	SMPL
J5057-01 - 1	HC-SOIL	1223	0.3034	-7471B	PPB	25 Sep 2018 16:21:02	U	SMPL
J5059-01 - 1	PHASE-2	1948	0.4779	-7471B	PPB	25 Sep 2018 16:23:18	U	SMPL
CCV76 - 1	CCV76	20616	4.9711	-7471B	PPB	25 Sep 2018 16:25:34	U	SMPL
CCB76 - 1	CCB76	-162	-0.0300	-7471B	PPB	25 Sep 2018 16:27:50	U	SMPL
J5061-01 - 1	230-KV-A	406	0.1068	-7471B	PPB	25 Sep 2018 16:30:07	U	SMPL
J5061-06 - 1	230-KV-B	752	0.1900	-7471B	PPB	25 Sep 2018 16:32:23	U	SMPL
J5061-11 - 1	230-KV-C	644	0.1640	-7471B	PPB	25 Sep 2018 16:34:40	U	SMPL
PB113175BL - 1	PBS	12	0.0119	-7471B	PPB	25 Sep 2018 16:36:56	U	SMPL
PB113175BS - 1	LCSS	15439	3.7251	-7471B	PPB	25 Sep 2018 16:44:15	U	SMPL
J4774-05 - 1	OK-03-092418	1607	0.3958	-7471B	PPB	25 Sep 2018 16:46:32	U	SMPL
J4775-09 - 1	OR-01-092418-A	1158	0.2878	-7471B	PPB	25 Sep 2018 16:48:51	U	SMPL
J4775-09DUP - 1	OR-01-092418-ADUP	1049	0.2615	-7471B	PPB	25 Sep 2018 16:51:09	U	SMPL
J4775-09MS - 1	OR-01-092418-AMS	17570	4.2380	-7471B	PPB	25 Sep 2018 16:56:48	U	SMPL
J4775-09MSD - 1	OR-01-092418-AMSD	17069	4.1174	-7471B	PPB	25 Sep 2018 16:59:04	U	SMPL
CCV77 - 1	CCV77	20562	4.9581	-7471B	PPB	25 Sep 2018 17:01:21	U	SMPL
CCB77 - 1	CCB77	-121	-0.0201	-7471B	PPB	25 Sep 2018 17:03:40	U	SMPL
J4775-11 - 1	OR-01-092418-B	2013	0.4935	-7471B	PPB	25 Sep 2018 17:05:57	U	SMPL
J4779-05 - 1	SU-02-092418-A	1464	0.3614	-7471B	PPB	25 Sep 2018 17:08:16	U	SMPL
J4779-07 - 1	SU-02-092418-B	1262	0.3128	-7471B	PPB	25 Sep 2018 17:10:33	U	SMPL
J4779-09 - 1	SU-3A-092418	2914	0.7104	-7471B	PPB	25 Sep 2018 17:12:49	U	SMPL
J5055-04 - 1	TP-A	532	0.1371	-7471B	PPB	25 Sep 2018 17:15:06	U	SMPL
J5055-08 - 1	TP-B	468	0.1217	-7471B	PPB	25 Sep 2018 17:17:23	U	SMPL
J5055-12 - 1	TP-C	93	0.0314	-7471B	PPB	25 Sep 2018 17:19:40	U	SMPL
J5055-16 - 1	TP-D	215	0.0608	-7471B	PPB	25 Sep 2018 17:21:57	U	SMPL
J5064-01 - 1	COMP	175	0.0512	-7471B	PPB	25 Sep 2018 17:24:16	U	SMPL
J5069-01 - 1	SLUDGE-BOX-A	121393	29.2273	-7471B	PPB	25 Sep 2018 17:37:25	U	SMPL
CCV78 - 1	CCV78	19627	4.7331	-7471B	PPB	25 Sep 2018 17:39:41	U	SMPL
CCB78 - 1	CCB78	-188	-0.0362	-7471B	PPB	25 Sep 2018 17:42:02	U	SMPL
J5069-05 - 1	SLUDGE-BOX-B	100915	24.2984	-7471B	PPB	25 Sep 2018 17:44:19	U	SMPL
J5070-01 - 1	CS-RBR-200027	1689	0.4156	-7471B	PPB	25 Sep 2018 17:50:43	U	SMPL
J5070-04 - 1	CS-RBR-200031	1452	0.3585	-7471B	PPB	25 Sep 2018 17:52:59	U	SMPL
J5070-07 - 1	CS-RT-4005	2078	0.5092	-7471B	PPB	25 Sep 2018 17:55:16	U	SMPL
J5070-10 - 1	CS-RBR-200028	3697	0.8989	-7471B	PPB	25 Sep 2018 17:57:33	U	SMPL
J4770-01LX5 - 1		2456	0.6002	-7471B	PPB	25 Sep 2018 17:59:50	U	SMPL
J4770-01A - 1		29553	7.1222	-7471B	PPB	25 Sep 2018 18:02:07	U	SMPL
J4775-09LX5 - 1		31	0.0165	-7471B	PPB	25 Sep 2018 18:04:25	U	SMPL
J4775-09A - 1		17136	4.1335	-7471B	PPB	25 Sep 2018 18:06:45	U	SMPL
CCV79 - 1	CCV79	20427	4.9256	-7471B	PPB	25 Sep 2018 18:09:01	U	SMPL
CCB79 - 1	CCB79	-176	-0.0333	-7471B	PPB	25 Sep 2018 18:11:20	U	SMPL
J5069-01DLX5 - 1		22577	5.4431	-7471B	PPB	25 Sep 2018 18:13:37	U	SMPL
J5069-05DLX5 - 1		18936	4.5668	-7471B	PPB	25 Sep 2018 18:15:53	U	SMPL
CCV80 - 1	CCV80	20546	4.9543	-7471B	PPB	25 Sep 2018 18:18:10	U	SMPL
CCB80 - 1	CCB80	-193	-0.0374	-7471B	PPB	25 Sep 2018 18:20:29	U	SMPL