


7470A

LB98153
INSTRUMENT ID: CVJ

Linear

 μ Abs.:

29282



A= 0.0000e+000

B= 3.4142e-004

C= -7.3590e-003

Rho= 0.9998708

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.034	0.034	122	0.000	122				
0.2	0.200	0.228	0.028	688	0.0 %	688				
2.5	2.500	2.377	-0.123	6983	0.0 %	6983				
5.0	5.000	5.062	0.062	14847	0.0 %	14847				
7.5	7.500	7.510	0.010	22017	0.0 %	22017				
10.0	10.000	9.990	-0.010	29282	0.0 %	29282				

LB98153 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Sep 2018 10:56:11

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	122	-	0.00007470A	PPB	26 Sep 2018 11:00:44	S	Std
0.2 - 1	502	688	-	0.20007470A	PPB	26 Sep 2018 11:02:52	S	Std
2.5 - 1	52.5	6983	-	2.50007470A	PPB	26 Sep 2018 11:05:01	S	Std
5.0 - 1	55	14847	-	5.00007470A	PPB	26 Sep 2018 11:07:10	S	Std
7.5 - 1	57.5	22017	-	7.50007470A	PPB	26 Sep 2018 11:09:18	S	Std
10.0 - 1	510	29282	-	10.00007470A	PPB	26 Sep 2018 11:11:26	S	Std
ICV83 - 1	ICV83	12616	4.3000	-7470A	PPB	26 Sep 2018 11:14:04	U	SMPL
ICB83 - 1	ICB83	28	0.0022	-7470A	PPB	26 Sep 2018 11:16:12	U	SMPL
CCV88 - 1	CCV88	15435	5.2624	-7470A	PPB	26 Sep 2018 11:21:55	U	SMPL
CCB88 - 1	CCB88	-123	-0.0494	-7470A	PPB	26 Sep 2018 11:24:01	U	SMPL
CRA - 1	CRA	649	0.2142	-7470A	PPB	26 Sep 2018 11:26:09	U	SMPL
HighStd - 1	HighStd	27571	9.4059	-7470A	PPB	26 Sep 2018 11:28:15	U	SMPL
ChkStd - 1	ChkStd	19347	6.5981	-7470A	PPB	26 Sep 2018 11:30:22	U	SMPL
PB113211BL - 1	PBW	-29	-0.0173	-7470A	PPB	26 Sep 2018 11:32:33	U	SMPL
PB113211BS - 1	LCSW	10984	3.7428	-7470A	PPB	26 Sep 2018 11:34:44	U	SMPL
PB113211TB - 1	PB113166TB	3	-0.0063	-7470A	PPB	26 Sep 2018 11:36:53	U	SMPL
J4774-06 - 1	OK-03-092418	58	0.0124	-7470A	PPB	26 Sep 2018 11:39:02	U	SMPL
J4774-06DUP - 1	OK-03-092418DUP	108	0.0295	-7470A	PPB	26 Sep 2018 11:43:51	U	SMPL
J4774-06MS - 1	OK-03-092418MS	13916	4.7438	-7470A	PPB	26 Sep 2018 11:47:34	U	SMPL
J4774-06MSD - 1	OK-03-092418MSD	13647	4.6520	-7470A	PPB	26 Sep 2018 11:49:42	U	SMPL
CCV89 - 1	CCV89	14290	4.8715	-7470A	PPB	26 Sep 2018 11:55:07	U	SMPL
CCB89 - 1	CCB89	-26	-0.0162	-7470A	PPB	26 Sep 2018 11:57:14	U	SMPL
J4775-10 - 1	OR-01-092418-A	88	0.0227	-7470A	PPB	26 Sep 2018 11:59:23	U	SMPL
J4775-12 - 1	OR-01-092418-B	87	0.0223	-7470A	PPB	26 Sep 2018 12:01:29	U	SMPL
J4779-06 - 1	SU-02-092418-A	129	0.0367	-7470A	PPB	26 Sep 2018 12:03:37	U	SMPL
J4779-08 - 1	SU-02-092418-B	255	0.0797	-7470A	PPB	26 Sep 2018 12:05:44	U	SMPL
J4779-10 - 1	SU-3A-092418	228	0.0705	-7470A	PPB	26 Sep 2018 12:07:51	U	SMPL
J5055-07 - 1	TP-A	119	0.0333	-7470A	PPB	26 Sep 2018 12:09:59	U	SMPL
J5055-11 - 1	TP-B	321	0.1022	-7470A	PPB	26 Sep 2018 12:12:09	U	SMPL
J5055-15 - 1	TP-C	105	0.0285	-7470A	PPB	26 Sep 2018 12:14:17	U	SMPL
J5055-19 - 1	TP-D	330	0.1053	-7470A	PPB	26 Sep 2018 12:16:28	U	SMPL
J5064-04 - 1	COMP	204	0.0623	-7470A	PPB	26 Sep 2018 12:18:35	U	SMPL
CCV90 - 1	CCV90	14011	4.7763	-7470A	PPB	26 Sep 2018 12:26:28	U	SMPL
CCB90 - 1	CCB90	-245	-0.0910	-7470A	PPB	26 Sep 2018 12:28:35	U	SMPL
J5069-04 - 1	SLUDGE-A	727	0.2409	-7470A	PPB	26 Sep 2018 12:30:44	U	SMPL
J5069-08 - 1	SLUDGE-B	248	0.0773	-7470A	PPB	26 Sep 2018 12:32:52	U	SMPL
J5070-03 - 1	CS-RBR-200027	61	0.0135	-7470A	PPB	26 Sep 2018 12:35:00	U	SMPL
J5070-06 - 1	CS-RBR-200031	255	0.0797	-7470A	PPB	26 Sep 2018 12:37:07	U	SMPL
J5070-09 - 1	CS-RT-4005	159	0.0469	-7470A	PPB	26 Sep 2018 12:39:15	U	SMPL
J5070-12 - 1	CS-RBR-200028	170	0.0507	-7470A	PPB	26 Sep 2018 12:41:24	U	SMPL
PB113212BL - 1	PBW	171	0.0510	-7470A	PPB	26 Sep 2018 12:43:34	U	SMPL
PB113212BS - 1	LCSW	13119	4.4717	-7470A	PPB	26 Sep 2018 12:45:42	U	SMPL
PB113212TB - 1	PB113176TB	-6	-0.0094	-7470A	PPB	26 Sep 2018 12:47:51	U	SMPL
J4281-03 - 1	DP-DRUM-3-OILY-DEBRIS	52	0.0104	-7470A	PPB	26 Sep 2018 12:49:58	U	SMPL
CCV91 - 1	CCV91	15803	5.3881	-7470A	PPB	26 Sep 2018 12:54:33	U	SMPL
CCB91 - 1	CCB91	-182	-0.0695	-7470A	PPB	26 Sep 2018 12:56:42	U	SMPL
J4281-03DUP - 1	DP-DRUM-3-OILY-DEBRISDUP	140	0.0404	-7470A	PPB	26 Sep 2018 12:58:50	U	SMPL
J4281-03MS - 1	DP-DRUM-3-OILY-DEBRISMS	13155	4.4840	-7470A	PPB	26 Sep 2018 13:00:59	U	SMPL
J4281-03MSD - 1	DP-DRUM-3-OILY-DEBRISMSD	12908	4.3997	-7470A	PPB	26 Sep 2018 13:03:06	U	SMPL
J4923-01 - 1	BS-PAINT-CHIPS	8377	2.8527	-7470A	PPB	26 Sep 2018 13:05:13	U	SMPL
PB113213BL - 1	PBW	3	-0.0063	-7470A	PPB	26 Sep 2018 13:07:20	U	SMPL
PB113213BS - 1	LCSW	13159	4.4854	-7470A	PPB	26 Sep 2018 13:12:06	U	SMPL
PB113213TB - 1	PB113193TB	32	0.0036	-7470A	PPB	26 Sep 2018 13:14:14	U	SMPL
J5055-21 - 1	OILY-WATER	170	0.0507	-7470A	PPB	26 Sep 2018 13:16:24	U	SMPL
J5056-02 - 1	RA-BASELINE-WATER	198	0.0602	-7470A	PPB	26 Sep 2018 13:18:33	U	SMPL
J5056-02DUP - 1	RA-BASELINE-WATERDUP	203	0.0619	-7470A	PPB	26 Sep 2018 13:20:42	U	SMPL
CCV92 - 1	CCV92	15139	5.1614	-7470A	PPB	26 Sep 2018 13:22:49	U	SMPL
CCB92 - 1	CCB92	-124	-0.0497	-7470A	PPB	26 Sep 2018 13:25:00	U	SMPL
J5056-02MS - 1	RA-BASELINE-WATERMS	14794	5.0436	-7470A	PPB	26 Sep 2018 13:27:08	U	SMPL
J5056-02MSD - 1	RA-BASELINE-WATERMSD	14569	4.9668	-7470A	PPB	26 Sep 2018 13:29:15	U	SMPL
J5072-07 - 1	MH-9	-9	-0.0104	-7470A	PPB	26 Sep 2018 13:31:23	U	SMPL
J5072-08 - 1	MH-8	177	0.0531	-7470A	PPB	26 Sep 2018 13:33:31	U	SMPL
J5072-09 - 1	MH-4	137	0.0394	-7470A	PPB	26 Sep 2018 13:35:41	U	SMPL
J5072-10 - 1	MH-5	82	0.0206	-7470A	PPB	26 Sep 2018 13:37:48	U	SMPL
J5072-11 - 1	MH-726	199	0.0606	-7470A	PPB	26 Sep 2018 13:39:56	U	SMPL
J5072-12 - 1	MH-725	74	0.0179	-7470A	PPB	26 Sep 2018 13:42:05	U	SMPL
PB113214BL - 1	PBW	185	0.0558	-7470A	PPB	26 Sep 2018 13:47:21	U	SMPL
PB113214BS - 1	LCSW	12248	4.1743	-7470A	PPB	26 Sep 2018 13:49:28	U	SMPL
CCV93 - 1	CCV93	14474	4.9343	-7470A	PPB	26 Sep 2018 13:51:35	U	SMPL
CCB93 - 1	CCB93	-51	-0.0248	-7470A	PPB	26 Sep 2018 13:53:44	U	SMPL
J5055-20 - 1	OILY-WATER	175	0.0524	-7470A	PPB	26 Sep 2018 13:55:54	U	SMPL
J5072-01 - 1	MH-9	242	0.0753	-7470A	PPB	26 Sep 2018 13:58:02	U	SMPL
J5072-01DUP - 1	MH-9DUP	286	0.0903	-7470A	PPB	26 Sep 2018 14:00:09	U	SMPL
J5072-01MS - 1	MH-9MS	13896	4.7370	-7470A	PPB	26 Sep 2018 14:02:18	U	SMPL
J5072-01MSD - 1	MH-9MSD	14414	4.9138	-7470A	PPB	26 Sep 2018 14:04:26	U	SMPL
J5072-02 - 1	MH-8	141	0.0408	-7470A	PPB	26 Sep 2018 14:06:34	U	SMPL
J5072-03 - 1	MH-4	194	0.0589	-7470A	PPB	26 Sep 2018 14:08:42	U	SMPL
J5072-04 - 1	MH-5	228	0.0705	-7470A	PPB	26 Sep 2018 14:10:52	U	SMPL
J5072-05 - 1	MH-726	249	0.0777	-7470A	PPB	26 Sep 2018 14:13:01	U	SMPL
J5072-06 - 1	MH-725	293	0.0927	-7470A	PPB	26 Sep 2018 14:15:08	U	SMPL
CCV94 - 1	CCV94	14220	4.8476	-7470A	PPB	26 Sep 2018 14:26:53	U	SMPL
CCB94 - 1	CCB94	28	0.0022	-7470A	PPB	26 Sep 2018 14:29:01	U	SMPL
J4774-06LX5 - 1		56	0.0118	-7470A	PPB	26 Sep 2018 14:31:10	U	SMPL
J4774-06A - 1		12944	4.4120	-7470A	PPB	26 Sep 2018 14:33:18	U	SMPL
J4281-03LX5 - 1		-13	-0.0118	-7470A	PPB	26 Sep 2018 14:35:25	U	SMPL
J4281-03A - 1		12219	4.1644	-7470A	PPB	26 Sep 2018 14:37:36	U	SMPL
J5056-02LX5 - 1		73	0.0176	-7470A	PPB	26 Sep 2018 14:39:44	U	SMPL
J5056-02A - 1		11712	3.9913	-7470A	PPB	26 Sep 2018 14:41:53	U	SMPL
J5072-01LX5 - 1		43	0.0073	-7470A	PPB	26 Sep 2018 14:44:01	U	SMPL
J5072-01A - 1		12604	4.2959	-7470A	PPB	26 Sep 2018 14:46:10	U	SMPL
CCV95 - 1	CCV95	15255	5.2010	-7470A	PPB	26 Sep 2018 14:50:45	U	SMPL
CCB95 - 1	CCB95	-23	-0.0152	-7470A	PPB	26 Sep 2018 14:52:52	U	SMPL