



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98154

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.9	4.01	09/26/2018	11:00
2	CAL2	1	Water			22.6	7.00	09/26/2018	11:02
3	CAL3	1	Water			22.8	10.03	09/26/2018	11:04
4	ICV	1	Water			22.7	7.01	09/26/2018	11:06
5	CCV1	1	Water			22.8	2.00	09/26/2018	11:08
6	J5085-01	1	Water			23.2	6.92	09/26/2018	11:10
7	J5085-01DUP	1	Water			23.3	6.96	09/26/2018	11:12
8	J5085-02	1	Water			23.0	6.87	09/26/2018	11:14
9	J5085-03	1	Water			22.9	7.28	09/26/2018	11:16
10	J5085-04	1	Water			23.2	8.04	09/26/2018	11:18
11	J5085-05	1	Water			23.1	7.48	09/26/2018	11:20
12	J5085-06	1	Water			23.0	8.08	09/26/2018	11:22
13	J5085-07	1	Water			23.0	7.76	09/26/2018	11:24
14	CCV2	1	Water			22.6	12.00	09/26/2018	11:26

WORKLIST(Hardcopy Internal Chain)

LB 98154

WorkList Name : pH092618

WorkList ID : 117194

Date : 9/26/2018 10:49:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/25/2018	Water	J5085-01 C	pH	Cool 4 deg C	TULL02	M61	MH-720	09/25/2018	SM 4500-PH B
09/25/2018	Water	J5085-02 C	pH	Cool 4 deg C	TULL02	M61	MH-719	09/25/2018	SM 4500-PH B
09/25/2018	Water	J5085-03 C	pH	Cool 4 deg C	TULL02	M61	MH-717	09/25/2018	SM 4500-PH B
09/25/2018	Water	J5085-04 C	pH	Cool 4 deg C	TULL02	M61	MH-10.25	09/25/2018	SM 4500-PH B
09/25/2018	Water	J5085-05 C	pH	Cool 4 deg C	TULL02	M61	MH-10.29	09/25/2018	SM 4500-PH B
09/25/2018	Water	J5085-06 C	pH	Cool 4 deg C	TULL02	M61	MH-714	09/25/2018	SM 4500-PH B
09/25/2018	Water	J5085-07 C	pH	Cool 4 deg C	TULL02	M61	MH-713	09/25/2018	SM 4500-PH B

Date/Time 09/26/18 10:55

Received by: SP

Relinquished by: CP

Date/Time 09/26/18 11:30

Received by: CP

Relinquished by: SP