



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB98155

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	09/26/2018	10:05
2	J5072-07		1	8	>100.0	>212.0	>212.0	09/26/2018	10:35
3	J5072-07DUP		1	8	>100.0	>212.0	>212.0	09/26/2018	10:55
4	J5072-08		1	6	>100.0	>212.0	>212.0	09/26/2018	11:15
5	J5072-09		1	7	>100.0	>212.0	>212.0	09/26/2018	11:40
6	J5072-10		1	9	>100.0	>212.0	>212.0	09/26/2018	12:05
7	J5072-11		1	7	>100.0	>212.0	>212.0	09/26/2018	12:35
8	J5072-12		1	8	>100.0	>212.0	>212.0	09/26/2018	12:50
9	J5078-02		1	8	>100.0	>212.0	>212.0	09/26/2018	13:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp92618

WorkList ID : 117183

Date : 9/26/2018 9:27:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/08/2018	Water	J5072-07 B	Flash Point	Cool 4 deg C	TULL02	M32	MH-9	09/24/2018	1010A
10/08/2018	Water	J5072-08 B	Flash Point	Cool 4 deg C	TULL02	M32	MH-8	09/24/2018	1010A
10/08/2018	Water	J5072-09 B	Flash Point	Cool 4 deg C	TULL02	M32	MH-4	09/24/2018	1010A
10/08/2018	Water	J5072-10 B	Flash Point	Cool 4 deg C	TULL02	M32	MH-5	09/24/2018	1010A
10/08/2018	Water	J5072-11 B	Flash Point	Cool 4 deg C	TULL02	M32	MH-726	09/24/2018	1010A
10/08/2018	Water	J5072-12 B	Flash Point	Cool 4 deg C	TULL02	M32	MH-725	09/24/2018	1010A
09/28/2018	Water	J5078-02 C	Flash Point	Cool 4 deg C	PSEG01	I21	600-WILSON-AVE-NEWARK	09/25/2018	1010A

Date/Time 09/26/18 9:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 09/26/18 13:20
 Received by: CP
 Relinquished by: KS