



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98162

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	09/26/2018	15:20
2	CAL2	1	Water			22.2	7.00	09/26/2018	15:22
3	CAL3	1	Water			22.4	10.04	09/26/2018	15:24
4	ICV	1	Water			22.6	7.00	09/26/2018	15:26
5	CCV1	1	Water			22.2	2.00	09/26/2018	15:42
6	J5105-01	1	Solid	20.02	20	23.8	6.76	09/26/2018	15:44
7	J5105-01DUP	1	Solid	20.03	20	23.6	6.77	09/26/2018	15:46
8	J5105-02	1	Solid	20.02	20	23.3	6.74	09/26/2018	15:48
9	J5105-03	1	Solid	20.02	20	23.2	7.11	09/26/2018	15:50
10	CCV2	1	Water			22.6	12.00	09/26/2018	15:52

LB 98162

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH092618-1

WorkList ID : 117217

Date : 9/26/2018 2:43:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2018	Solid	J5105-01	pH	Cool 4 deg C	CTIA01	M23	WJ-18-09-26-TSHG-09	09/04/2018	9045D
09/04/2018	Solid	J5105-02	pH	Cool 4 deg C	CTIA01	M23	WJ-18-09-26-TSHG-10	09/04/2018	9045D
09/04/2018	Solid	J5105-03	pH	Cool 4 deg C	CTIA01	M23	WJ-18-09-26-TSHG-11	09/04/2018	9045D

Date/Time 09/26/18 2:45

Received by: SP

Relinquished by: CP

Date/Time 09/26/18 4:00

Received by: SP

Relinquished by: SP