



TOTAL SOLIDS - SM2540B

SUPERVISOR: AMANDEEP

ANALYST: Aparna

Date: 09/27/2018

OVEN TEMP1: 103 °C

TEMP: 105 °C

DATE & TIME IN: 09/26/2018 16:00 OUT: 09/27/2018 07:00

OVEN TEMP2: 105 °C

TEMP: 105 °C

DATE & TIME IN: 09/27/2018 08:20 OUT: 09/27/2018 10:00

Run Number: LB98164

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB98164BL	LB98164BL	82.6138	82.6138	100	82.6138	82.6138	82.6138	0.0000	0
2	J5090-01	MW-1	97.3443	97.3443	100	97.5190	97.5190	97.5190	0.1747	1747
3	J5090-02	MW-2	83.4606	83.4606	100	83.6539	83.6539	83.6539	0.1933	1933
4	J5090-03	MW-4	95.0994	95.0994	100	95.5199	95.5199	95.5199	0.4205	4205
5	J5090-03DUP	MW-4DUP	93.0105	93.0105	100	93.4410	93.4410	93.4410	0.4305	4305

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts092618

WorkList ID : 117198

Date : 9/26/2018 11:07:28 AM

21398164

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/02/2018	Water	J5090-01	TS	Cool 4 deg C	TULL02	M63	MW-1	09/25/2018	SM2540 B
10/02/2018	Water	J5090-02	TS	Cool 4 deg C	TULL02	M63	MW-2	09/25/2018	SM2540 B
10/02/2018	Water	J5090-03	TS	Cool 4 deg C	TULL02	M63	MW-4	09/25/2018	SM2540 B

Date/Time

Received by:

Relinquished by:

09/26/18 14:00
AS
CS

Date/Time

Received by:

Relinquished by:

09/26/18 16:00
CS
AS