



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98168

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	09/26/2018	15:20
2	CAL2	1	Water			22.2	7.00	09/26/2018	15:22
3	CAL3	1	Water			22.4	10.04	09/26/2018	15:24
4	ICV	1	Water			22.6	7.01	09/26/2018	15:26
5	CCV1	1	Water			22.3	2.00	09/26/2018	15:28
6	J5083-04	1	Water			18.9	7.71	09/26/2018	15:30
7	J5083-05	1	Water			19.2	7.28	09/26/2018	15:32
8	J5083-06	1	Water			19.3	7.46	09/26/2018	15:34
9	J5101-02	1	Water			23.8	6.83	09/26/2018	15:36
10	J5101-02DUP	1	Water			23.9	6.84	09/26/2018	15:38
11	CCV2	1	Water			22.4	12.00	09/26/2018	15:40

WORKLIST(Hardcopy Internal Chain)

LB 98168

WorkList Name : corrosivity092618-1

WorkList ID : 117219

Date : 9/26/2018 2:58:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/25/2018	Water	J5083-04	Corrosivity	Cool 4 deg C	ENVI77	M63	TKA-092518	09/25/2018	9040C
09/25/2018	Water	J5083-05	Corrosivity	Cool 4 deg C	ENVI77	M63	TKC-092518	09/25/2018	9040C
09/25/2018	Water	J5083-06	Corrosivity	Cool 4 deg C	ENVI77	M63	TKB-092518	09/25/2018	9040C
09/26/2018	Water	J5101-02	Corrosivity	Cool 4 deg C	PSEG01	I33	C-1329	09/26/2018	9040C

Date/Time 09/26/18 3:00

Received by: SP

Relinquished by: CP

Date/Time 09/26/18 4:00

Received by: CP

Relinquished by: SP