



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB98169

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	09/26/2018	15:20
2	CAL2	1	Water			22.2	7.00	09/26/2018	15:22
3	CAL3	1	Water			22.4	10.04	09/26/2018	15:24
4	ICV	1	Water			22.6	7.01	09/26/2018	15:26
5	CCV1	1	Water			22.3	2.00	09/26/2018	15:28
6	J5101-01	1	Water			23.8	6.83	09/26/2018	15:36
7	J5101-01DUP	1	Water			23.9	6.84	09/26/2018	15:38
8	CCV2	1	Water			22.4	12.00	09/26/2018	15:40

LB 98169

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH092618-3

WorkList ID : 117220

Date : 9/26/2018 2:59:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/26/2018	Water	J5101-01	pH	Cool 4 deg C	PSEG01	I33	C-1329	09/26/2018	9040C

Date/Time 09/26/18 3:00

Received by: spRelinquished by: CP

Date/Time 09/26/18 4:00

Received by: CPRelinquished by: SP