



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB98170

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.6	4.01	09/26/2018	15:20
2	CAL2	1	Water			22.2	7.00	09/26/2018	15:22
3	CAL3	1	Water			22.4	10.04	09/26/2018	15:24
4	ICV	1	Water			22.6	7.00	09/26/2018	15:26
5	CCV1	1	Water			22.4	2.00	09/26/2018	16:00
6	J5096-02	1	Solid	20.02	20	23.6	6.21	09/26/2018	16:02
7	J5096-02DUP	1	Solid	20.04	20	23.7	6.22	09/26/2018	16:04
8	J5097-02	1	Solid	20.02	20	23.5	6.39	09/26/2018	16:06
9	J5097-04	1	Solid	20.03	20	23.4	5.15	09/26/2018	16:08
10	J5097-06	1	Solid	20.02	20	23.6	6.23	09/26/2018	16:10
11	J5097-08	1	Solid	20.02	20	23.4	5.90	09/26/2018	16:12
12	J5099-02	1	Solid	20.04	20	23.2	8.33	09/26/2018	16:14
13	J5100-04	1	Solid	20.02	20	23.6	9.25	09/26/2018	16:16
14	CCV2	1	Water			22.6	12.00	09/26/2018	16:20

LB98170

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity092618

WorkList ID : 117218

Date : 9/26/2018 2:58:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/26/2018	Solid	J5096-02	Corrosivity	Cool 4 deg C	PSEG01	I31	MT-T-5-2	09/26/2018	9045D
09/26/2018	Solid	J5097-02	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-1	09/26/2018	9045D
09/26/2018	Solid	J5097-04	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-2	09/26/2018	9045D
09/26/2018	Solid	J5097-06	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-3	09/26/2018	9045D
09/26/2018	Solid	J5097-08	Corrosivity	Cool 4 deg C	PSEG01	I31	TP-4	09/26/2018	9045D
09/26/2018	Solid	J5099-02	Corrosivity	Cool 4 deg C	PSEG01	I31	7211975	09/26/2018	9045D
09/26/2018	Solid	J5100-04	Corrosivity	Cool 4 deg C	PSEG01	I31	HS-STONE-RO-227	09/26/2018	9045D

Date/Time 09/26/18 3:00

Received by: SP

Relinquished by: CP

Date/Time 09/26/18 4:30

Received by: CP

Relinquished by: SP